

Data Path : Z:\voasrv\HPCHEM1\MSVOA_V\Data\VW121123\
 Data File : VW033234.D
 Acq On : 11 Dec 2023 23:52
 Operator : SY/MD
 Sample : 05751-01
 Misc : 5.0mL/MSVOA_V/WATER
 ALS Vial : 35 Sample Multiplier: 1

Instrument :
 MSVOA_V
 ClientSampleId :
 YCH96

Quant Time: Dec 12 03:27:54 2023
 Quant Method : Z:\voasrv\HPCHEM1\MSVOA_V\Method\SFAMVLM120123WMA.M
 Quant Title : VOC Analysis
 QLast Update : Fri Dec 08 21:51:12 2023
 Response via : Initial Calibration

Compound	R.T.	QIon	Response	Conc	Units	Dev(Min)

Internal Standards						
1) 1,4-Difluorobenzene	5.532	114	277731	50.000	ug/L	0.00
28) Chlorobenzene-d5	8.783	117	279010	50.000	ug/L	0.00
58) 1,4-Dichlorobenzene-d4	11.185	152	144840	50.000	ug/L	0.00
System Monitoring Compounds						
4) Vinyl Chloride-d3	1.278	65	55962	33.706	ug/L	0.00
Spiked Amount	50.000	Range	60 - 135	Recovery	=	67.420%
7) Chloroethane-d5	1.532	69	46538	40.229	ug/L	0.00
Spiked Amount	50.000	Range	70 - 130	Recovery	=	80.460%
11) 1,1-Dichloroethene-d2	2.060	65	24843	39.333	ug/L	0.00
Spiked Amount	50.000	Range	60 - 125	Recovery	=	78.660%
21) 2-Butanone-d5	3.780	46	80528	97.654	ug/L	0.00
Spiked Amount	100.000	Range	40 - 130	Recovery	=	97.650%
24) Chloroform-d	4.246	84	157311	44.647	ug/L	0.00
Spiked Amount	50.000	Range	70 - 125	Recovery	=	89.300%
26) 1,2-Dichloroethane-d4	4.941	65	104413	47.664	ug/L	0.00
Spiked Amount	50.000	Range	70 - 125	Recovery	=	95.320%
32) Benzene-d6	4.960	84	287629	44.078	ug/L	0.00
Spiked Amount	50.000	Range	70 - 125	Recovery	=	88.160%
36) 1,2-Dichloropropane-d6	5.986	67	78542	44.765	ug/L	0.00
Spiked Amount	50.000	Range	70 - 120	Recovery	=	89.540%
41) Toluene-d8	7.243	98	260074	41.780	ug/L	0.00
Spiked Amount	50.000	Range	80 - 120	Recovery	=	83.560%
43) trans-1,3-Dichloroprop...	7.551	79	44081	39.732	ug/L	0.00
Spiked Amount	50.000	Range	60 - 125	Recovery	=	79.460%
47) 2-Hexanone-d5	8.021	63	62699	88.953	ug/L	0.00
Spiked Amount	100.000	Range	45 - 130	Recovery	=	88.950%
56) 1,1,2,2-Tetrachloroeth...	10.153	84	115200	46.188	ug/L	0.00
Spiked Amount	50.000	Range	65 - 120	Recovery	=	92.380%
66) 1,2-Dichlorobenzene-d4	11.561	152	131875	51.173	ug/L	0.00
Spiked Amount	50.000	Range	80 - 120	Recovery	=	102.340%
Target Compounds						
					Qvalue	
13) Acetone	2.108	43	41795	53.091	ug/L	98
22) 2-Butanone	3.870	43	31369	31.311	ug/L	92

(#) = qualifier out of range (m) = manual integration (+) = signals summed

Data Path : Z:\voasrv\HPCHEM1\MSVOA_V\Data\VV121123\
Data File : VV033234.D
Acq On : 11 Dec 2023 23:52
Operator : SY/MD
Sample : 05751-01
Misc : 5.0mL/MSVOA_V/WATER
ALS Vial : 35 Sample Multiplier: 1

Instrument :
MSVOA_V
ClientSampleId :
YCH96

Quant Time: Dec 12 03:27:54 2023
Quant Method : Z:\voasrv\HPCHEM1\MSVOA_V\Method\SFAMVLM120123WMA.M
Quant Title : VOC Analysis
QLast Update : Fri Dec 08 21:51:12 2023
Response via : Initial Calibration

