

Data Path : Z:\voasrv\HPCHEM1\MSVOA_V\Data\VW121123\
 Data File : VW033232.D
 Acq On : 11 Dec 2023 23:05
 Operator : SY/MD
 Sample : 05730-09
 Misc : 5.0mL/MSVOA_V/WATER
 ALS Vial : 33 Sample Multiplier: 1

Instrument :
 MSVOA_V
 ClientSampleId :
 C0GC4

Quant Time: Dec 12 03:27:38 2023
 Quant Method : Z:\voasrv\HPCHEM1\MSVOA_V\Method\SFAMVLM120123WMA.M
 Quant Title : VOC Analysis
 QLast Update : Fri Dec 08 21:51:12 2023
 Response via : Initial Calibration

Compound	R.T.	QIon	Response	Conc	Units	Dev(Min)

Internal Standards						
1) 1,4-Difluorobenzene	5.532	114	275622	50.000	ug/L	0.00
28) Chlorobenzene-d5	8.783	117	275789	50.000	ug/L	0.00
58) 1,4-Dichlorobenzene-d4	11.185	152	140493	50.000	ug/L	0.00
System Monitoring Compounds						
4) Vinyl Chloride-d3	1.278	65	57147	34.684	ug/L	0.00
Spiked Amount	50.000	Range	60 - 135	Recovery	=	69.360%
7) Chloroethane-d5	1.532	69	47158	41.077	ug/L	0.00
Spiked Amount	50.000	Range	70 - 130	Recovery	=	82.160%
11) 1,1-Dichloroethene-d2	2.060	65	24392	38.914	ug/L	0.00
Spiked Amount	50.000	Range	60 - 125	Recovery	=	77.820%
21) 2-Butanone-d5	3.780	46	79346	96.957	ug/L	0.00
Spiked Amount	100.000	Range	40 - 130	Recovery	=	96.960%
24) Chloroform-d	4.249	84	158116	45.219	ug/L	0.00
Spiked Amount	50.000	Range	70 - 125	Recovery	=	90.440%
26) 1,2-Dichloroethane-d4	4.941	65	104944	48.273	ug/L	0.00
Spiked Amount	50.000	Range	70 - 125	Recovery	=	96.540%
32) Benzene-d6	4.960	84	286843	44.471	ug/L	0.00
Spiked Amount	50.000	Range	70 - 125	Recovery	=	88.940%
36) 1,2-Dichloropropane-d6	5.989	67	79730	45.973	ug/L	0.00
Spiked Amount	50.000	Range	70 - 120	Recovery	=	91.940%
41) Toluene-d8	7.243	98	262236	42.619	ug/L	0.00
Spiked Amount	50.000	Range	80 - 120	Recovery	=	85.240%
43) trans-1,3-Dichloroprop...	7.551	79	45344	41.347	ug/L	0.00
Spiked Amount	50.000	Range	60 - 125	Recovery	=	82.700%
47) 2-Hexanone-d5	8.021	63	64398	92.431	ug/L	0.00
Spiked Amount	100.000	Range	45 - 130	Recovery	=	92.430%
56) 1,1,2,2-Tetrachloroeth...	10.153	84	118236	47.959	ug/L	0.00
Spiked Amount	50.000	Range	65 - 120	Recovery	=	95.920%
66) 1,2-Dichlorobenzene-d4	11.561	152	128823	51.535	ug/L	0.00
Spiked Amount	50.000	Range	80 - 120	Recovery	=	103.080%

Target Compounds Qvalue

(#) = qualifier out of range (m) = manual integration (+) = signals summed

Data Path : Z:\voasrv\HPCHEM1\MSVOA_V\Data\VV121123\
Data File : VV033232.D
Acq On : 11 Dec 2023 23:05
Operator : SY/MD
Sample : 05730-09
Misc : 5.0mL/MSVOA_V/WATER
ALS Vial : 33 Sample Multiplier: 1

Instrument :
MSVOA_V
ClientSampleId :
C0GC4

Quant Time: Dec 12 03:27:38 2023
Quant Method : Z:\voasrv\HPCHEM1\MSVOA_V\Method\SFAMVLM120123WMA.M
Quant Title : VOC Analysis
QLast Update : Fri Dec 08 21:51:12 2023
Response via : Initial Calibration

