

Data Path : Z:\voasrv\HPCHEM1\MSVOA_V\DATA\VV121125\
 Data File : VV039535.D
 Acq On : 11 Dec 2025 12:04
 Operator : SY/MD
 Sample : VV1211WBL01
 Misc : 25.0 mL/MSVOA_V/WATER
 ALS Vial : 3 Sample Multiplier: 1

Instrument :
 MSVOA_V
 ClientSampleId :
 VBLK240

Quant Time: Dec 12 00:48:01 2025
 Quant Method : Z:\voasrv\HPCHEM1\MSVOA_V\Method\SFAMVSIM120325MA.M
 Quant Title : TRACE VOA SOM01.0
 QLast Update : Tue Dec 09 08:40:48 2025
 Response via : Initial Calibration

Compound	R.T.	QIon	Response	Conc	Units	Dev(Min)

Internal Standards						
1) 1,4-Difluorobenzene	5.566	114	5311	0.500	ug/L	0.00
5) Chlorobenzene-d5	8.800	117	5097	0.500	ug/L	0.02
12) 1,4-Dichlorobenzene-d4	11.197	152	2635	0.500	ug/L	0.02
System Monitoring Compounds						
2) Vinyl Chloride-d3	1.295	65	3180	0.610	ug/L	0.00
Spiked Amount	0.500	Range	40 - 130	Recovery	= 122.000%	
4) 1,2-Dichloroethane-d4	4.970	65	1762	0.498	ug/L	0.00
Spiked Amount	0.500	Range	70 - 130	Recovery	= 100.000%	
7) 1,2-Dichloropropane-d6	5.996	67	2019	0.469	ug/L	-0.02
Spiked Amount	0.500	Range	60 - 140	Recovery	= 94.000%	
8) Toluene-d8	7.254	98	3786	0.522	ug/L	0.02
Spiked Amount	0.500	Range	70 - 130	Recovery	= 104.000%	
10) 1,1,2,2-Tetrachloroeth...	10.133	84	1427	0.589	ug/L	0.00
Spiked Amount	0.500	Range	65 - 120	Recovery	= 118.000%	

Target Compounds Qvalue

 (#) = qualifier out of range (m) = manual integration (+) = signals summed

Data Path : Z:\voasrv\HPCHEM1\MSVOA_V\DATA\VV121125\
Data File : VV039535.D
Acq On : 11 Dec 2025 12:04
Operator : SY/MD
Sample : VV1211WBL01
Misc : 25.0 mL/MSVOA_V/WATER
ALS Vial : 3 Sample Multiplier: 1

Instrument :
MSVOA_V
ClientSampleId :
VBLK240

Quant Time: Dec 12 00:48:01 2025
Quant Method : Z:\voasrv\HPCHEM1\MSVOA_V\Method\SFAMVSIM120325MA.M
Quant Title : TRACE VOA SOM01.0
QLast Update : Tue Dec 09 08:40:48 2025
Response via : Initial Calibration

