

Data Path : Z:\voasrv\HPCHEM1\MSVOA_V\DATA\VV121125\
 Data File : VV039538.D
 Acq On : 11 Dec 2025 13:46
 Operator : SY/MD
 Sample : Q3828-02
 Misc : 25.0 mL/MSVOA_V/WATER
 ALS Vial : 6 Sample Multiplier: 1

Instrument :
 MSVOA_V
 ClientSampleId :
 BR-02-133-120925-SIM

Quant Time: Dec 12 00:48:35 2025
 Quant Method : Z:\voasrv\HPCHEM1\MSVOA_V\Method\SFAMVSIM120325MA.M
 Quant Title : TRACE VOA SOM01.0
 QLast Update : Tue Dec 09 08:40:48 2025
 Response via : Initial Calibration

Compound	R.T.	QIon	Response	Conc	Units	Dev(Min)

Internal Standards						
1) 1,4-Difluorobenzene	5.566	114	6385	0.500	ug/L	0.00
5) Chlorobenzene-d5	8.800	117	6179	0.500	ug/L	0.02
12) 1,4-Dichlorobenzene-d4	11.197	152	3284	0.500	ug/L	0.02
System Monitoring Compounds						
2) Vinyl Chloride-d3	1.294	65	3756	0.599	ug/L	0.00
Spiked Amount	0.500	Range	40 - 130	Recovery	=	120.000%
4) 1,2-Dichloroethane-d4	4.970	65	2020	0.475	ug/L	0.00
Spiked Amount	0.500	Range	70 - 130	Recovery	=	96.000%
7) 1,2-Dichloropropane-d6	5.996	67	2319	0.445	ug/L	-0.02
Spiked Amount	0.500	Range	60 - 140	Recovery	=	88.000%
8) Toluene-d8	7.254	98	4292	0.488	ug/L	0.02
Spiked Amount	0.500	Range	70 - 130	Recovery	=	98.000%
10) 1,1,2,2-Tetrachloroeth...	10.133	84	1666	0.567	ug/L	0.00
Spiked Amount	0.500	Range	65 - 120	Recovery	=	114.000%

Target Compounds Qvalue

 (#) = qualifier out of range (m) = manual integration (+) = signals summed

Data Path : Z:\voasrv\HPCHEM1\MSVOA_V\DATA\VV121125\
Data File : VV039538.D
Acq On : 11 Dec 2025 13:46
Operator : SY/MD
Sample : Q3828-02
Misc : 25.0 mL/MSVOA_V/WATER
ALS Vial : 6 Sample Multiplier: 1

Instrument :
MSVOA_V
ClientSampleId :
BR-02-133-120925-SIM

Quant Time: Dec 12 00:48:35 2025
Quant Method : Z:\voasrv\HPCHEM1\MSVOA_V\Method\SFAMVSIM120325MA.M
Quant Title : TRACE VOA SOM01.0
QLast Update : Tue Dec 09 08:40:48 2025
Response via : Initial Calibration

