

Data Path : Z:\voasrv\HPCHEM1\MSVOA_V\DATA\VV121125\
 Data File : VV039541.D
 Acq On : 11 Dec 2025 14:50
 Operator : SY/MD
 Sample : Q3828-09
 Misc : 25.0 mL/MSVOA_V/WATER
 ALS Vial : 9 Sample Multiplier: 1

Instrument :
 MSVOA_V
 ClientSampleId :
 TB01-120925

Quant Time: Dec 12 00:49:09 2025
 Quant Method : Z:\voasrv\HPCHEM1\MSVOA_V\Method\SFAMVSIM120325MA.M
 Quant Title : TRACE VOA SOM01.0
 QLast Update : Tue Dec 09 08:40:48 2025
 Response via : Initial Calibration

Compound	R.T.	QIon	Response	Conc	Units	Dev(Min)

Internal Standards						
1) 1,4-Difluorobenzene	5.566	114	5761	0.500	ug/L	0.00
5) Chlorobenzene-d5	8.800	117	5746	0.500	ug/L	0.02
12) 1,4-Dichlorobenzene-d4	11.197	152	2968	0.500	ug/L	0.02
System Monitoring Compounds						
2) Vinyl Chloride-d3	1.294	65	3125	0.553	ug/L	0.00
Spiked Amount	0.500	Range	40 - 130	Recovery	=	110.000%
4) 1,2-Dichloroethane-d4	4.993	65	1775	0.463	ug/L	0.02
Spiked Amount	0.500	Range	70 - 130	Recovery	=	92.000%
7) 1,2-Dichloropropane-d6	5.996	67	1989	0.410	ug/L	-0.02
Spiked Amount	0.500	Range	60 - 140	Recovery	=	82.000%
8) Toluene-d8	7.254	98	3724	0.455	ug/L	0.02
Spiked Amount	0.500	Range	70 - 130	Recovery	=	92.000%
10) 1,1,2,2-Tetrachloroeth...	10.133	84	1474	0.540	ug/L	0.00
Spiked Amount	0.500	Range	65 - 120	Recovery	=	108.000%

Target Compounds Qvalue

(#) = qualifier out of range (m) = manual integration (+) = signals summed

Data Path : Z:\voasrv\HPCHEM1\MSVOA_V\DATA\VV121125\
Data File : VV039541.D
Acq On : 11 Dec 2025 14:50
Operator : SY/MD
Sample : Q3828-09
Misc : 25.0 mL/MSVOA_V/WATER
ALS Vial : 9 Sample Multiplier: 1

Instrument :
MSVOA_V
ClientSampleId :
TB01-120925

Quant Time: Dec 12 00:49:09 2025
Quant Method : Z:\voasrv\HPCHEM1\MSVOA_V\Method\SFAMVSIM120325MA.M
Quant Title : TRACE VOA SOM01.0
QLast Update : Tue Dec 09 08:40:48 2025
Response via : Initial Calibration

