

Data Path : Z:\VOASRV\HPCHEM1\MSVOA V\DATA\VV121220\
 Data File : VV019550.D
 Acq On : 12 Dec 2020 12:45
 Operator : SY/MD
 Sample : L5056-04 200X
 Misc : 5.0mL/MSVOA V/WATER
 ALS Vial : 1 Sample Multiplier: 1

Instrument :
 MSVOA_V
 ClientSampleId :
 BG4N1

Quant Time: Dec 13 05:30:17 2020
 Quant Method : Z:\VOASRV\HPCHEM1\MSVOA_V\METHOD\SOMVLM121020WMA.M
 Quant Title : VOC Analysis
 QLast Update : Sun Dec 13 04:37:51 2020
 Response via : Initial Calibration

Internal Standards	R.T.	QIon	Response	Conc	Units	Dev(Min)
1) 1,4-Difluorobenzene	5.62	114	625967	50.00	ug/L	0.00
28) Chlorobenzene-d5	8.86	117	577010	50.00	ug/L	0.00
60) 1,4-Dichlorobenzene-d4	11.25	152	264969	50.00	ug/L	0.00

System Monitoring Compounds

4) Vinyl Chloride-d3	1.31	65	184659	41.14	ug/L	0.00
Spiked Amount	50.000	Range	60 - 135	Recovery	=	82.28%
7) Chloroethane-d5	1.57	69	160730	45.40	ug/L	0.00
Spiked Amount	50.000	Range	70 - 130	Recovery	=	90.80%
11) 1,1-Dichloroethene-d2	2.11	63	289557	34.76	ug/L	0.00
Spiked Amount	50.000	Range	60 - 125	Recovery	=	69.52%
21) 2-Butanone-d5	3.89	46	259582	106.40	ug/L	0.00
Spiked Amount	100.000	Range	40 - 130	Recovery	=	106.40%
24) Chloroform-d	4.35	84	376916	46.59	ug/L	0.00
Spiked Amount	50.000	Range	70 - 125	Recovery	=	93.18%
26) 1,2-Dichloroethane-d4	5.04	65	267264	50.55	ug/L	0.00
Spiked Amount	50.000	Range	70 - 125	Recovery	=	101.10%
32) Benzene-d6	5.05	84	747192	48.55	ug/L	0.00
Spiked Amount	50.000	Range	70 - 125	Recovery	=	97.10%
36) 1,2-Dichloropropane-d6	6.07	67	240270	49.84	ug/L	0.00
Spiked Amount	50.000	Range	70 - 120	Recovery	=	99.68%
41) Toluene-d8	7.32	98	669319	48.31	ug/L	0.00
Spiked Amount	50.000	Range	80 - 120	Recovery	=	96.62%
43) trans-1,3-Dichloropropene-	7.62	79	125286	45.69	ug/L	0.00
Spiked Amount	50.000	Range	60 - 125	Recovery	=	91.38%
47) 2-Hexanone-d5	8.09	63	194818	102.97	ug/L	0.00
Spiked Amount	100.000	Range	45 - 130	Recovery	=	102.97%
57) 1,1,2,2-Tetrachloroethane-	10.22	84	319269	48.96	ug/L	0.00
Spiked Amount	50.000	Range	65 - 120	Recovery	=	97.92%
64) 1,2-Dichlorobenzene-d4	11.63	152	245715	49.61	ug/L	0.00
Spiked Amount	50.000	Range	80 - 120	Recovery	=	99.22%


Target Compounds					Ovalue
34) Trichloroethene	5.92	95	13361	2.999 ug/L	96
46) Tetrachloroethene	7.98	164	1138372	369.307 ug/L	98


(#) = qualifier out of range (m) = manual integration (+) = signals summed

Data Path : Z:\V0ASRV\HPCHEM1\MSVOA V\DATA\VV121220\
Data File : VV019550.D
Acq On : 12 Dec 2020 12:45
Operator : SY/MD
Sample : L5056-04 200X
Misc : 5.0mL/MSVOA V/WATER
ALS Vial : 1 Sample Multiplier: 1

Instrument :
MSVOA_V
ClientSampleId :
BG4N1

Quant Time: Dec 13 05:30:17 2020
Quant Method : Z:\V0ASRV\HPCHEM1\MSVOA_V\METHOD\SOMVLM121020WMA.M
Quant Title : VOC Analysis
QLast Update : Sun Dec 13 04:37:51 2020
Response via : Initial Calibration





#34

Trichloroethene

Concen: 2.999 ug/L

RT: 5.92 min Scan# 1503

Delta R.T. 0.01 min

Lab File: VV019550.D

Acq: 12 Dec 2020 12:45

Instrument :

MSVOA_V

ClientSampled :

BG4N1

Tot Ion: 95 Resp: 13361



Ion Ratio Lower Upper

95 100

97 59.5 44.4 82.4

132 86.4 64.3 119.3

130 99.4 67.8 125.8





Abundance Ion 95.00 (94.70 to 95.70): VV0

Ion 97.00 (96.70 to 97.70): VV0

Ion 132.00 (131.70 to 132.70): V

Ion 130.00 (129.70 to 130.70): V



#46

Tetrachloroethene

Concen: 369.307 ug/L

RT: 7.98 min Scan# 2143

Delta R.T. 0.00 min

Lab File: VV019550.D

Acq: 12 Dec 2020 12:45

Tgt Ion:164 Resp: 1138372



Ion Ratio Lower Upper

164 100

129 101.9 70.6 131.0

131 99.2 68.0 126.4

166 127.7 86.9 161.5



Abundance Ion 164.00 (163.70 to 164.70): V

Ion 129.00 (128.70 to 129.70): V

Ion 131.00 (130.70 to 131.70): V

Ion 166.00 (165.70 to 166.70): V

