

Data Path : Z:\voasrv\HPCHEM1\MSVOA_V\Data\VW121223\
 Data File : VW033266.D
 Acq On : 12 Dec 2023 16:33
 Operator : SY/MD
 Sample : 05751-20
 Misc : 5.0mL/MSVOA_V/WATER
 ALS Vial : 15 Sample Multiplier: 1

Instrument :
 MSVOA_V
 ClientSampleId :
 VHBLK001

Quant Time: Dec 13 00:43:50 2023
 Quant Method : Z:\voasrv\HPCHEM1\MSVOA_V\Method\SFAMVLM120123WMA.M
 Quant Title : VOC Analysis
 QLast Update : Fri Dec 08 21:51:12 2023
 Response via : Initial Calibration

Compound	R.T.	QIon	Response	Conc	Units	Dev(Min)

Internal Standards						
1) 1,4-Difluorobenzene	5.532	114	272859	50.000	ug/L	0.00
28) Chlorobenzene-d5	8.786	117	276726	50.000	ug/L	0.00
58) 1,4-Dichlorobenzene-d4	11.185	152	147584	50.000	ug/L	0.00
System Monitoring Compounds						
4) Vinyl Chloride-d3	1.278	65	79570	48.782	ug/L	0.00
Spiked Amount	50.000	Range	60 - 135	Recovery	=	97.560%
7) Chloroethane-d5	1.532	69	58179	51.190	ug/L	0.00
Spiked Amount	50.000	Range	70 - 130	Recovery	=	102.380%
11) 1,1-Dichloroethene-d2	2.060	65	31035	50.014	ug/L	0.00
Spiked Amount	50.000	Range	60 - 125	Recovery	=	100.020%
21) 2-Butanone-d5	3.780	46	70227	86.683	ug/L	0.00
Spiked Amount	100.000	Range	40 - 130	Recovery	=	86.680%
24) Chloroform-d	4.249	84	170446	49.239	ug/L	0.00
Spiked Amount	50.000	Range	70 - 125	Recovery	=	98.480%
26) 1,2-Dichloroethane-d4	4.940	65	112429	52.239	ug/L	0.00
Spiked Amount	50.000	Range	70 - 125	Recovery	=	104.480%
32) Benzene-d6	4.960	84	328464	50.752	ug/L	0.00
Spiked Amount	50.000	Range	70 - 125	Recovery	=	101.500%
36) 1,2-Dichloropropane-d6	5.989	67	85100	48.903	ug/L	0.00
Spiked Amount	50.000	Range	70 - 120	Recovery	=	97.800%
41) Toluene-d8	7.243	98	311518	50.457	ug/L	0.00
Spiked Amount	50.000	Range	80 - 120	Recovery	=	100.920%
43) trans-1,3-Dichloroprop...	7.551	79	51810	47.084	ug/L	0.00
Spiked Amount	50.000	Range	60 - 125	Recovery	=	94.160%
47) 2-Hexanone-d5	8.024	63	59984	85.804	ug/L	0.00
Spiked Amount	100.000	Range	45 - 130	Recovery	=	85.800%
56) 1,1,2,2-Tetrachloroeth...	10.152	84	116318	47.022	ug/L	0.00
Spiked Amount	50.000	Range	65 - 120	Recovery	=	94.040%
66) 1,2-Dichlorobenzene-d4	11.561	152	144837	55.158	ug/L	0.00
Spiked Amount	50.000	Range	80 - 120	Recovery	=	110.320%

Target Compounds Qvalue

(#) = qualifier out of range (m) = manual integration (+) = signals summed

Data Path : Z:\voasrv\HPCHEM1\MSVOA_V\Data\VV121223\
Data File : VV033266.D
Acq On : 12 Dec 2023 16:33
Operator : SY/MD
Sample : 05751-20
Misc : 5.0mL/MSVOA_V/WATER
ALS Vial : 15 Sample Multiplier: 1

Instrument :
MSVOA_V
ClientSampleId :
VHBLK001

Quant Time: Dec 13 00:43:50 2023
Quant Method : Z:\voasrv\HPCHEM1\MSVOA_V\Method\SFAMVLM120123WMA.M
Quant Title : VOC Analysis
QLast Update : Fri Dec 08 21:51:12 2023
Response via : Initial Calibration

