

Data Path : Z:\VOASRV\HPCHEM1\MSVOA V\DATA\VV121318\
 Data File : VV008911.D
 Acq On : 13 Dec 2018 16:13
 Operator : SY/MD
 Sample : J6367-01
 Misc : 5.0 mL/MSVOA V/WATER
 ALS Vial : 11 Sample Multiplier: 1

Instrument :
 MSVOA_V
 ClientSampleId :
 COL36

Quant Time: Dec 20 05:10:24 2018
 Quant Method : Z:\VOASRV\HPCHEM1\MSVOA_V\METHOD\SOMVLM121318WMA.M
 Quant Title : VOC Analysis
 QLast Update : Fri Dec 14 01:09:42 2018
 Response via : Initial Calibration

Internal Standards	R.T.	QIon	Response	Conc	Units	Dev(Min)
1) 1,4-Difluorobenzene	5.67	114	253767	50.00	ug/L	0.00
28) Chlorobenzene-d5	8.90	117	234461	50.00	ug/L	0.00
60) 1,4-Dichlorobenzene-d4	11.30	152	100711	50.00	ug/L	0.00

System Monitoring Compounds

4) Vinyl Chloride-d3	1.32	65	65114	43.94	ug/L	0.00
Spiked Amount	50.000	Range	60 - 135	Recovery	=	87.88%
7) Chloroethane-d5	1.57	69	46473	49.83	ug/L	0.00
Spiked Amount	50.000	Range	70 - 130	Recovery	=	99.66%
11) 1,1-Dichloroethene-d2	2.13	63	120228	34.35	ug/L	0.00
Spiked Amount	50.000	Range	60 - 125	Recovery	=	68.70%
21) 2-Butanone-d5	3.93	46	96603	97.63	ug/L	0.00
Spiked Amount	100.000	Range	40 - 130	Recovery	=	97.63%
24) Chloroform-d	4.40	84	146774	46.42	ug/L	0.00
Spiked Amount	50.000	Range	70 - 125	Recovery	=	92.84%
26) 1,2-Dichloroethane-d4	5.08	65	96112	47.72	ug/L	0.00
Spiked Amount	50.000	Range	70 - 125	Recovery	=	95.44%
32) Benzene-d6	5.10	84	301027	47.64	ug/L	0.00
Spiked Amount	50.000	Range	70 - 125	Recovery	=	95.28%
36) 1,2-Dichloropropane-d6	6.12	67	88959	46.85	ug/L	0.00
Spiked Amount	50.000	Range	70 - 120	Recovery	=	93.70%
41) Toluene-d8	7.36	98	276567	46.63	ug/L	0.00
Spiked Amount	50.000	Range	80 - 120	Recovery	=	93.26%
43) trans-1,3-Dichloropropene-	7.67	79	44344	46.05	ug/L	0.00
Spiked Amount	50.000	Range	60 - 125	Recovery	=	92.10%
47) 2-Hexanone-d5	8.13	63	61606	87.67	ug/L	0.00
Spiked Amount	100.000	Range	45 - 130	Recovery	=	87.67%
57) 1,1,2,2-Tetrachloroethane-	10.26	84	111565	45.51	ug/L	0.00
Spiked Amount	50.000	Range	65 - 120	Recovery	=	91.02%
64) 1,2-Dichlorobenzene-d4	11.68	152	97548	46.96	ug/L	0.00
Spiked Amount	50.000	Range	80 - 120	Recovery	=	93.92%

Target Compounds	Ovalue

(#) = qualifier out of range (m) = manual integration (+) = signals summed

Data Path : Z:\V0ASRV\HPCHEM1\MSVOA V\DATA\VV121318\
Data File : VV008911.D
Acq On : 13 Dec 2018 16:13
Operator : SY/MD
Sample : J6367-01
Misc : 5.0 mL/MSVOA V/WATER
ALS Vial : 11 Sample Multiplier: 1

Instrument :
MSVOA_V
ClientSampleId :
COL36

Quant Time: Dec 20 05:10:24 2018
Quant Method : Z:\V0ASRV\HPCHEM1\MSVOA_V\METHOD\SOMVLM121318WMA.M
Quant Title : VOC Analysis
QLast Update : Fri Dec 14 01:09:42 2018
Response via : Initial Calibration

