

Data Path : Z:\VOASRV\HPCHEM1\MSVOA V\DATA\VV121318\
 Data File : VV008915.D
 Acq On : 13 Dec 2018 17:49
 Operator : SY/MD
 Sample : J6367-05
 Misc : 5.0 mL/MSVOA V/WATER
 ALS Vial : 15 Sample Multiplier: 1

Instrument :
 MSVOA_V
 ClientSampleId :
 COL41

Quant Time: Dec 20 05:12:59 2018
 Quant Method : Z:\VOASRV\HPCHEM1\MSVOA_V\METHOD\SOMVLM121318WMA.M
 Quant Title : VOC Analysis
 QLast Update : Fri Dec 14 01:09:42 2018
 Response via : Initial Calibration

Internal Standards	R.T.	QIon	Response	Conc	Units	Dev(Min)
1) 1,4-Difluorobenzene	5.67	114	210616	50.00	ug/L	0.00
28) Chlorobenzene-d5	8.90	117	193159	50.00	ug/L	0.00
60) 1,4-Dichlorobenzene-d4	11.30	152	80349	50.00	ug/L	0.00

System Monitoring Compounds

4) Vinyl Chloride-d3	1.32	65	57653	46.87	ug/L	0.00
Spiked Amount	50.000	Range	60 - 135	Recovery	=	93.74%
7) Chloroethane-d5	1.57	69	43170	55.77	ug/L	0.00
Spiked Amount	50.000	Range	70 - 130	Recovery	=	111.54%
11) 1,1-Dichloroethene-d2	2.13	63	109629	37.74	ug/L	0.00
Spiked Amount	50.000	Range	60 - 125	Recovery	=	75.48%
21) 2-Butanone-d5	3.93	46	84604	103.02	ug/L	0.00
Spiked Amount	100.000	Range	40 - 130	Recovery	=	103.02%
24) Chloroform-d	4.41	84	127129	48.44	ug/L	0.00
Spiked Amount	50.000	Range	70 - 125	Recovery	=	96.88%
26) 1,2-Dichloroethane-d4	5.08	65	85703	51.27	ug/L	0.00
Spiked Amount	50.000	Range	70 - 125	Recovery	=	102.54%
32) Benzene-d6	5.10	84	263767	50.67	ug/L	0.00
Spiked Amount	50.000	Range	70 - 125	Recovery	=	101.34%
36) 1,2-Dichloropropane-d6	6.12	67	79486	50.82	ug/L	0.00
Spiked Amount	50.000	Range	70 - 120	Recovery	=	101.64%
41) Toluene-d8	7.36	98	242644	49.66	ug/L	0.00
Spiked Amount	50.000	Range	80 - 120	Recovery	=	99.32%
43) trans-1,3-Dichloropropene-	7.67	79	38118	48.05	ug/L	0.00
Spiked Amount	50.000	Range	60 - 125	Recovery	=	96.10%
47) 2-Hexanone-d5	8.13	63	57268	98.92	ug/L	0.00
Spiked Amount	100.000	Range	45 - 130	Recovery	=	98.92%
57) 1,1,2,2-Tetrachloroethane-	10.26	84	97769	48.41	ug/L	0.00
Spiked Amount	50.000	Range	65 - 120	Recovery	=	96.82%
64) 1,2-Dichlorobenzene-d4	11.67	152	85822	51.79	ug/L	0.00
Spiked Amount	50.000	Range	80 - 120	Recovery	=	103.58%

Target Compounds					Ovalue
34) Trichloroethene	5.96	95	5143	3.346 ug/L	95

(#) = qualifier out of range (m) = manual integration (+) = signals summed

Data Path : Z:\V0ASRV\HPCHEM1\MSVOA V\DATA\VV121318\
Data File : VV008915.D
Acq On : 13 Dec 2018 17:49
Operator : SY/MD
Sample : J6367-05
Misc : 5.0 mL/MSVOA V/WATER
ALS Vial : 15 Sample Multiplier: 1

Instrument :
MSVOA_V
ClientSampleId :
COL41

Quant Time: Dec 20 05:12:59 2018
Quant Method : Z:\V0ASRV\HPCHEM1\MSVOA_V\METHOD\SOMVLM121318WMA.M
Quant Title : VOC Analysis
QLast Update : Fri Dec 14 01:09:42 2018
Response via : Initial Calibration

