

Data Path : Z:\VOASRV\HPCHEM1\MSVOA V\DATA\VV121318\
 Data File : VV008919.D
 Acq On : 13 Dec 2018 19:25
 Operator : SY/MD
 Sample : J6367-09
 Misc : 5.0 mL/MSVOA V/WATER
 ALS Vial : 19 Sample Multiplier: 1

Instrument :
 MSVOA_V
 ClientSampleId :
 COL45

Quant Time: Dec 20 05:13:40 2018
 Quant Method : Z:\VOASRV\HPCHEM1\MSVOA_V\METHOD\SOMVLM121318WMA.M
 Quant Title : VOC Analysis
 QLast Update : Fri Dec 14 01:09:42 2018
 Response via : Initial Calibration

Internal Standards	R.T.	QIon	Response	Conc	Units	Dev(Min)
1) 1,4-Difluorobenzene	5.67	114	202761	50.00	ug/L	0.00
28) Chlorobenzene-d5	8.90	117	184091	50.00	ug/L	0.00
60) 1,4-Dichlorobenzene-d4	11.30	152	76669	50.00	ug/L	0.00

System Monitoring Compounds

4) Vinyl Chloride-d3	1.32	65	55161	46.58	ug/L	0.00
Spiked Amount	50.000	Range	60 - 135	Recovery	=	93.16%
7) Chloroethane-d5	1.56	69	41138	55.20	ug/L	0.00
Spiked Amount	50.000	Range	70 - 130	Recovery	=	110.40%
11) 1,1-Dichloroethene-d2	2.13	63	106872	38.22	ug/L	0.00
Spiked Amount	50.000	Range	60 - 125	Recovery	=	76.44%
21) 2-Butanone-d5	3.93	46	86908	109.93	ug/L	0.00
Spiked Amount	100.000	Range	40 - 130	Recovery	=	109.93%
24) Chloroform-d	4.40	84	123886	49.03	ug/L	0.00
Spiked Amount	50.000	Range	70 - 125	Recovery	=	98.06%
26) 1,2-Dichloroethane-d4	5.08	65	81323	50.53	ug/L	0.00
Spiked Amount	50.000	Range	70 - 125	Recovery	=	101.06%
32) Benzene-d6	5.10	84	255271	51.46	ug/L	0.00
Spiked Amount	50.000	Range	70 - 125	Recovery	=	102.92%
36) 1,2-Dichloropropane-d6	6.12	67	76498	51.32	ug/L	0.00
Spiked Amount	50.000	Range	70 - 120	Recovery	=	102.64%
41) Toluene-d8	7.36	98	233398	50.12	ug/L	0.00
Spiked Amount	50.000	Range	80 - 120	Recovery	=	100.24%
43) trans-1,3-Dichloropropene-	7.67	79	36583	48.39	ug/L	0.00
Spiked Amount	50.000	Range	60 - 125	Recovery	=	96.78%
47) 2-Hexanone-d5	8.13	63	60216	109.14	ug/L	0.00
Spiked Amount	100.000	Range	45 - 130	Recovery	=	109.14%
57) 1,1,2,2-Tetrachloroethane-	10.26	84	96998	50.40	ug/L	0.00
Spiked Amount	50.000	Range	65 - 120	Recovery	=	100.80%
64) 1,2-Dichlorobenzene-d4	11.68	152	82731	52.32	ug/L	0.00
Spiked Amount	50.000	Range	80 - 120	Recovery	=	104.64%

Target Compounds	Ovalue

(#) = qualifier out of range (m) = manual integration (+) = signals summed

Data Path : Z:\V0ASRV\HPCHEM1\MSVOA V\DATA\VV121318\
Data File : VV008919.D
Acq On : 13 Dec 2018 19:25
Operator : SY/MD
Sample : J6367-09
Misc : 5.0 mL/MSVOA V/WATER
ALS Vial : 19 Sample Multiplier: 1

Instrument :
MSVOA_V
ClientSampleId :
COL45

Quant Time: Dec 20 05:13:40 2018
Quant Method : Z:\V0ASRV\HPCHEM1\MSVOA_V\METHOD\SOMVLM121318WMA.M
Quant Title : VOC Analysis
QLast Update : Fri Dec 14 01:09:42 2018
Response via : Initial Calibration

