

Data Path : Z:\VOASRV\HPCHEM1\MSVOA V\DATA\VV121318\
 Data File : VV008910.D
 Acq On : 13 Dec 2018 15:49
 Operator : SY/MD
 Sample : VV1213WBL01
 Misc : 5.0 mL/MSVOA V/WATER
 ALS Vial : 10 Sample Multiplier: 1

Instrument :
 MSVOA_V
 ClientSampleId :
 VBLK65

Quant Time: Dec 14 01:20:10 2018
 Quant Method : Z:\VOASRV\HPCHEM1\MSVOA_V\METHOD\SOMVLM121318WMA.M
 Quant Title : VOC Analysis
 QLast Update : Fri Dec 14 01:09:42 2018
 Response via : Initial Calibration

Internal Standards	R.T.	QIon	Response	Conc	Units	Dev(Min)
1) 1,4-Difluorobenzene	5.66	114	215962	50.00	ug/L	0.00
28) Chlorobenzene-d5	8.90	117	198103	50.00	ug/L	0.00
60) 1,4-Dichlorobenzene-d4	11.30	152	81227	50.00	ug/L	0.00

System Monitoring Compounds

4) Vinyl Chloride-d3	1.32	65	61790	48.99	ug/L	0.00
Spiked Amount	50.000	Range	60 - 135	Recovery	=	97.98%
7) Chloroethane-d5	1.57	69	45179	56.92	ug/L	0.00
Spiked Amount	50.000	Range	70 - 130	Recovery	=	113.84%
11) 1,1-Dichloroethene-d2	2.13	63	115644	38.83	ug/L	0.00
Spiked Amount	50.000	Range	60 - 125	Recovery	=	77.66%
21) 2-Butanone-d5	3.93	46	87421	103.82	ug/L	0.00
Spiked Amount	100.000	Range	40 - 130	Recovery	=	103.82%
24) Chloroform-d	4.40	84	130588	48.53	ug/L	0.00
Spiked Amount	50.000	Range	70 - 125	Recovery	=	97.06%
26) 1,2-Dichloroethane-d4	5.08	65	87268	50.91	ug/L	0.00
Spiked Amount	50.000	Range	70 - 125	Recovery	=	101.82%
32) Benzene-d6	5.10	84	274132	51.35	ug/L	0.00
Spiked Amount	50.000	Range	70 - 125	Recovery	=	102.70%
36) 1,2-Dichloropropane-d6	6.12	67	81307	50.68	ug/L	0.00
Spiked Amount	50.000	Range	70 - 120	Recovery	=	101.36%
41) Toluene-d8	7.36	98	253219	50.53	ug/L	0.00
Spiked Amount	50.000	Range	80 - 120	Recovery	=	101.06%
43) trans-1,3-Dichloropropene-	7.66	79	40172	49.38	ug/L	0.00
Spiked Amount	50.000	Range	60 - 125	Recovery	=	98.76%
47) 2-Hexanone-d5	8.13	63	53164	89.54	ug/L	0.00
Spiked Amount	100.000	Range	45 - 130	Recovery	=	89.54%
57) 1,1,2,2-Tetrachloroethane-	10.26	84	98442	47.53	ug/L	0.00
Spiked Amount	50.000	Range	65 - 120	Recovery	=	95.06%
64) 1,2-Dichlorobenzene-d4	11.68	152	85457	51.01	ug/L	0.00
Spiked Amount	50.000	Range	80 - 120	Recovery	=	102.02%

Target Compounds	Ovalue

(#) = qualifier out of range (m) = manual integration (+) = signals summed

Data Path : Z:\V0ASRV\HPCHEM1\MSVOA V\DATA\VV121318\
Data File : VV008910.D
Acq On : 13 Dec 2018 15:49
Operator : SY/MD
Sample : VV1213WBL01
Misc : 5.0 mL/MSVOA V/WATER
ALS Vial : 10 Sample Multiplier: 1

Instrument :
MSVOA_V
ClientSampleId :
VBLK65

Quant Time: Dec 14 01:20:10 2018
Quant Method : Z:\V0ASRV\HPCHEM1\MSVOA_V\METHOD\SOMVLM121318WMA.M
Quant Title : VOC Analysis
QLast Update : Fri Dec 14 01:09:42 2018
Response via : Initial Calibration

