

Data Path : Z:\VOASRV\HPCHEM1\MSVOA V\DATA\VV121319\
 Data File : VV014048.D
 Acq On : 13 Dec 2019 13:01
 Operator : SY/MD
 Sample : K6200-03
 Misc : 25.0mL/MSVOA V/WATER
 ALS Vial : 4 Sample Multiplier: 1

Instrument :
 MSVOA_V
 ClientSampleId :
 BFF41

Quant Time: Dec 14 03:42:14 2019
 Quant Method : Z:\VOASRV\HPCHEM1\MSVOA_V\METHOD\SOMVTR112219WMA.M
 Quant Title : TRACE VOA SOM01.0
 QLast Update : Sat Dec 14 03:25:09 2019
 Response via : Initial Calibration

Internal Standards	R.T.	QIon	Response	Conc	Units	Dev(Min)
1) 1,4-Difluorobenzene	5.66	114	362138	5.00	ug/L	0.00
28) Chlorobenzene-d5	8.89	117	356969	5.00	ug/L	0.00
60) 1,4-Dichlorobenzene-d4	11.29	152	140978	5.00	ug/L	0.00

System Monitoring Compounds

4) Vinyl Chloride-d3	1.32	65	109394	4.88	ug/L	0.00
Spiked Amount	5.000	Range	40 - 130	Recovery	=	97.60%
7) Chloroethane-d5	1.58	69	100145	5.26	ug/L	0.00
Spiked Amount	5.000	Range	65 - 130	Recovery	=	105.20%
11) 1,1-Dichloroethene-d2	2.13	63	139037	3.68	ug/L	0.00
Spiked Amount	5.000	Range	60 - 125	Recovery	=	73.60%
20) 2-Butanone-d5	3.95	46	240980	51.72	ug/L	0.00
Spiked Amount	50.000	Range	40 - 130	Recovery	=	103.44%
24) Chloroform-d	4.40	84	202697	4.49	ug/L	0.00
Spiked Amount	5.000	Range	70 - 125	Recovery	=	89.80%
26) 1,2-Dichloroethane-d4	5.08	65	102856	4.78	ug/L	0.00
Spiked Amount	5.000	Range	70 - 130	Recovery	=	95.60%
32) Benzene-d6	5.10	84	379442	4.12	ug/L	0.00
Spiked Amount	5.000	Range	70 - 125	Recovery	=	82.40%
36) 1,2-Dichloropropane-d6	6.12	67	125854	4.59	ug/L	0.00
Spiked Amount	5.000	Range	60 - 140	Recovery	=	91.80%
41) Toluene-d8	7.36	98	279162	3.23	ug/L	0.00
Spiked Amount	5.000	Range	70 - 130	Recovery	=	64.60%#
43) trans-1,3-Dichloropropene-	7.66	79	39304	3.68	ug/L	0.00
Spiked Amount	5.000	Range	55 - 130	Recovery	=	73.60%
46) 2-Hexanone-d5	8.14	63	130728	33.18	ug/L	0.00
Spiked Amount	50.000	Range	45 - 130	Recovery	=	66.36%
57) 1,1,2,2-Tetrachloroethane-	10.26	84	89805	4.72	ug/L	0.00
Spiked Amount	5.000	Range	65 - 120	Recovery	=	94.40%
64) 1,2-Dichlorobenzene-d4	11.67	152	119692	4.81	ug/L	0.00
Spiked Amount	5.000	Range	80 - 120	Recovery	=	96.20%

Target Compounds	Ovalue

(#) = qualifier out of range (m) = manual integration (+) = signals summed

Data Path : Z:\V0ASRV\HPCHEM1\MSVOA V\DATA\VV121319\
Data File : VV014048.D
Acq On : 13 Dec 2019 13:01
Operator : SY/MD
Sample : K6200-03
Misc : 25.0mL/MSVOA V/WATER
ALS Vial : 4 Sample Multiplier: 1

Instrument :
MSVOA_V
ClientSampleId :
BFF41

Quant Time: Dec 14 03:42:14 2019
Quant Method : Z:\V0ASRV\HPCHEM1\MSVOA_V\METHOD\SOMVTR112219WMA.M
Quant Title : TRACE VOA SOM01.0
QLast Update : Sat Dec 14 03:25:09 2019
Response via : Initial Calibration

