

Data Path : Z:\VOASRV\HPCHEM1\MSVOA V\DATA\VV121319\
 Data File : VV014053.D
 Acq On : 13 Dec 2019 15:02
 Operator : SY/MD
 Sample : K6275-01
 Misc : 25.0mL/MSVOA V/WATER
 ALS Vial : 9 Sample Multiplier: 1

Instrument :
 MSVOA_V
 ClientSampleId :
 BFDR8

Quant Time: Dec 14 04:01:45 2019
 Quant Method : Z:\VOASRV\HPCHEM1\MSVOA_V\METHOD\SOMVTR112219WMA.M
 Quant Title : TRACE VOA SOM01.0
 QLast Update : Sat Dec 14 03:25:09 2019
 Response via : Initial Calibration

Internal Standards	R.T.	QIon	Response	Conc	Units	Dev(Min)
1) 1,4-Difluorobenzene	5.66	114	381659	5.00	ug/L	0.00
28) Chlorobenzene-d5	8.89	117	377885	5.00	ug/L	0.00
60) 1,4-Dichlorobenzene-d4	11.29	152	167885	5.00	ug/L	0.00

System Monitoring Compounds

4) Vinyl Chloride-d3	1.32	65	105659	4.47	ug/L	0.00
Spiked Amount	5.000	Range	40 - 130	Recovery	=	89.40%
7) Chloroethane-d5	1.58	69	97044	4.84	ug/L	0.00
Spiked Amount	5.000	Range	65 - 130	Recovery	=	96.80%
11) 1,1-Dichloroethene-d2	2.13	63	130658	3.29	ug/L	0.00
Spiked Amount	5.000	Range	60 - 125	Recovery	=	65.80%
20) 2-Butanone-d5	3.93	46	271185	55.22	ug/L	-0.03
Spiked Amount	50.000	Range	40 - 130	Recovery	=	110.44%
24) Chloroform-d	4.40	84	191889	4.03	ug/L	0.00
Spiked Amount	5.000	Range	70 - 125	Recovery	=	80.60%
26) 1,2-Dichloroethane-d4	5.08	65	103530	4.57	ug/L	0.00
Spiked Amount	5.000	Range	70 - 130	Recovery	=	91.40%
32) Benzene-d6	5.10	84	363666	3.73	ug/L	0.00
Spiked Amount	5.000	Range	70 - 125	Recovery	=	74.60%
36) 1,2-Dichloropropane-d6	6.11	67	123193	4.24	ug/L	0.00
Spiked Amount	5.000	Range	60 - 140	Recovery	=	84.80%
41) Toluene-d8	7.36	98	266441	2.91	ug/L	0.00
Spiked Amount	5.000	Range	70 - 130	Recovery	=	58.20%#
43) trans-1,3-Dichloropropene-	7.66	79	40796	3.61	ug/L	0.00
Spiked Amount	5.000	Range	55 - 130	Recovery	=	72.20%
46) 2-Hexanone-d5	8.13	63	151493	36.32	ug/L	0.00
Spiked Amount	50.000	Range	45 - 130	Recovery	=	72.64%
57) 1,1,2,2-Tetrachloroethane-	10.26	84	95159	4.72	ug/L	0.00
Spiked Amount	5.000	Range	65 - 120	Recovery	=	94.40%
64) 1,2-Dichlorobenzene-d4	11.67	152	129993	4.39	ug/L	0.00
Spiked Amount	5.000	Range	80 - 120	Recovery	=	87.80%

Target Compounds

					Ovalue
13) Acetone	2.19	43	8563	2.494 ug/L #	39
14) Carbon disulfide	2.32	76	35354	0.835 ug/L	97
25) Chloroform	4.43	83	14535	0.331 ug/L	98
42) Toluene	7.43	91	20122	0.202 ug/L	96

(#) = qualifier out of range (m) = manual integration (+) = signals summed

Data Path : Z:\V0ASRV\HPCHEM1\MSVOA V\DATA\VV121319\
Data File : VV014053.D
Acq On : 13 Dec 2019 15:02
Operator : SY/MD
Sample : K6275-01
Misc : 25.0mL/MSVOA V/WATER
ALS Vial : 9 Sample Multiplier: 1

Instrument :
MSVOA_V
ClientSampleId :
BFDR8

Quant Time: Dec 14 04:01:45 2019
Quant Method : Z:\V0ASRV\HPCHEM1\MSVOA_V\METHOD\SOMVTR112219WMA.M
Quant Title : TRACE VOA SOM01.0
QLast Update : Sat Dec 14 03:25:09 2019
Response via : Initial Calibration

