

Data Path : Z:\voasrv\HPCHEM1\MSVOA\_V\Data\VW121321\  
 Data File : VV023946.D  
 Acq On : 13 Dec 2021 13:57  
 Operator : SY/MD  
 Sample : M4971-16  
 Misc : 25.0mL/MSVOA\_V/WATER  
 ALS Vial : 9 Sample Multiplier: 1

Instrument :  
 MSVOA\_V  
 ClientSampleId :

Quant Time: Dec 14 01:08:03 2021  
 Quant Method : Z:\voasrv\HPCHEM1\MSVOA\_V\Method\SFAMVSIM112221.M  
 Quant Title : TRACE VOA SOM01.0  
 QLast Update : Tue Nov 23 02:04:29 2021  
 Response via : Initial Calibration

| Compound                      | R.T.   | QIon  | Response | Conc     | Units | Dev(Min) |
|-------------------------------|--------|-------|----------|----------|-------|----------|
| -----                         |        |       |          |          |       |          |
| Internal Standards            |        |       |          |          |       |          |
| 1) 1,4-Difluorobenzene        | 5.639  | 114   | 8521     | 0.500    | ug/L  | 0.00     |
| 5) Chlorobenzene-d5           | 8.863  | 117   | 9262     | 0.500    | ug/L  | 0.00     |
| 11) 1,4-Dichlorobenzene-d4    | 11.244 | 152   | 6623     | 0.500    | ug/L  | -0.02    |
| System Monitoring Compounds   |        |       |          |          |       |          |
| 2) Vinyl Chloride-d3          | 1.310  | 65    | 4569     | 0.528    | ug/L  | 0.00     |
| Spiked Amount                 | 0.500  | Range | 40 - 130 | Recovery | =     | 106.000% |
| 4) 1,2-Dichloroethane-d4      | 5.042  | 65    | 3039     | 0.517    | ug/L  | 0.00     |
| Spiked Amount                 | 0.500  | Range | 70 - 130 | Recovery | =     | 104.000% |
| 7) 1,2-Dichloropropane-d6     | 6.068  | 67    | 3203     | 0.398    | ug/L  | 0.00     |
| Spiked Amount                 | 0.500  | Range | 60 - 140 | Recovery | =     | 80.000%  |
| 8) Toluene-d8                 | 7.317  | 98    | 10891    | 0.523    | ug/L  | 0.00     |
| Spiked Amount                 | 0.500  | Range | 70 - 130 | Recovery | =     | 104.000% |
| 10) 1,1,2,2-Tetrachloroeth... | 10.216 | 84    | 2268     | 0.464    | ug/L  | 0.00     |
| Spiked Amount                 | 0.500  | Range | 65 - 120 | Recovery | =     | 92.000%  |

Target Compounds Qvalue

-----  
 (#) = qualifier out of range (m) = manual integration (+) = signals summed

Data Path : Z:\voasrv\HPCHEM1\MSVOA\_V\Data\VV121321\  
Data File : VV023946.D  
Acq On : 13 Dec 2021 13:57  
Operator : SY/MD  
Sample : M4971-16  
Misc : 25.0mL/MSVOA\_V/WATER  
ALS Vial : 9 Sample Multiplier: 1

Instrument :  
MSVOA\_V  
ClientSampleId :

Quant Time: Dec 14 01:08:03 2021  
Quant Method : Z:\voasrv\HPCHEM1\MSVOA\_V\Method\SFAMVSIM112221.M  
Quant Title : TRACE VOA SOM01.0  
QLast Update : Tue Nov 23 02:04:29 2021  
Response via : Initial Calibration

