

Data Path : Z:\voasrv\HPCHEM1\MSVOA_V\Data\VW121321\
 Data File : VV023954.D
 Acq On : 13 Dec 2021 17:19
 Operator : SY/MD
 Sample : M4971-05
 Misc : 25.0mL/MSVOA_V/WATER
 ALS Vial : 17 Sample Multiplier: 1

Instrument :
 MSVOA_V
ClientSampleId :
 VHBLK001

Quant Time: Dec 14 01:08:38 2021
 Quant Method : Z:\voasrv\HPCHEM1\MSVOA_V\Method\SFAMVSIM112221.M
 Quant Title : TRACE VOA SOM01.0
 QLast Update : Tue Nov 23 02:04:29 2021
 Response via : Initial Calibration

| Compound | R.T. | QIon | Response | Conc | Units | Dev(Min) |
|-------------------------------|--------|----------------|----------|-------|----------|----------|
| Internal Standards | | | | | | |
| 1) 1,4-Difluorobenzene | 5.639 | 114 | 7538 | 0.500 | ug/L | 0.00 |
| 5) Chlorobenzene-d5 | 8.863 | 117 | 7386 | 0.500 | ug/L | 0.00 |
| 11) 1,4-Dichlorobenzene-d4 | 11.261 | 152 | 4107 | 0.500 | ug/L | 0.00 |
| System Monitoring Compounds | | | | | | |
| 2) Vinyl Chloride-d3 | 1.303 | 65 | 4506 | 0.589 | ug/L | 0.00 |
| Spiked Amount | 0.500 | Range 40 - 130 | Recovery | = | 118.000% | |
| 4) 1,2-Dichloroethane-d4 | 5.042 | 65 | 2848 | 0.548 | ug/L | 0.00 |
| Spiked Amount | 0.500 | Range 70 - 130 | Recovery | = | 110.000% | |
| 7) 1,2-Dichloropropane-d6 | 6.068 | 67 | 3137 | 0.489 | ug/L | 0.00 |
| Spiked Amount | 0.500 | Range 60 - 140 | Recovery | = | 98.000% | |
| 8) Toluene-d8 | 7.317 | 98 | 9690 | 0.584 | ug/L | 0.00 |
| Spiked Amount | 0.500 | Range 70 - 130 | Recovery | = | 116.000% | |
| 10) 1,1,2,2-Tetrachloroeth... | 10.216 | 84 | 2103 | 0.540 | ug/L | 0.00 |
| Spiked Amount | 0.500 | Range 65 - 120 | Recovery | = | 108.000% | |

Target Compounds Qvalue

(#) = qualifier out of range (m) = manual integration (+) = signals summed

Data Path : Z:\voasrv\HPCHEM1\MSVOA_V\Data\VV121321\
 Data File : VV023954.D
 Acq On : 13 Dec 2021 17:19
 Operator : SY/MD
 Sample : M4971-05
 Misc : 25.0mL/MSVOA_V/WATER
 ALS Vial : 17 Sample Multiplier: 1

Instrument :
 MSVOA_V
 ClientSampleId :
 VHBLK001

Quant Time: Dec 14 01:08:38 2021
 Quant Method : Z:\voasrv\HPCHEM1\MSVOA_V\Method\SFAMVSIM112221.M
 Quant Title : TRACE VOA SOM01.0
 QLast Update : Tue Nov 23 02:04:29 2021
 Response via : Initial Calibration

