

Data Path : Z:\voasrv\HPCHEM1\MSVOA_V\Data\VW121322\
 Data File : VW029645.D
 Acq On : 13 Dec 2022 12:52
 Operator : SY/MD
 Sample : N5928-06DL 10X
 Misc : 5.0mL/MSVOA_V/WATER
 ALS Vial : 7 Sample Multiplier: 1

Instrument :
 MSVOA_V
 ClientSampleId :
 EXZ69DL

Quant Time: Dec 13 22:51:17 2022
 Quant Method : Z:\voasrv\HPCHEM1\MSVOA_V\Method\SFAMVLM120722WMA.M
 Quant Title : VOC Analysis
 QLast Update : Tue Dec 13 22:49:59 2022
 Response via : Initial Calibration

Compound	R.T.	QIon	Response	Conc	Units	Dev(Min)

Internal Standards						
1) 1,4-Difluorobenzene	5.609	114	498200	50.000	ug/L	0.00
28) Chlorobenzene-d5	8.844	117	454637	50.000	ug/L	0.00
58) 1,4-Dichlorobenzene-d4	11.239	152	242222	50.000	ug/L	0.00
System Monitoring Compounds						
4) Vinyl Chloride-d3	1.304	65	142289	46.640	ug/L	0.00
Spiked Amount	50.000	Range	60 - 135	Recovery	=	93.280%
7) Chloroethane-d5	1.561	69	113991	49.001	ug/L	0.00
Spiked Amount	50.000	Range	70 - 130	Recovery	=	98.000%
11) 1,1-Dichloroethene-d2	2.104	63	177543	37.312	ug/L	0.00
Spiked Amount	50.000	Range	60 - 125	Recovery	=	74.620%
21) 2-Butanone-d5	3.873	46	166680	89.562	ug/L	0.00
Spiked Amount	100.000	Range	40 - 130	Recovery	=	89.560%
24) Chloroform-d	4.339	84	244122	42.523	ug/L	0.00
Spiked Amount	50.000	Range	70 - 125	Recovery	=	85.040%
26) 1,2-Dichloroethane-d4	5.024	65	145463	44.364	ug/L	0.00
Spiked Amount	50.000	Range	70 - 125	Recovery	=	88.720%
32) Benzene-d6	5.043	84	525844	43.773	ug/L	0.00
Spiked Amount	50.000	Range	70 - 125	Recovery	=	87.540%
36) 1,2-Dichloropropane-d6	6.063	67	160729	43.802	ug/L	0.00
Spiked Amount	50.000	Range	70 - 120	Recovery	=	87.600%
41) Toluene-d8	7.307	98	476881	42.131	ug/L	0.00
Spiked Amount	50.000	Range	80 - 120	Recovery	=	84.260%
43) trans-1,3-Dichloroprop...	7.612	79	67757	36.568	ug/L	0.00
Spiked Amount	50.000	Range	60 - 125	Recovery	=	73.140%
47) 2-Hexanone-d5	8.082	63	138634	88.475	ug/L	0.00
Spiked Amount	100.000	Range	45 - 130	Recovery	=	88.470%
56) 1,1,2,2-Tetrachloroeth...	10.207	84	220880	42.124	ug/L	0.00
Spiked Amount	50.000	Range	65 - 120	Recovery	=	84.240%
66) 1,2-Dichlorobenzene-d4	11.612	152	194584	42.200	ug/L	0.00
Spiked Amount	50.000	Range	80 - 120	Recovery	=	84.400%
Target Compounds						
					Qvalue	
5) Vinyl chloride	1.310	62	22933	6.126	ug/L	89
8) Chloroethane	1.580	64	1925	0.924	ug/L	94
12) 1,1-Dichloroethene	2.114	96	3280	1.232	ug/L #	1
19) 1,1-Dichloroethane	3.185	63	79975	15.214	ug/L	99
20) cis-1,2-Dichloroethene	3.902	96	146812	43.512	ug/L	94
30) 1,1,1-Trichloroethane	4.600	97	60865	12.288	ug/L	99
46) Tetrachloroethene	7.969	164	4616	1.788	ug/L	92

(#) = qualifier out of range (m) = manual integration (+) = signals summed

Data Path : Z:\voasrv\HPCHEM1\MSVOA_V\Data\VV121322\
Data File : VV029645.D
Acq On : 13 Dec 2022 12:52
Operator : SY/MD
Sample : N5928-06DL 10X
Misc : 5.0mL/MSVOA_V/WATER
ALS Vial : 7 Sample Multiplier: 1

Instrument :
MSVOA_V
ClientSampleId :
EXZ69DL

Quant Time: Dec 13 22:51:17 2022
Quant Method : Z:\voasrv\HPCHEM1\MSVOA_V\Method\SFAMVLM120722WMA.M
Quant Title : VOC Analysis
QLast Update : Tue Dec 13 22:49:59 2022
Response via : Initial Calibration

