

Data Path : Z:\voasrv\HPCHEM1\MSVOA_V\Data\VW121322\
 Data File : VW029661.D
 Acq On : 13 Dec 2022 19:09
 Operator : SY/MD
 Sample : N5980-03
 Misc : 5.0mL/MSVOA_V/WATER
 ALS Vial : 23 Sample Multiplier: 1

Instrument :
 MSVOA_V
 ClientSampleId :
 C0EP6

Quant Time: Dec 13 22:54:26 2022
 Quant Method : Z:\voasrv\HPCHEM1\MSVOA_V\Method\SFAMVLM120722WMA.M
 Quant Title : VOC Analysis
 QLast Update : Tue Dec 13 22:49:59 2022
 Response via : Initial Calibration

Compound	R.T.	QIon	Response	Conc	Units	Dev(Min)


Internal Standards						
1) 1,4-Difluorobenzene	5.609	114	420226	50.000	ug/L	0.00
28) Chlorobenzene-d5	8.844	117	385729	50.000	ug/L	0.00
58) 1,4-Dichlorobenzene-d4	11.239	152	207445	50.000	ug/L	0.00
System Monitoring Compounds						
4) Vinyl Chloride-d3	1.304	65	118812	46.171	ug/L	0.00
Spiked Amount 50.000	Range 60 - 135		Recovery =	92.340%		
7) Chloroethane-d5	1.564	69	98552	50.225	ug/L	0.00
Spiked Amount 50.000	Range 70 - 130		Recovery =	100.460%		
11) 1,1-Dichloroethene-d2	2.105	63	154709	38.546	ug/L	0.00
Spiked Amount 50.000	Range 60 - 125		Recovery =	77.100%		
21) 2-Butanone-d5	3.873	46	155539	99.084	ug/L	0.00
Spiked Amount 100.000	Range 40 - 130		Recovery =	99.080%		
24) Chloroform-d	4.339	84	220096	45.451	ug/L	0.00
Spiked Amount 50.000	Range 70 - 125		Recovery =	90.900%		
26) 1,2-Dichloroethane-d4	5.024	65	131342	47.490	ug/L	0.00
Spiked Amount 50.000	Range 70 - 125		Recovery =	94.980%		
32) Benzene-d6	5.043	84	471136	46.225	ug/L	0.00
Spiked Amount 50.000	Range 70 - 125		Recovery =	92.440%		
36) 1,2-Dichloropropane-d6	6.063	67	145638	46.780	ug/L	0.00
Spiked Amount 50.000	Range 70 - 120		Recovery =	93.560%		
41) Toluene-d8	7.307	98	418202	43.547	ug/L	0.00
Spiked Amount 50.000	Range 80 - 120		Recovery =	87.100%		
43) trans-1,3-Dichloroprop...	7.612	79	60209	38.299	ug/L	0.00
Spiked Amount 50.000	Range 60 - 125		Recovery =	76.600%		
47) 2-Hexanone-d5	8.085	63	129126	97.128	ug/L	0.00
Spiked Amount 100.000	Range 45 - 130		Recovery =	97.130%		
56) 1,1,2,2-Tetrachloroeth...	10.207	84	206249	46.360	ug/L	0.00
Spiked Amount 50.000	Range 65 - 120		Recovery =	92.720%		
66) 1,2-Dichlorobenzene-d4	11.612	152	174913	44.294	ug/L	0.00
Spiked Amount 50.000	Range 80 - 120		Recovery =	88.580%		
Target Compounds					Qvalue	
13) Acetone	2.162	43	3023	2.956	ug/L	94


(#) = qualifier out of range (m) = manual integration (+) = signals summed

Data Path : Z:\voasrv\HPCHEM1\MSVOA_V\Data\VV121322\
Data File : VV029661.D
Acq On : 13 Dec 2022 19:09
Operator : SY/MD
Sample : N5980-03
Misc : 5.0mL/MSVOA_V/WATER
ALS Vial : 23 Sample Multiplier: 1

Instrument :
MSVOA_V
ClientSampleId :
C0EP6

Quant Time: Dec 13 22:54:26 2022
Quant Method : Z:\voasrv\HPCHEM1\MSVOA_V\Method\SFAMVLM120722WMA.M
Quant Title : VOC Analysis
QLast Update : Tue Dec 13 22:49:59 2022
Response via : Initial Calibration





#13
 Acetone
 Concen: 2.956 ug/L
 RT: 2.162 min Scan# 34
 Delta R.T. -0.000 min
 Lab File: VV029661.D
 Acq: 13 Dec 2022 19:09

Instrument :
 MSVOA_V
 ClientSampleId :
 C0EP6

Tgt Ion: 43 Resp: 3023
 Ion Ratio Lower Upper
 43 100
 58 41.1 0.0 74.8

