

Data Path : Z:\voasrv\HPCHEM1\MSVOA_V\Data\VW121323\
 Data File : VW033309.D
 Acq On : 14 Dec 2023 04:43
 Operator : SY/MD
 Sample : 05770-09
 Misc : 25.0mL/MSVOA_V/WATER
 ALS Vial : 44 Sample Multiplier: 1

Instrument :
 MSVOA_V
 ClientSampleId :
 E0Y81

Quant Time: Dec 14 05:46:19 2023
 Quant Method : Z:\voasrv\HPCHEM1\MSVOA_V\Method\SFAMVTR121323WMA.M
 Quant Title : TRACE VOA SFAM1.0
 QLast Update : Thu Dec 14 01:07:25 2023
 Response via : Initial Calibration

Compound		R.T.	QIon	Response	Conc	Units	Dev(Min)
Internal Standards							
1) 1,4-Difluorobenzene		5.535	114	122656	5.000	ug/L	0.00
28) Chlorobenzene-d5		8.786	117	121668	5.000	ug/L	0.00
58) 1,4-Dichlorobenzene-d4		11.188	152	63638	5.000	ug/L	0.00
System Monitoring Compounds							
4) Vinyl Chloride-d3		1.278	65	40854	5.196	ug/L	0.00
Spiked Amount	5.000	Range	40 - 130	Recovery	=	104.000%	
7) Chloroethane-d5		1.532	69	31468	5.381	ug/L	0.00
Spiked Amount	5.000	Range	65 - 130	Recovery	=	107.600%	
11) 1,1-Dichloroethene-d2		2.060	65	16072	5.136	ug/L	0.00
Spiked Amount	5.000	Range	60 - 125	Recovery	=	102.800%	
20) 2-Butanone-d5		3.809	46	52695	56.631	ug/L	0.01
Spiked Amount	50.000	Range	40 - 130	Recovery	=	113.260%	
24) Chloroform-d		4.249	84	90650	5.527	ug/L	0.00
Spiked Amount	5.000	Range	70 - 125	Recovery	=	110.600%	
26) 1,2-Dichloroethane-d4		4.944	65	41169	5.777	ug/L	0.00
Spiked Amount	5.000	Range	70 - 130	Recovery	=	115.600%	
32) Benzene-d6		4.960	84	172365	5.477	ug/L	0.00
Spiked Amount	5.000	Range	70 - 125	Recovery	=	109.600%	
36) 1,2-Dichloropropane-d6		5.989	67	40314	5.627	ug/L	0.00
Spiked Amount	5.000	Range	60 - 140	Recovery	=	112.600%	
41) Toluene-d8		7.243	98	161120	5.253	ug/L	0.00
Spiked Amount	5.000	Range	70 - 130	Recovery	=	105.000%	
43) trans-1,3-Dichloroprop...		7.558	79	17410	5.327	ug/L	0.00
Spiked Amount	5.000	Range	55 - 130	Recovery	=	106.600%	
46) 2-Hexanone-d5		8.030	63	55270	60.578	ug/L	0.00
Spiked Amount	50.000	Range	45 - 130	Recovery	=	121.160%	
56) 1,1,2,2-Tetrachloroeth...		10.156	84	28071	5.794	ug/L	0.00
Spiked Amount	5.000	Range	65 - 120	Recovery	=	115.800%	
66) 1,2-Dichlorobenzene-d4		11.561	152	65304	5.826	ug/L	0.00
Spiked Amount	5.000	Range	80 - 120	Recovery	=	116.600%	
Target Compounds							Qvalue
14) Carbon disulfide		2.243	76	4063	0.178	ug/L	98
22) cis-1,2-Dichloroethene		3.821	96	4580	0.521	ug/L	91
34) Trichloroethene		5.841	95	7528	0.792	ug/L	84
42) Toluene		7.317	91	72145	1.932	ug/L	99
47) Tetrachloroethene		7.911	164	1045	0.116	ug/L	88
51) Chlorobenzene		8.818	112	7635	0.299	ug/L	97
65) 1,4-Dichlorobenzene		11.217	146	1739	0.081	ug/L	85
67) 1,2-Dichlorobenzene		11.583	146	2637	0.143	ug/L #	84

(#) = qualifier out of range (m) = manual integration (+) = signals summed

Data Path : Z:\voasrv\HPCHEM1\MSVOA_V\Data\VV121323\
Data File : VV033309.D
Acq On : 14 Dec 2023 04:43
Operator : SY/MD
Sample : 05770-09
Misc : 25.0mL/MSVOA_V/WATER
ALS Vial : 44 Sample Multiplier: 1

Instrument :
MSVOA_V
ClientSampleId :
E0Y81

Quant Time: Dec 14 05:46:19 2023
Quant Method : Z:\voasrv\HPCHEM1\MSVOA_V\Method\SFAMVTR121323WMA.M
Quant Title : TRACE VOA SFAM1.0
QLast Update : Thu Dec 14 01:07:25 2023
Response via : Initial Calibration

