

Data Path : Z:\voasrv\HPCHEM1\MSVOA_V\Data\VW121323\
 Data File : VW033318.D
 Acq On : 14 Dec 2023 08:19
 Operator : SY/MD
 Sample : 05770-20
 Misc : 25.0mL/MSVOA_V/WATER
 ALS Vial : 53 Sample Multiplier: 1

Instrument :
 MSVOA_V
 ClientSampleId :
 EOY90

Quant Time: Dec 14 11:18:43 2023
 Quant Method : Z:\voasrv\HPCHEM1\MSVOA_V\Method\SFAMVTR121323WMA.M
 Quant Title : TRACE VOA SFAM1.0
 QLast Update : Thu Dec 14 01:07:25 2023
 Response via : Initial Calibration

Compound	R.T.	QIon	Response	Conc	Units	Dev(Min)

Internal Standards						
1) 1,4-Difluorobenzene	5.535	114	124425	5.000	ug/L	0.00
28) Chlorobenzene-d5	8.786	117	128702	5.000	ug/L	0.00
58) 1,4-Dichlorobenzene-d4	11.188	152	65231	5.000	ug/L	0.00
System Monitoring Compounds						
4) Vinyl Chloride-d3	1.278	65	38257	4.796	ug/L	0.00
Spiked Amount	5.000	Range 40 - 130	Recovery	=	96.000%	
7) Chloroethane-d5	1.532	69	28805	4.856	ug/L	0.00
Spiked Amount	5.000	Range 65 - 130	Recovery	=	97.200%	
11) 1,1-Dichloroethene-d2	2.059	65	14787	4.658	ug/L	0.00
Spiked Amount	5.000	Range 60 - 125	Recovery	=	93.200%	
20) 2-Butanone-d5	3.812	46	48941	51.848	ug/L	0.02
Spiked Amount	50.000	Range 40 - 130	Recovery	=	103.700%	
24) Chloroform-d	4.249	84	83618	5.026	ug/L	0.00
Spiked Amount	5.000	Range 70 - 125	Recovery	=	100.600%	
26) 1,2-Dichloroethane-d4	4.944	65	40085	5.545	ug/L	0.00
Spiked Amount	5.000	Range 70 - 130	Recovery	=	110.800%	
32) Benzene-d6	4.960	84	156413	4.698	ug/L	0.00
Spiked Amount	5.000	Range 70 - 125	Recovery	=	94.000%	
36) 1,2-Dichloropropane-d6	5.992	67	38237	5.046	ug/L	0.00
Spiked Amount	5.000	Range 60 - 140	Recovery	=	101.000%	
41) Toluene-d8	7.246	98	148715	4.584	ug/L	0.00
Spiked Amount	5.000	Range 70 - 130	Recovery	=	91.600%	
43) trans-1,3-Dichloroprop...	7.554	79	15519	4.489	ug/L	0.00
Spiked Amount	5.000	Range 55 - 130	Recovery	=	89.800%	
46) 2-Hexanone-d5	8.034	63	53081	54.999	ug/L	0.00
Spiked Amount	50.000	Range 45 - 130	Recovery	=	110.000%	
56) 1,1,2,2-Tetrachloroeth...	10.152	84	27568	5.379	ug/L	0.00
Spiked Amount	5.000	Range 65 - 120	Recovery	=	107.600%	
66) 1,2-Dichlorobenzene-d4	11.564	152	63547	5.531	ug/L	0.00
Spiked Amount	5.000	Range 80 - 120	Recovery	=	110.600%	
Target Compounds						Qvalue
3) Chloromethane	1.214	50	1740	0.196	ug/L	99
14) Carbon disulfide	2.243	76	4307	0.186	ug/L	98
22) cis-1,2-Dichloroethene	3.815	96	25421	2.853	ug/L	99
34) Trichloroethene	5.834	95	9791	0.974	ug/L	80
42) Toluene	7.320	91	36235	0.917	ug/L	99
51) Chlorobenzene	8.818	112	3135	0.116	ug/L	90
65) 1,4-Dichlorobenzene	11.213	146	1687	0.077	ug/L	89
67) 1,2-Dichlorobenzene	11.577	146	2751	0.146	ug/L #	85

(#) = qualifier out of range (m) = manual integration (+) = signals summed

Data Path : Z:\voasrv\HPCHEM1\MSVOA_V\Data\VV121323\
Data File : VV033318.D
Acq On : 14 Dec 2023 08:19
Operator : SY/MD
Sample : 05770-20
Misc : 25.0mL/MSVOA_V/WATER
ALS Vial : 53 Sample Multiplier: 1

Instrument :
MSVOA_V
ClientSampleId :
EOY90

Quant Time: Dec 14 11:18:43 2023
Quant Method : Z:\voasrv\HPCHEM1\MSVOA_V\Method\SFAMVTR121323WMA.M
Quant Title : TRACE VOA SFAM1.0
QLast Update : Thu Dec 14 01:07:25 2023
Response via : Initial Calibration

