

Data Path : Z:\voasrv\HPCHEM1\MSVOA_V\Data\VV121323\
 Data File : VV033300.D
 Acq On : 14 Dec 2023 01:08
 Operator : SY/MD
 Sample : VV1213WBL02
 Misc : 25.0mL/MSVOA_V/WATER
 ALS Vial : 35 Sample Multiplier: 1

Instrument :
 MSVOA_V
 ClientSampleId :
 VBLK389

Quant Time: Dec 14 01:26:42 2023
 Quant Method : Z:\voasrv\HPCHEM1\MSVOA_V\Method\SFAMVTR121323WMA.M
 Quant Title : TRACE VOA SFAM1.0
 QLast Update : Thu Dec 14 01:07:25 2023
 Response via : Initial Calibration

Compound	R.T.	QIon	Response	Conc	Units	Dev(Min)

Internal Standards						
1) 1,4-Difluorobenzene	5.535	114	131149	5.000	ug/L	0.00
28) Chlorobenzene-d5	8.786	117	128448	5.000	ug/L	0.00
58) 1,4-Dichlorobenzene-d4	11.188	152	63797	5.000	ug/L	0.00
System Monitoring Compounds						
4) Vinyl Chloride-d3	1.278	65	42131	5.011	ug/L	0.00
Spiked Amount	5.000	Range 40 - 130	Recovery	= 100.200%		
7) Chloroethane-d5	1.532	69	31164	4.984	ug/L	0.00
Spiked Amount	5.000	Range 65 - 130	Recovery	= 99.600%		
11) 1,1-Dichloroethene-d2	2.060	65	16279	4.865	ug/L	0.00
Spiked Amount	5.000	Range 60 - 125	Recovery	= 97.200%		
20) 2-Butanone-d5	3.799	46	46710	46.948	ug/L	0.00
Spiked Amount	50.000	Range 40 - 130	Recovery	= 93.900%		
24) Chloroform-d	4.249	84	88026	5.020	ug/L	0.00
Spiked Amount	5.000	Range 70 - 125	Recovery	= 100.400%		
26) 1,2-Dichloroethane-d4	4.944	65	39743	5.216	ug/L	0.00
Spiked Amount	5.000	Range 70 - 130	Recovery	= 104.400%		
32) Benzene-d6	4.960	84	167569	5.043	ug/L	0.00
Spiked Amount	5.000	Range 70 - 125	Recovery	= 100.800%		
36) 1,2-Dichloropropane-d6	5.989	67	39748	5.255	ug/L	0.00
Spiked Amount	5.000	Range 60 - 140	Recovery	= 105.200%		
41) Toluene-d8	7.243	98	158216	4.886	ug/L	0.00
Spiked Amount	5.000	Range 70 - 130	Recovery	= 97.800%		
43) trans-1,3-Dichloroprop...	7.558	79	16661	4.829	ug/L	0.00
Spiked Amount	5.000	Range 55 - 130	Recovery	= 96.600%		
46) 2-Hexanone-d5	8.027	63	48082	49.918	ug/L	0.00
Spiked Amount	50.000	Range 45 - 130	Recovery	= 99.840%		
56) 1,1,2,2-Tetrachloroeth...	10.153	84	26052	5.093	ug/L	0.00
Spiked Amount	5.000	Range 65 - 120	Recovery	= 101.800%		
66) 1,2-Dichlorobenzene-d4	11.564	152	59927	5.333	ug/L	0.00
Spiked Amount	5.000	Range 80 - 120	Recovery	= 106.600%		

Target Compounds Qvalue

(#) = qualifier out of range (m) = manual integration (+) = signals summed

Data Path : Z:\voasrv\HPCHEM1\MSVOA_V\Data\VV121323\
Data File : VV033300.D
Acq On : 14 Dec 2023 01:08
Operator : SY/MD
Sample : VV1213WBL02
Misc : 25.0mL/MSVOA_V/WATER
ALS Vial : 35 Sample Multiplier: 1

Instrument :
MSVOA_V
ClientSampleId :
VBLK389

Quant Time: Dec 14 01:26:42 2023
Quant Method : Z:\voasrv\HPCHEM1\MSVOA_V\Method\SFAMVTR121323WMA.M
Quant Title : TRACE VOA SFAM1.0
QLast Update : Thu Dec 14 01:07:25 2023
Response via : Initial Calibration

