

Data Path : Z:\VOASRV\HPCHEM1\MSVOA V\DATA\VV121418\
 Data File : VV008930.D
 Acq On : 14 Dec 2018 15:29
 Operator : SY/MD
 Sample : J6367-16
 Misc : 5.0 mL/MSVOA V/WATER
 ALS Vial : 9 Sample Multiplier: 1

Instrument :
 MSVOA_V
 ClientSampleId :
 C0L52

Quant Time: Dec 20 05:46:48 2018
 Quant Method : Z:\VOASRV\HPCHEM1\MSVOA_V\METHOD\SOMVLM121318WMA.M
 Quant Title : VOC Analysis
 QLast Update : Fri Dec 14 23:39:04 2018
 Response via : Initial Calibration

Internal Standards	R.T.	QIon	Response	Conc	Units	Dev(Min)
1) 1,4-Difluorobenzene	5.67	114	205201	50.00	ug/L	0.00
28) Chlorobenzene-d5	8.90	117	185176	50.00	ug/L	0.00
60) 1,4-Dichlorobenzene-d4	11.30	152	78233	50.00	ug/L	0.00

System Monitoring Compounds

4) Vinyl Chloride-d3	1.32	65	53291	44.47	ug/L	0.00
Spiked Amount	50.000	Range	60 - 135	Recovery	=	88.94%
7) Chloroethane-d5	1.56	69	41577	55.13	ug/L	0.00
Spiked Amount	50.000	Range	70 - 130	Recovery	=	110.26%
11) 1,1-Dichloroethene-d2	2.13	63	109500	38.69	ug/L	0.00
Spiked Amount	50.000	Range	60 - 125	Recovery	=	77.38%
21) 2-Butanone-d5	3.93	46	89543	111.91	ug/L	0.00
Spiked Amount	100.000	Range	40 - 130	Recovery	=	111.91%
24) Chloroform-d	4.41	84	126821	49.60	ug/L	0.00
Spiked Amount	50.000	Range	70 - 125	Recovery	=	99.20%
26) 1,2-Dichloroethane-d4	5.08	65	85846	52.71	ug/L	0.00
Spiked Amount	50.000	Range	70 - 125	Recovery	=	105.42%
32) Benzene-d6	5.10	84	260802	52.26	ug/L	0.00
Spiked Amount	50.000	Range	70 - 125	Recovery	=	104.52%
36) 1,2-Dichloropropane-d6	6.12	67	78388	52.28	ug/L	0.00
Spiked Amount	50.000	Range	70 - 120	Recovery	=	104.56%
41) Toluene-d8	7.36	98	239824	51.20	ug/L	0.00
Spiked Amount	50.000	Range	80 - 120	Recovery	=	102.40%
43) trans-1,3-Dichloropropene-	7.67	79	38444	50.55	ug/L	0.00
Spiked Amount	50.000	Range	60 - 125	Recovery	=	101.10%
47) 2-Hexanone-d5	8.14	63	61300	110.45	ug/L	0.00
Spiked Amount	100.000	Range	45 - 130	Recovery	=	110.45%
57) 1,1,2,2-Tetrachloroethane-	10.26	84	95776	49.47	ug/L	0.00
Spiked Amount	50.000	Range	65 - 120	Recovery	=	98.94%
64) 1,2-Dichlorobenzene-d4	11.68	152	86175	53.41	ug/L	0.00
Spiked Amount	50.000	Range	80 - 120	Recovery	=	106.82%

Target Compounds	Ovalue

(#) = qualifier out of range (m) = manual integration (+) = signals summed

Data Path : Z:\V0ASRV\HPCHEM1\MSVOA V\DATA\VV121418\
Data File : VV008930.D
Acq On : 14 Dec 2018 15:29
Operator : SY/MD
Sample : J6367-16
Misc : 5.0 mL/MSVOA V/WATER
ALS Vial : 9 Sample Multiplier: 1

Instrument :
MSVOA_V
ClientSampleId :
COL52

Quant Time: Dec 20 05:46:48 2018
Quant Method : Z:\V0ASRV\HPCHEM1\MSVOA_V\METHOD\SOMVLM121318WMA.M
Quant Title : VOC Analysis
QLast Update : Fri Dec 14 23:39:04 2018
Response via : Initial Calibration

