

Data Path : Z:\VOASRV\HPCHEM1\MSVOA V\DATA\VV121418\
 Data File : VV008954.D
 Acq On : 15 Dec 2018 01:04
 Operator : SY/MD
 Sample : J6387-07
 Misc : 5.0 mL/MSVOA V/WATER
 ALS Vial : 33 Sample Multiplier: 1

Instrument :
 MSVOA_V
 ClientSampleId :
 C0LA9

Quant Time: Dec 15 04:17:35 2018
 Quant Method : Z:\VOASRV\HPCHEM1\MSVOA_V\METHOD\SOMVLM121318WMA.M
 Quant Title : VOC Analysis
 QLast Update : Sat Dec 15 01:01:00 2018
 Response via : Initial Calibration

Internal Standards	R.T.	QIon	Response	Conc	Units	Dev(Min)
1) 1,4-Difluorobenzene	5.67	114	203155	50.00	ug/L	0.00
28) Chlorobenzene-d5	8.90	117	191978	50.00	ug/L	0.00
60) 1,4-Dichlorobenzene-d4	11.30	152	78684	50.00	ug/L	0.00

System Monitoring Compounds

4) Vinyl Chloride-d3	1.32	65	48297	40.71	ug/L	0.00
Spiked Amount	50.000	Range	60 - 135	Recovery	=	81.42%
7) Chloroethane-d5	1.57	69	38974	52.20	ug/L	0.00
Spiked Amount	50.000	Range	70 - 130	Recovery	=	104.40%
11) 1,1-Dichloroethene-d2	2.13	63	98097	35.01	ug/L	0.00
Spiked Amount	50.000	Range	60 - 125	Recovery	=	70.02%
21) 2-Butanone-d5	3.93	46	84075	106.14	ug/L	0.00
Spiked Amount	100.000	Range	40 - 130	Recovery	=	106.14%
24) Chloroform-d	4.40	84	118623	46.86	ug/L	0.00
Spiked Amount	50.000	Range	70 - 125	Recovery	=	93.72%
26) 1,2-Dichloroethane-d4	5.08	65	79061	49.03	ug/L	0.00
Spiked Amount	50.000	Range	70 - 125	Recovery	=	98.06%
32) Benzene-d6	5.10	84	242298	46.83	ug/L	0.00
Spiked Amount	50.000	Range	70 - 125	Recovery	=	93.66%
36) 1,2-Dichloropropane-d6	6.12	67	74309	47.80	ug/L	0.00
Spiked Amount	50.000	Range	70 - 120	Recovery	=	95.60%
41) Toluene-d8	7.36	98	221869	45.69	ug/L	0.00
Spiked Amount	50.000	Range	80 - 120	Recovery	=	91.38%
43) trans-1,3-Dichloropropene-	7.67	79	34409	43.64	ug/L	0.00
Spiked Amount	50.000	Range	60 - 125	Recovery	=	87.28%
47) 2-Hexanone-d5	8.13	63	54399	94.55	ug/L	0.00
Spiked Amount	100.000	Range	45 - 130	Recovery	=	94.55%
57) 1,1,2,2-Tetrachloroethane-	10.26	84	90084	44.88	ug/L	0.00
Spiked Amount	50.000	Range	65 - 120	Recovery	=	89.76%
64) 1,2-Dichlorobenzene-d4	11.68	152	78800	48.56	ug/L	0.00
Spiked Amount	50.000	Range	80 - 120	Recovery	=	97.12%

Target Compounds					Ovalue
34) Trichloroethene	5.96	95	6431	4.210 ug/L	98

(#) = qualifier out of range (m) = manual integration (+) = signals summed

Data Path : Z:\V0ASRV\HPCHEM1\MSVOA V\DATA\VV121418\
Data File : VV008954.D
Acq On : 15 Dec 2018 01:04
Operator : SY/MD
Sample : J6387-07
Misc : 5.0 mL/MSVOA V/WATER
ALS Vial : 33 Sample Multiplier: 1

Instrument :
MSVOA_V
ClientSampleId :
C0LA9

Quant Time: Dec 15 04:17:35 2018
Quant Method : Z:\V0ASRV\HPCHEM1\MSVOA_V\METHOD\SOMVLM121318WMA.M
Quant Title : VOC Analysis
QLast Update : Sat Dec 15 01:01:00 2018
Response via : Initial Calibration

