

Data Path : Z:\VOASRV\HPCHEM1\MSVOA V\DATA\VV121418\
 Data File : VV008965.D
 Acq On : 15 Dec 2018 05:27
 Operator : SY/MD
 Sample : J6383-01
 Misc : 5.0 mL/MSVOA V/WATER
 ALS Vial : 44 Sample Multiplier: 1

Instrument :
 MSVOA_V
 ClientSampleId :
 C0934

Quant Time: Dec 22 05:57:02 2018
 Quant Method : Z:\VOASRV\HPCHEM1\MSVOA_V\METHOD\SOMVLM121318WMA.M
 Quant Title : VOC Analysis
 QLast Update : Sat Dec 15 01:01:00 2018
 Response via : Initial Calibration

Internal Standards	R.T.	QIon	Response	Conc	Units	Dev(Min)
1) 1,4-Difluorobenzene	5.67	114	202503	50.00	ug/L	0.00
28) Chlorobenzene-d5	8.90	117	188422	50.00	ug/L	0.00
60) 1,4-Dichlorobenzene-d4	11.30	152	80971	50.00	ug/L	0.00

System Monitoring Compounds

4) Vinyl Chloride-d3	1.32	65	47626	40.27	ug/L	0.00
Spiked Amount	50.000	Range	60 - 135	Recovery	=	80.54%
7) Chloroethane-d5	1.56	69	38399	51.59	ug/L	0.00
Spiked Amount	50.000	Range	70 - 130	Recovery	=	103.18%
11) 1,1-Dichloroethene-d2	2.13	63	99924	35.78	ug/L	0.00
Spiked Amount	50.000	Range	60 - 125	Recovery	=	71.56%
21) 2-Butanone-d5	3.93	46	90539	114.67	ug/L	0.00
Spiked Amount	100.000	Range	40 - 130	Recovery	=	114.67%
24) Chloroform-d	4.40	84	119849	47.50	ug/L	0.00
Spiked Amount	50.000	Range	70 - 125	Recovery	=	95.00%
26) 1,2-Dichloroethane-d4	5.08	65	82791	51.51	ug/L	0.00
Spiked Amount	50.000	Range	70 - 125	Recovery	=	103.02%
32) Benzene-d6	5.10	84	250151	49.26	ug/L	0.00
Spiked Amount	50.000	Range	70 - 125	Recovery	=	98.52%
36) 1,2-Dichloropropane-d6	6.12	67	74444	48.79	ug/L	0.00
Spiked Amount	50.000	Range	70 - 120	Recovery	=	97.58%
41) Toluene-d8	7.36	98	223886	46.97	ug/L	0.00
Spiked Amount	50.000	Range	80 - 120	Recovery	=	93.94%
43) trans-1,3-Dichloropropene-	7.67	79	34702	44.84	ug/L	0.00
Spiked Amount	50.000	Range	60 - 125	Recovery	=	89.68%
47) 2-Hexanone-d5	8.13	63	54904	97.22	ug/L	0.00
Spiked Amount	100.000	Range	45 - 130	Recovery	=	97.22%
57) 1,1,2,2-Tetrachloroethane-	10.26	84	88538	44.94	ug/L	0.00
Spiked Amount	50.000	Range	65 - 120	Recovery	=	89.88%
64) 1,2-Dichlorobenzene-d4	11.68	152	82515	49.41	ug/L	0.00
Spiked Amount	50.000	Range	80 - 120	Recovery	=	98.82%

Target Compounds	Ovalue

(#) = qualifier out of range (m) = manual integration (+) = signals summed

Data Path : Z:\VOASRV\HPCHEM1\MSVOA V\DATA\VV121418\
Data File : VV008965.D
Acq On : 15 Dec 2018 05:27
Operator : SY/MD
Sample : J6383-01
Misc : 5.0 mL/MSVOA V/WATER
ALS Vial : 44 Sample Multiplier: 1

Instrument :
MSVOA_V
ClientSampleId :
C0934

Quant Time: Dec 22 05:57:02 2018
Quant Method : Z:\VOASRV\HPCHEM1\MSVOA_V\METHOD\SOMVLM121318WMA.M
Quant Title : VOC Analysis
QLast Update : Sat Dec 15 01:01:00 2018
Response via : Initial Calibration

