

Data Path : Z:\VOASRV\HPCHEM1\MSVOA V\DATA\VV121418\
 Data File : VV008926.D
 Acq On : 14 Dec 2018 13:54
 Operator : SY/MD
 Sample : J6367-12
 Misc : 5.0 mL/MSVOA V/WATER
 ALS Vial : 5 Sample Multiplier: 1

Instrument :
 MSVOA_V
 ClientSampleId :
 COL48

Quant Time: Dec 15 00:02:30 2018
 Quant Method : Z:\VOASRV\HPCHEM1\MSVOA_V\METHOD\SOMVLM121318WMA.M
 Quant Title : VOC Analysis
 QLast Update : Fri Dec 14 23:39:04 2018
 Response via : Initial Calibration

Internal Standards	R.T.	QIon	Response	Conc	Units	Dev(Min)
1) 1,4-Difluorobenzene	5.67	114	206942	50.00	ug/L	0.00
28) Chlorobenzene-d5	8.90	117	189163	50.00	ug/L	0.00
60) 1,4-Dichlorobenzene-d4	11.30	152	81155	50.00	ug/L	0.00

System Monitoring Compounds

4) Vinyl Chloride-d3	1.32	65	54119	44.78	ug/L	0.00
Spiked Amount	50.000	Range	60 - 135	Recovery	=	89.56%
7) Chloroethane-d5	1.57	69	39304	51.68	ug/L	0.00
Spiked Amount	50.000	Range	70 - 130	Recovery	=	103.36%
11) 1,1-Dichloroethene-d2	2.13	63	104356	36.56	ug/L	0.00
Spiked Amount	50.000	Range	60 - 125	Recovery	=	73.12%
21) 2-Butanone-d5	3.93	46	83326	103.27	ug/L	0.00
Spiked Amount	100.000	Range	40 - 130	Recovery	=	103.27%
24) Chloroform-d	4.40	84	125235	48.57	ug/L	0.00
Spiked Amount	50.000	Range	70 - 125	Recovery	=	97.14%
26) 1,2-Dichloroethane-d4	5.08	65	83199	50.65	ug/L	0.00
Spiked Amount	50.000	Range	70 - 125	Recovery	=	101.30%
32) Benzene-d6	5.10	84	251205	49.28	ug/L	0.00
Spiked Amount	50.000	Range	70 - 125	Recovery	=	98.56%
36) 1,2-Dichloropropane-d6	6.12	67	76421	49.89	ug/L	0.00
Spiked Amount	50.000	Range	70 - 120	Recovery	=	99.78%
41) Toluene-d8	7.36	98	232928	48.68	ug/L	0.00
Spiked Amount	50.000	Range	80 - 120	Recovery	=	97.36%
43) trans-1,3-Dichloropropene-	7.67	79	37119	47.78	ug/L	0.00
Spiked Amount	50.000	Range	60 - 125	Recovery	=	95.56%
47) 2-Hexanone-d5	8.13	63	55105	97.20	ug/L	0.00
Spiked Amount	100.000	Range	45 - 130	Recovery	=	97.20%
57) 1,1,2,2-Tetrachloroethane-	10.26	84	94901	47.99	ug/L	0.00
Spiked Amount	50.000	Range	65 - 120	Recovery	=	95.98%
64) 1,2-Dichlorobenzene-d4	11.68	152	85419	51.03	ug/L	0.00
Spiked Amount	50.000	Range	80 - 120	Recovery	=	102.06%

Target Compounds	Ovalue

(#) = qualifier out of range (m) = manual integration (+) = signals summed

Data Path : Z:\V0ASRV\HPCHEM1\MSVOA V\DATA\VV121418\
Data File : VV008926.D
Acq On : 14 Dec 2018 13:54
Operator : SY/MD
Sample : J6367-12
Misc : 5.0 mL/MSVOA V/WATER
ALS Vial : 5 Sample Multiplier: 1

Instrument :
MSVOA_V
ClientSampleId :
COL48

Quant Time: Dec 15 00:02:30 2018
Quant Method : Z:\V0ASRV\HPCHEM1\MSVOA_V\METHOD\SOMVLM121318WMA.M
Quant Title : VOC Analysis
QLast Update : Fri Dec 14 23:39:04 2018
Response via : Initial Calibration

