

Data Path : Z:\VOASRV\HPCHEM1\MSVOA V\DATA\VV121418\
 Data File : VV008945.D
 Acq On : 14 Dec 2018 21:28
 Operator : SY/MD
 Sample : J6369-12
 Misc : 5.0 mL/MSVOA V/WATER
 ALS Vial : 24 Sample Multiplier: 1

Instrument :
 MSVOA_V
 ClientSampleId :
 COL96

Quant Time: Dec 15 00:59:33 2018
 Quant Method : Z:\VOASRV\HPCHEM1\MSVOA_V\METHOD\SOMVLM121318WMA.M
 Quant Title : VOC Analysis
 QLast Update : Fri Dec 14 23:39:04 2018
 Response via : Initial Calibration

Internal Standards	R.T.	QIon	Response	Conc	Units	Dev(Min)
1) 1,4-Difluorobenzene	5.67	114	206977	50.00	ug/L	0.00
28) Chlorobenzene-d5	8.90	117	189036	50.00	ug/L	0.00
60) 1,4-Dichlorobenzene-d4	11.30	152	82470	50.00	ug/L	0.00

System Monitoring Compounds

4) Vinyl Chloride-d3	1.32	65	51184	42.35	ug/L	0.00
Spiked Amount	50.000	Range	60 - 135	Recovery	=	84.70%
7) Chloroethane-d5	1.57	69	40159	52.79	ug/L	0.00
Spiked Amount	50.000	Range	70 - 130	Recovery	=	105.58%
11) 1,1-Dichloroethene-d2	2.13	63	106515	37.31	ug/L	0.00
Spiked Amount	50.000	Range	60 - 125	Recovery	=	74.62%
21) 2-Butanone-d5	3.93	46	88652	109.85	ug/L	0.00
Spiked Amount	100.000	Range	40 - 130	Recovery	=	109.85%
24) Chloroform-d	4.41	84	123690	47.96	ug/L	0.00
Spiked Amount	50.000	Range	70 - 125	Recovery	=	95.92%
26) 1,2-Dichloroethane-d4	5.08	65	83633	50.91	ug/L	0.00
Spiked Amount	50.000	Range	70 - 125	Recovery	=	101.82%
32) Benzene-d6	5.10	84	257304	50.51	ug/L	0.00
Spiked Amount	50.000	Range	70 - 125	Recovery	=	101.02%
36) 1,2-Dichloropropane-d6	6.12	67	77738	50.78	ug/L	0.00
Spiked Amount	50.000	Range	70 - 120	Recovery	=	101.56%
41) Toluene-d8	7.36	98	237607	49.69	ug/L	0.00
Spiked Amount	50.000	Range	80 - 120	Recovery	=	99.38%
43) trans-1,3-Dichloropropene-	7.67	79	36786	47.38	ug/L	0.00
Spiked Amount	50.000	Range	60 - 125	Recovery	=	94.76%
47) 2-Hexanone-d5	8.14	63	61296	108.19	ug/L	0.00
Spiked Amount	100.000	Range	45 - 130	Recovery	=	108.19%
57) 1,1,2,2-Tetrachloroethane-	10.26	84	94812	47.97	ug/L	0.00
Spiked Amount	50.000	Range	65 - 120	Recovery	=	95.94%
64) 1,2-Dichlorobenzene-d4	11.68	152	84760	49.83	ug/L	0.00
Spiked Amount	50.000	Range	80 - 120	Recovery	=	99.66%

Target Compounds	Ovalue

(#) = qualifier out of range (m) = manual integration (+) = signals summed

Data Path : Z:\V0ASRV\HPCHEM1\MSVOA V\DATA\VV121418\
Data File : VV008945.D
Acq On : 14 Dec 2018 21:28
Operator : SY/MD
Sample : J6369-12
Misc : 5.0 mL/MSVOA V/WATER
ALS Vial : 24 Sample Multiplier: 1

Instrument :
MSVOA_V
ClientSampleId :
COL96

Quant Time: Dec 15 00:59:33 2018
Quant Method : Z:\V0ASRV\HPCHEM1\MSVOA_V\METHOD\SOMVLM121318WMA.M
Quant Title : VOC Analysis
QLast Update : Fri Dec 14 23:39:04 2018
Response via : Initial Calibration

