

Data Path : Z:\VOASRV\HPCHEM1\MSVOA V\DATA\VV121418\
 Data File : VV008950.D
 Acq On : 14 Dec 2018 23:28
 Operator : SY/MD
 Sample : J6387-03
 Misc : 5.0 mL/MSVOA V/WATER
 ALS Vial : 29 Sample Multiplier: 1

Instrument :
 MSVOA_V
 ClientSampleId :
 C0LA5

Quant Time: Dec 24 07:41:19 2018
 Quant Method : Z:\VOASRV\HPCHEM1\MSVOA_V\METHOD\SOMVLM121318WMA.M
 Quant Title : VOC Analysis
 QLast Update : Sat Dec 15 01:01:00 2018
 Response via : Initial Calibration

Internal Standards	R.T.	QIon	Response	Conc	Units	Dev(Min)
1) 1,4-Difluorobenzene	5.67	114	208915	50.00	ug/L	0.00
28) Chlorobenzene-d5	8.90	117	192685	50.00	ug/L	0.00
60) 1,4-Dichlorobenzene-d4	11.30	152	80401	50.00	ug/L	0.00

System Monitoring Compounds

4) Vinyl Chloride-d3	1.32	65	48478	39.73	ug/L	0.00
Spiked Amount	50.000	Range	60 - 135	Recovery	=	79.46%
7) Chloroethane-d5	1.56	69	37457	48.78	ug/L	0.00
Spiked Amount	50.000	Range	70 - 130	Recovery	=	97.56%
11) 1,1-Dichloroethene-d2	2.13	63	104920	36.41	ug/L	0.00
Spiked Amount	50.000	Range	60 - 125	Recovery	=	72.82%
21) 2-Butanone-d5	3.93	46	86727	106.47	ug/L	0.00
Spiked Amount	100.000	Range	40 - 130	Recovery	=	106.47%
24) Chloroform-d	4.41	84	117066	44.97	ug/L	0.00
Spiked Amount	50.000	Range	70 - 125	Recovery	=	89.94%
26) 1,2-Dichloroethane-d4	5.08	65	78933	47.60	ug/L	0.00
Spiked Amount	50.000	Range	70 - 125	Recovery	=	95.20%
32) Benzene-d6	5.10	84	240164	46.25	ug/L	0.00
Spiked Amount	50.000	Range	70 - 125	Recovery	=	92.50%
36) 1,2-Dichloropropane-d6	6.12	67	74177	47.54	ug/L	0.00
Spiked Amount	50.000	Range	70 - 120	Recovery	=	95.08%
41) Toluene-d8	7.36	98	222205	45.59	ug/L	0.00
Spiked Amount	50.000	Range	80 - 120	Recovery	=	91.18%
43) trans-1,3-Dichloropropene-	7.67	79	34221	43.24	ug/L	0.00
Spiked Amount	50.000	Range	60 - 125	Recovery	=	86.48%
47) 2-Hexanone-d5	8.13	63	55600	96.28	ug/L	0.00
Spiked Amount	100.000	Range	45 - 130	Recovery	=	96.28%
57) 1,1,2,2-Tetrachloroethane-	10.26	84	91925	45.63	ug/L	0.00
Spiked Amount	50.000	Range	65 - 120	Recovery	=	91.26%
64) 1,2-Dichlorobenzene-d4	11.68	152	79434	47.90	ug/L	0.00
Spiked Amount	50.000	Range	80 - 120	Recovery	=	95.80%

Target Compounds	Ovalue

(#) = qualifier out of range (m) = manual integration (+) = signals summed

Data Path : Z:\V0ASRV\HPCHEM1\MSVOA V\DATA\VV121418\
Data File : VV008950.D
Acq On : 14 Dec 2018 23:28
Operator : SY/MD
Sample : J6387-03
Misc : 5.0 mL/MSVOA V/WATER
ALS Vial : 29 Sample Multiplier: 1

Instrument :
MSVOA_V
ClientSampleId :
C0LA5

Quant Time: Dec 24 07:41:19 2018
Quant Method : Z:\V0ASRV\HPCHEM1\MSVOA_V\METHOD\SOMVLM121318WMA.M
Quant Title : VOC Analysis
QLast Update : Sat Dec 15 01:01:00 2018
Response via : Initial Calibration

