

Data Path : Z:\VOASRV\HPCHEM1\MSVOA V\DATA\VV121418\
 Data File : VV008963.D
 Acq On : 15 Dec 2018 04:39
 Operator : SY/MD
 Sample : J6369-17
 Misc : 5.0 mL/MSVOA V/WATER
 ALS Vial : 42 Sample Multiplier: 1

Instrument :
 MSVOA_V
 ClientSampleId :
 C0LA1

Quant Time: Dec 15 05:53:29 2018
 Quant Method : Z:\VOASRV\HPCHEM1\MSVOA_V\METHOD\SOMVLM121318WMA.M
 Quant Title : VOC Analysis
 QLast Update : Sat Dec 15 01:01:00 2018
 Response via : Initial Calibration

Internal Standards	R.T.	QIon	Response	Conc	Units	Dev(Min)
1) 1,4-Difluorobenzene	5.67	114	202203	50.00	ug/L	0.00
28) Chlorobenzene-d5	8.90	117	187691	50.00	ug/L	0.00
60) 1,4-Dichlorobenzene-d4	11.30	152	78762	50.00	ug/L	0.00

System Monitoring Compounds

4) Vinyl Chloride-d3	1.32	65	46680	39.53	ug/L	0.00
Spiked Amount	50.000	Range	60 - 135	Recovery	=	79.06%
7) Chloroethane-d5	1.57	69	38116	51.29	ug/L	0.00
Spiked Amount	50.000	Range	70 - 130	Recovery	=	102.58%
11) 1,1-Dichloroethene-d2	2.13	63	96767	34.70	ug/L	0.00
Spiked Amount	50.000	Range	60 - 125	Recovery	=	69.40%
21) 2-Butanone-d5	3.93	46	86925	110.25	ug/L	0.00
Spiked Amount	100.000	Range	40 - 130	Recovery	=	110.25%
24) Chloroform-d	4.41	84	118156	46.90	ug/L	0.00
Spiked Amount	50.000	Range	70 - 125	Recovery	=	93.80%
26) 1,2-Dichloroethane-d4	5.08	65	80463	50.14	ug/L	0.00
Spiked Amount	50.000	Range	70 - 125	Recovery	=	100.28%
32) Benzene-d6	5.10	84	243049	48.05	ug/L	0.00
Spiked Amount	50.000	Range	70 - 125	Recovery	=	96.10%
36) 1,2-Dichloropropane-d6	6.12	67	73720	48.50	ug/L	0.00
Spiked Amount	50.000	Range	70 - 120	Recovery	=	97.00%
41) Toluene-d8	7.36	98	224726	47.33	ug/L	0.00
Spiked Amount	50.000	Range	80 - 120	Recovery	=	94.66%
43) trans-1,3-Dichloropropene-	7.67	79	33613	43.61	ug/L	0.00
Spiked Amount	50.000	Range	60 - 125	Recovery	=	87.22%
47) 2-Hexanone-d5	8.14	63	56583	100.59	ug/L	0.00
Spiked Amount	100.000	Range	45 - 130	Recovery	=	100.59%
57) 1,1,2,2-Tetrachloroethane-	10.26	84	90920	46.33	ug/L	0.00
Spiked Amount	50.000	Range	65 - 120	Recovery	=	92.66%
64) 1,2-Dichlorobenzene-d4	11.68	152	79437	48.90	ug/L	0.00
Spiked Amount	50.000	Range	80 - 120	Recovery	=	97.80%

Target Compounds	Ovalue

(#) = qualifier out of range (m) = manual integration (+) = signals summed

Data Path : Z:\V0ASRV\HPCHEM1\MSVOA V\DATA\VV121418\
Data File : VV008963.D
Acq On : 15 Dec 2018 04:39
Operator : SY/MD
Sample : J6369-17
Misc : 5.0 mL/MSVOA V/WATER
ALS Vial : 42 Sample Multiplier: 1

Instrument :
MSVOA_V
ClientSampleId :
C0LA1

Quant Time: Dec 15 05:53:29 2018
Quant Method : Z:\V0ASRV\HPCHEM1\MSVOA_V\METHOD\SOMVLM121318WMA.M
Quant Title : VOC Analysis
QLast Update : Sat Dec 15 01:01:00 2018
Response via : Initial Calibration

