

Data Path : Z:\VOASRV\HPCHEM1\MSVOA V\DATA\VV121420\
 Data File : VV019633.D
 Acq On : 15 Dec 2020 03:28
 Operator : SY/MD
 Sample : L5024-03
 Misc : 5.0mL/MSVOA V/WATER
 ALS Vial : 4 Sample Multiplier: 1

Instrument :
 MSVOA_V
 ClientSampleId :
 C0AX3

Quant Time: Dec 15 11:45:50 2020
 Quant Method : Z:\VOASRV\HPCHEM1\MSVOA_V\METHOD\SOMVLM121020WMA.M
 Quant Title : VOC Analysis
 QLast Update : Tue Dec 15 10:26:59 2020
 Response via : Initial Calibration

Internal Standards	R.T.	QIon	Response	Conc	Units	Dev(Min)
1) 1,4-Difluorobenzene	5.62	114	601607	50.00	ug/L	0.00
28) Chlorobenzene-d5	8.86	117	559905	50.00	ug/L	0.00
60) 1,4-Dichlorobenzene-d4	11.25	152	253671	50.00	ug/L	0.00

System Monitoring Compounds

4) Vinyl Chloride-d3	1.31	65	190830	44.24	ug/L	0.00
Spiked Amount	50.000	Range	60 - 135	Recovery	=	88.48%
7) Chloroethane-d5	1.57	69	166772	49.01	ug/L	0.00
Spiked Amount	50.000	Range	70 - 130	Recovery	=	98.02%
11) 1,1-Dichloroethene-d2	2.11	63	287668	35.93	ug/L	0.00
Spiked Amount	50.000	Range	60 - 125	Recovery	=	71.86%
21) 2-Butanone-d5	3.89	46	253496	108.11	ug/L	0.00
Spiked Amount	100.000	Range	40 - 130	Recovery	=	108.11%
24) Chloroform-d	4.35	84	384455	49.45	ug/L	0.00
Spiked Amount	50.000	Range	70 - 125	Recovery	=	98.90%
26) 1,2-Dichloroethane-d4	5.04	65	269676	53.07	ug/L	0.00
Spiked Amount	50.000	Range	70 - 125	Recovery	=	106.14%
32) Benzene-d6	5.05	84	756810	50.68	ug/L	0.00
Spiked Amount	50.000	Range	70 - 125	Recovery	=	101.36%
36) 1,2-Dichloropropane-d6	6.07	67	244912	52.36	ug/L	0.00
Spiked Amount	50.000	Range	70 - 120	Recovery	=	104.72%
41) Toluene-d8	7.32	98	670613	49.88	ug/L	0.00
Spiked Amount	50.000	Range	80 - 120	Recovery	=	99.76%
43) trans-1,3-Dichloropropene-	7.62	79	113896	42.81	ug/L	0.00
Spiked Amount	50.000	Range	60 - 125	Recovery	=	85.62%
47) 2-Hexanone-d5	8.09	63	178474	97.21	ug/L	0.00
Spiked Amount	100.000	Range	45 - 130	Recovery	=	97.21%
57) 1,1,2,2-Tetrachloroethane-	10.22	84	316508	50.02	ug/L	0.00
Spiked Amount	50.000	Range	65 - 120	Recovery	=	100.04%
64) 1,2-Dichlorobenzene-d4	11.63	152	247235	52.14	ug/L	0.00
Spiked Amount	50.000	Range	80 - 120	Recovery	=	104.28%

Target Compounds	Ovalue

(#) = qualifier out of range (m) = manual integration (+) = signals summed

Data Path : Z:\V0ASRV\HPCHEM1\MSVOA V\DATA\VV121420\
Data File : VV019633.D
Acq On : 15 Dec 2020 03:28
Operator : SY/MD
Sample : L5024-03
Misc : 5.0mL/MSVOA V/WATER
ALS Vial : 4 Sample Multiplier: 1

Instrument :
MSVOA_V
ClientSampleId :
C0AX3

Quant Time: Dec 15 11:45:50 2020
Quant Method : Z:\V0ASRV\HPCHEM1\MSVOA_V\METHOD\SOMVLM121020WMA.M
Quant Title : VOC Analysis
QLast Update : Tue Dec 15 10:26:59 2020
Response via : Initial Calibration

