

Data Path : Z:\VOASRV\HPCHEM1\MSVOA V\DATA\VV121420\
 Data File : VV019604.D
 Acq On : 14 Dec 2020 15:04
 Operator : SY/MD
 Sample : VIBLK71
 Misc : 5.0mL/MSVOA V/WATER
 ALS Vial : 4 Sample Multiplier: 1

Instrument :
 MSVOA_V
 ClientSampleId :
 VIBLK71

Quant Time: Dec 15 07:54:22 2020
 Quant Method : Z:\VOASRV\HPCHEM1\MSVOA_V\METHOD\SOMVLM121020WMA.M
 Quant Title : VOC Analysis
 QLast Update : Tue Dec 15 07:21:43 2020
 Response via : Initial Calibration

Internal Standards	R.T.	QIon	Response	Conc	Units	Dev(Min)
1) 1,4-Difluorobenzene	5.62	114	681270	50.00	ug/L	0.00
28) Chlorobenzene-d5	8.86	117	639748	50.00	ug/L	0.00
60) 1,4-Dichlorobenzene-d4	11.25	152	296647	50.00	ug/L	0.00

System Monitoring Compounds

4) Vinyl Chloride-d3	1.31	65	221591	45.36	ug/L	0.00
Spiked Amount	50.000	Range	60 - 135	Recovery	=	90.72%
7) Chloroethane-d5	1.57	69	188758	48.99	ug/L	0.00
Spiked Amount	50.000	Range	70 - 130	Recovery	=	97.98%
11) 1,1-Dichloroethene-d2	2.11	63	323768	35.71	ug/L	0.00
Spiked Amount	50.000	Range	60 - 125	Recovery	=	71.42%
21) 2-Butanone-d5	3.90	46	273110	102.86	ug/L	0.00
Spiked Amount	100.000	Range	40 - 130	Recovery	=	102.86%
24) Chloroform-d	4.35	84	404529	45.95	ug/L	0.00
Spiked Amount	50.000	Range	70 - 125	Recovery	=	91.90%
26) 1,2-Dichloroethane-d4	5.04	65	282836	49.16	ug/L	0.00
Spiked Amount	50.000	Range	70 - 125	Recovery	=	98.32%
32) Benzene-d6	5.05	84	811177	47.54	ug/L	0.00
Spiked Amount	50.000	Range	70 - 125	Recovery	=	95.08%
36) 1,2-Dichloropropane-d6	6.07	67	258368	48.34	ug/L	0.00
Spiked Amount	50.000	Range	70 - 120	Recovery	=	96.68%
41) Toluene-d8	7.32	98	719514	46.84	ug/L	0.00
Spiked Amount	50.000	Range	80 - 120	Recovery	=	93.68%
43) trans-1,3-Dichloropropene-	7.62	79	131681	43.31	ug/L	0.00
Spiked Amount	50.000	Range	60 - 125	Recovery	=	86.62%
47) 2-Hexanone-d5	8.09	63	198256	94.51	ug/L	0.00
Spiked Amount	100.000	Range	45 - 130	Recovery	=	94.51%
57) 1,1,2,2-Tetrachloroethane-	10.22	84	336563	46.55	ug/L	0.00
Spiked Amount	50.000	Range	65 - 120	Recovery	=	93.10%
64) 1,2-Dichlorobenzene-d4	11.63	152	261593	47.17	ug/L	0.00
Spiked Amount	50.000	Range	80 - 120	Recovery	=	94.34%

Target Compounds	Ovalue

(#) = qualifier out of range (m) = manual integration (+) = signals summed

Data Path : Z:\V0ASRV\HPCHEM1\MSVOA V\DATA\VV121420\
Data File : VV019604.D
Acq On : 14 Dec 2020 15:04
Operator : SY/MD
Sample : VIBLK71
Misc : 5.0mL/MSVOA V/WATER
ALS Vial : 4 Sample Multiplier: 1

Instrument :
MSVOA_V
ClientSampleId :
VIBLK71

Quant Time: Dec 15 07:54:22 2020
Quant Method : Z:\V0ASRV\HPCHEM1\MSVOA_V\METHOD\SOMVLM121020WMA.M
Quant Title : VOC Analysis
QLast Update : Tue Dec 15 07:21:43 2020
Response via : Initial Calibration

