

Data Path : Z:\VOASRV\HPCHEM1\MSVOA V\DATA\VV121420\
 Data File : VV019622.D
 Acq On : 14 Dec 2020 23:05
 Operator : SY/MD
 Sample : L4928-07
 Misc : 5.0mL/MSVOA V/WATER
 ALS Vial : 4 Sample Multiplier: 1

Instrument :
 MSVOA_V
 ClientSampleId :
 VHBLK01

Quant Time: Dec 15 11:00:20 2020
 Quant Method : Z:\VOASRV\HPCHEM1\MSVOA_V\METHOD\SOMVLM121020WMA.M
 Quant Title : VOC Analysis
 QLast Update : Tue Dec 15 10:26:59 2020
 Response via : Initial Calibration

Internal Standards	R.T.	QIon	Response	Conc	Units	Dev(Min)
1) 1,4-Difluorobenzene	5.62	114	598856	50.00	ug/L	0.00
28) Chlorobenzene-d5	8.86	117	553028	50.00	ug/L	0.00
60) 1,4-Dichlorobenzene-d4	11.25	152	253371	50.00	ug/L	0.00

System Monitoring Compounds

4) Vinyl Chloride-d3	1.31	65	201308	46.88	ug/L	0.00
Spiked Amount	50.000	Range	60 - 135	Recovery	=	93.76%
7) Chloroethane-d5	1.57	69	173855	51.33	ug/L	0.00
Spiked Amount	50.000	Range	70 - 130	Recovery	=	102.66%
11) 1,1-Dichloroethene-d2	2.11	63	300848	37.75	ug/L	0.00
Spiked Amount	50.000	Range	60 - 125	Recovery	=	75.50%
21) 2-Butanone-d5	3.89	46	269578	115.50	ug/L	0.00
Spiked Amount	100.000	Range	40 - 130	Recovery	=	115.50%
24) Chloroform-d	4.35	84	389256	50.30	ug/L	0.00
Spiked Amount	50.000	Range	70 - 125	Recovery	=	100.60%
26) 1,2-Dichloroethane-d4	5.04	65	272221	53.82	ug/L	0.00
Spiked Amount	50.000	Range	70 - 125	Recovery	=	107.64%
32) Benzene-d6	5.05	84	775441	52.57	ug/L	0.00
Spiked Amount	50.000	Range	70 - 125	Recovery	=	105.14%
36) 1,2-Dichloropropane-d6	6.07	67	249411	53.98	ug/L	0.00
Spiked Amount	50.000	Range	70 - 120	Recovery	=	107.96%
41) Toluene-d8	7.32	98	686458	51.69	ug/L	0.00
Spiked Amount	50.000	Range	80 - 120	Recovery	=	103.38%
43) trans-1,3-Dichloropropene-	7.62	79	118480	45.08	ug/L	0.00
Spiked Amount	50.000	Range	60 - 125	Recovery	=	90.16%
47) 2-Hexanone-d5	8.09	63	192318	106.06	ug/L	0.00
Spiked Amount	100.000	Range	45 - 130	Recovery	=	106.06%
57) 1,1,2,2-Tetrachloroethane-	10.22	84	328846	52.62	ug/L	0.00
Spiked Amount	50.000	Range	65 - 120	Recovery	=	105.24%
64) 1,2-Dichlorobenzene-d4	11.63	152	254504	53.74	ug/L	0.00
Spiked Amount	50.000	Range	80 - 120	Recovery	=	107.48%

Target Compounds	Ovalue

(#) = qualifier out of range (m) = manual integration (+) = signals summed

Data Path : Z:\V0ASRV\HPCHEM1\MSVOA V\DATA\VV121420\
Data File : VV019622.D
Acq On : 14 Dec 2020 23:05
Operator : SY/MD
Sample : L4928-07
Misc : 5.0mL/MSVOA V/WATER
ALS Vial : 4 Sample Multiplier: 1

Instrument :
MSVOA_V
ClientSampleId :
VHBLK01

Quant Time: Dec 15 11:00:20 2020
Quant Method : Z:\V0ASRV\HPCHEM1\MSVOA_V\METHOD\SOMVLM121020WMA.M
Quant Title : VOC Analysis
QLast Update : Tue Dec 15 10:26:59 2020
Response via : Initial Calibration

