

Data Path : Z:\voasrv\HPCHEM1\MSVOA_V\Data\VW121423\
 Data File : VW033332.D
 Acq On : 14 Dec 2023 14:03
 Operator : SY/MD
 Sample : 05773-08
 Misc : 5.0mL/MSVOA_V/WATER
 ALS Vial : 10 Sample Multiplier: 1

Instrument :
 MSVOA_V
 ClientSampleId :
 VHBLK001

Quant Time: Dec 15 01:04:56 2023
 Quant Method : Z:\voasrv\HPCHEM1\MSVOA_V\Method\SFAMVLM120123WMA.M
 Quant Title : VOC Analysis
 QLast Update : Fri Dec 08 21:51:12 2023
 Response via : Initial Calibration

Compound	R.T.	QIon	Response	Conc	Units	Dev(Min)

Internal Standards						
1) 1,4-Difluorobenzene	5.532	114	237128	50.000	ug/L	0.00
28) Chlorobenzene-d5	8.786	117	241841	50.000	ug/L	0.00
58) 1,4-Dichlorobenzene-d4	11.185	152	131429	50.000	ug/L	0.00
System Monitoring Compounds						
4) Vinyl Chloride-d3	1.278	65	59648	42.078	ug/L	0.00
Spiked Amount	50.000	Range	60 - 135	Recovery	=	84.160%
7) Chloroethane-d5	1.532	69	47725	48.319	ug/L	0.00
Spiked Amount	50.000	Range	70 - 130	Recovery	=	96.640%
11) 1,1-Dichloroethene-d2	2.060	65	25248	46.819	ug/L	0.00
Spiked Amount	50.000	Range	60 - 125	Recovery	=	93.640%
21) 2-Butanone-d5	3.783	46	55770	79.211	ug/L	0.00
Spiked Amount	100.000	Range	40 - 130	Recovery	=	79.210%
24) Chloroform-d	4.249	84	151571	50.384	ug/L	0.00
Spiked Amount	50.000	Range	70 - 125	Recovery	=	100.760%
26) 1,2-Dichloroethane-d4	4.941	65	98625	52.731	ug/L	0.00
Spiked Amount	50.000	Range	70 - 125	Recovery	=	105.460%
32) Benzene-d6	4.960	84	279312	49.382	ug/L	0.00
Spiked Amount	50.000	Range	70 - 125	Recovery	=	98.760%
36) 1,2-Dichloropropane-d6	5.989	67	71079	46.738	ug/L	0.00
Spiked Amount	50.000	Range	70 - 120	Recovery	=	93.480%
41) Toluene-d8	7.243	98	264910	49.097	ug/L	0.00
Spiked Amount	50.000	Range	80 - 120	Recovery	=	98.200%
43) trans-1,3-Dichloroprop...	7.555	79	44216	45.978	ug/L	0.00
Spiked Amount	50.000	Range	60 - 125	Recovery	=	91.960%
47) 2-Hexanone-d5	8.024	63	52359	85.700	ug/L	0.00
Spiked Amount	100.000	Range	45 - 130	Recovery	=	85.700%
56) 1,1,2,2-Tetrachloroeth...	10.153	84	108637	50.251	ug/L	0.00
Spiked Amount	50.000	Range	65 - 120	Recovery	=	100.500%
66) 1,2-Dichlorobenzene-d4	11.561	152	131730	56.333	ug/L	0.00
Spiked Amount	50.000	Range	80 - 120	Recovery	=	112.660%

Target Compounds Qvalue

(#) = qualifier out of range (m) = manual integration (+) = signals summed

Data Path : Z:\voasrv\HPCHEM1\MSVOA_V\Data\VV121423\
Data File : VV033332.D
Acq On : 14 Dec 2023 14:03
Operator : SY/MD
Sample : 05773-08
Misc : 5.0mL/MSVOA_V/WATER
ALS Vial : 10 Sample Multiplier: 1

Instrument :
MSVOA_V
ClientSampleId :
VHBLK001

Quant Time: Dec 15 01:04:56 2023
Quant Method : Z:\voasrv\HPCHEM1\MSVOA_V\Method\SFAMVLM120123WMA.M
Quant Title : VOC Analysis
QLast Update : Fri Dec 08 21:51:12 2023
Response via : Initial Calibration

