

Data Path : Z:\voasrv\HPCHEM1\MSVOA_V\Data\VW121423\
 Data File : VW033335.D
 Acq On : 14 Dec 2023 15:13
 Operator : SY/MD
 Sample : 05794-01
 Misc : 4.70g/5.0mL/100uL/5.00mL/MSVOA_V/MEOH
 ALS Vial : 13 Sample Multiplier: 1

Instrument :
 MSVOA_V
 ClientSampleId :
 BGRF7

Quant Time: Dec 15 01:05:23 2023
 Quant Method : Z:\voasrv\HPCHEM1\MSVOA_V\Method\SFAMVLM120123WMA.M
 Quant Title : VOC Analysis
 QLast Update : Fri Dec 08 21:51:12 2023
 Response via : Initial Calibration

Compound	R.T.	QIon	Response	Conc	Units	Dev(Min)

Internal Standards						
1) 1,4-Difluorobenzene	5.532	114	240646	50.000	ug/L	0.00
28) Chlorobenzene-d5	8.783	117	248032	50.000	ug/L	0.00
58) 1,4-Dichlorobenzene-d4	11.185	152	135458	50.000	ug/L	0.00
System Monitoring Compounds						
4) Vinyl Chloride-d3	1.278	65	59927	41.657	ug/L	0.00
Spiked Amount	50.000	Range	60 - 135	Recovery	=	83.320%
7) Chloroethane-d5	1.532	69	46955	46.845	ug/L	0.00
Spiked Amount	50.000	Range	70 - 130	Recovery	=	93.680%
11) 1,1-Dichloroethene-d2	2.060	65	24396	44.578	ug/L	0.00
Spiked Amount	50.000	Range	60 - 125	Recovery	=	89.160%
21) 2-Butanone-d5	3.783	46	55724	77.989	ug/L	0.00
Spiked Amount	100.000	Range	40 - 130	Recovery	=	77.990%
24) Chloroform-d	4.249	84	148126	48.519	ug/L	0.00
Spiked Amount	50.000	Range	70 - 125	Recovery	=	97.040%
26) 1,2-Dichloroethane-d4	4.941	65	97061	51.136	ug/L	0.00
Spiked Amount	50.000	Range	70 - 125	Recovery	=	102.280%
32) Benzene-d6	4.960	84	280439	48.344	ug/L	0.00
Spiked Amount	50.000	Range	70 - 125	Recovery	=	96.680%
36) 1,2-Dichloropropane-d6	5.989	67	70785	45.383	ug/L	0.00
Spiked Amount	50.000	Range	70 - 120	Recovery	=	90.760%
41) Toluene-d8	7.243	98	267764	48.388	ug/L	0.00
Spiked Amount	50.000	Range	80 - 120	Recovery	=	96.780%
43) trans-1,3-Dichloroprop...	7.551	79	42183	42.770	ug/L	0.00
Spiked Amount	50.000	Range	60 - 125	Recovery	=	85.540%
47) 2-Hexanone-d5	8.024	63	50971	81.346	ug/L	0.00
Spiked Amount	100.000	Range	45 - 130	Recovery	=	81.350%
56) 1,1,2,2-Tetrachloroeth...	10.153	84	108239	48.818	ug/L	0.00
Spiked Amount	50.000	Range	65 - 120	Recovery	=	97.640%
66) 1,2-Dichlorobenzene-d4	11.561	152	131338	54.495	ug/L	0.00
Spiked Amount	50.000	Range	80 - 120	Recovery	=	108.980%

Target Compounds Qvalue

(#) = qualifier out of range (m) = manual integration (+) = signals summed

Data Path : Z:\voasrv\HPCHEM1\MSVOA_V\Data\VV121423\
Data File : VV033335.D
Acq On : 14 Dec 2023 15:13
Operator : SY/MD
Sample : 05794-01
Misc : 4.70g/5.0mL/100uL/5.00mL/MSVOA_V/MEOH
ALS Vial : 13 Sample Multiplier: 1

Instrument :
MSVOA_V
ClientSampleId :
BGRF7

Quant Time: Dec 15 01:05:23 2023
Quant Method : Z:\voasrv\HPCHEM1\MSVOA_V\Method\SFAMVLM120123WMA.M
Quant Title : VOC Analysis
QLast Update : Fri Dec 08 21:51:12 2023
Response via : Initial Calibration

