

Data Path : Z:\voasrv\HPCHEM1\MSVOA_V\Data\VW121423\
 Data File : VW033345.D
 Acq On : 14 Dec 2023 19:04
 Operator : SY/MD
 Sample : 05794-12
 Misc : 3.98g/5.0mL/100uL/5.00mL/MSVOA_V/MEOH
 ALS Vial : 23 Sample Multiplier: 1

Instrument :
 MSVOA_V
 ClientSampleId :
 BH3Q7

Quant Time: Dec 15 01:06:59 2023
 Quant Method : Z:\voasrv\HPCHEM1\MSVOA_V\Method\SFAMVLM120123WMA.M
 Quant Title : VOC Analysis
 QLast Update : Fri Dec 08 21:51:12 2023
 Response via : Initial Calibration

Compound	R.T.	QIon	Response	Conc	Units	Dev(Min)

Internal Standards						
1) 1,4-Difluorobenzene	5.532	114	240067	50.000	ug/L	0.00
28) Chlorobenzene-d5	8.786	117	243045	50.000	ug/L	0.00
58) 1,4-Dichlorobenzene-d4	11.185	152	134906	50.000	ug/L	0.00
System Monitoring Compounds						
4) Vinyl Chloride-d3	1.278	65	53572	37.329	ug/L	0.00
Spiked Amount 50.000	Range 60 - 135		Recovery =	74.660%		
7) Chloroethane-d5	1.529	69	42945	42.947	ug/L	0.00
Spiked Amount 50.000	Range 70 - 130		Recovery =	85.900%		
11) 1,1-Dichloroethene-d2	2.060	65	23711	43.430	ug/L	0.00
Spiked Amount 50.000	Range 60 - 125		Recovery =	86.860%		
21) 2-Butanone-d5	3.786	46	56069	78.661	ug/L	0.00
Spiked Amount 100.000	Range 40 - 130		Recovery =	78.660%		
24) Chloroform-d	4.249	84	145181	47.669	ug/L	0.00
Spiked Amount 50.000	Range 70 - 125		Recovery =	95.340%		
26) 1,2-Dichloroethane-d4	4.941	65	98918	52.240	ug/L	0.00
Spiked Amount 50.000	Range 70 - 125		Recovery =	104.480%		
32) Benzene-d6	4.960	84	267418	47.045	ug/L	0.00
Spiked Amount 50.000	Range 70 - 125		Recovery =	94.100%		
36) 1,2-Dichloropropane-d6	5.989	67	68102	44.559	ug/L	0.00
Spiked Amount 50.000	Range 70 - 120		Recovery =	89.120%		
41) Toluene-d8	7.243	98	260210	47.987	ug/L	0.00
Spiked Amount 50.000	Range 80 - 120		Recovery =	95.980%		
43) trans-1,3-Dichloroprop...	7.555	79	42221	43.686	ug/L	0.00
Spiked Amount 50.000	Range 60 - 125		Recovery =	87.380%		
47) 2-Hexanone-d5	8.024	63	54250	88.355	ug/L	0.00
Spiked Amount 100.000	Range 45 - 130		Recovery =	88.360%		
56) 1,1,2,2-Tetrachloroeth...	10.153	84	104573	48.132	ug/L	0.00
Spiked Amount 50.000	Range 65 - 120		Recovery =	96.260%		
66) 1,2-Dichlorobenzene-d4	11.561	152	133509	55.622	ug/L	0.00
Spiked Amount 50.000	Range 80 - 120		Recovery =	111.240%		

Target Compounds Qvalue

(#) = qualifier out of range (m) = manual integration (+) = signals summed

Data Path : Z:\voasrv\HPCHEM1\MSVOA_V\Data\VV121423\
Data File : VV033345.D
Acq On : 14 Dec 2023 19:04
Operator : SY/MD
Sample : 05794-12
Misc : 3.98g/5.0mL/100uL/5.00mL/MSVOA_V/MEOH
ALS Vial : 23 Sample Multiplier: 1

Instrument :
MSVOA_V
ClientSampleId :
BH3Q7

Quant Time: Dec 15 01:06:59 2023
Quant Method : Z:\voasrv\HPCHEM1\MSVOA_V\Method\SFAMVLM120123WMA.M
Quant Title : VOC Analysis
QLast Update : Fri Dec 08 21:51:12 2023
Response via : Initial Calibration

