

Data Path : Z:\VOASRV\HPCHEM1\MSVOA V\DATA\VV121520\
 Data File : VV019658.D
 Acq On : 15 Dec 2020 18:47
 Operator : SY/MD
 Sample : L5024-19
 Misc : 5.0mL/MSVOA V/WATER
 ALS Vial : 1 Sample Multiplier: 1

Instrument :
 MSVOA_V
 ClientSampleId :
 C0AZ5

Quant Time: Dec 16 01:24:23 2020
 Quant Method : Z:\VOASRV\HPCHEM1\MSVOA_V\METHOD\SOMVLM121020WMA.M
 Quant Title : VOC Analysis
 QLast Update : Wed Dec 16 00:27:25 2020
 Response via : Initial Calibration

Internal Standards	R.T.	QIon	Response	Conc	Units	Dev(Min)
1) 1,4-Difluorobenzene	5.62	114	586759	50.00	ug/L	0.00
28) Chlorobenzene-d5	8.86	117	543516	50.00	ug/L	0.00
60) 1,4-Dichlorobenzene-d4	11.25	152	249802	50.00	ug/L	0.00

System Monitoring Compounds

4) Vinyl Chloride-d3	1.31	65	223556	53.13	ug/L	0.00
Spiked Amount	50.000	Range	60 - 135	Recovery	=	106.26%
7) Chloroethane-d5	1.57	69	177807	53.58	ug/L	0.00
Spiked Amount	50.000	Range	70 - 130	Recovery	=	107.16%
11) 1,1-Dichloroethene-d2	2.11	63	318396	40.77	ug/L	0.00
Spiked Amount	50.000	Range	60 - 125	Recovery	=	81.54%
21) 2-Butanone-d5	3.88	46	279238	122.10	ug/L	0.00
Spiked Amount	100.000	Range	40 - 130	Recovery	=	122.10%
24) Chloroform-d	4.35	84	383349	50.55	ug/L	0.00
Spiked Amount	50.000	Range	70 - 125	Recovery	=	101.10%
26) 1,2-Dichloroethane-d4	5.04	65	268965	54.27	ug/L	0.00
Spiked Amount	50.000	Range	70 - 125	Recovery	=	108.54%
32) Benzene-d6	5.05	84	788542	54.40	ug/L	0.00
Spiked Amount	50.000	Range	70 - 125	Recovery	=	108.80%
36) 1,2-Dichloropropane-d6	6.07	67	248963	54.83	ug/L	0.00
Spiked Amount	50.000	Range	70 - 120	Recovery	=	109.66%
41) Toluene-d8	7.32	98	692981	53.10	ug/L	0.00
Spiked Amount	50.000	Range	80 - 120	Recovery	=	106.20%
43) trans-1,3-Dichloropropene-	7.62	79	119264	46.18	ug/L	0.00
Spiked Amount	50.000	Range	60 - 125	Recovery	=	92.36%
47) 2-Hexanone-d5	8.09	63	200596	112.56	ug/L	0.00
Spiked Amount	100.000	Range	45 - 130	Recovery	=	112.56%
57) 1,1,2,2-Tetrachloroethane-	10.22	84	302458	49.24	ug/L	0.00
Spiked Amount	50.000	Range	65 - 120	Recovery	=	98.48%
64) 1,2-Dichlorobenzene-d4	11.63	152	251996	53.97	ug/L	0.00
Spiked Amount	50.000	Range	80 - 120	Recovery	=	107.94%

Target Compounds	Ovalue

(#) = qualifier out of range (m) = manual integration (+) = signals summed

Data Path : Z:\V0ASRV\HPCHEM1\MSVOA V\DATA\VV121520\
Data File : VV019658.D
Acq On : 15 Dec 2020 18:47
Operator : SY/MD
Sample : L5024-19
Misc : 5.0mL/MSVOA V/WATER
ALS Vial : 1 Sample Multiplier: 1

Instrument :
MSVOA_V
ClientSampleId :
C0AZ5

Quant Time: Dec 16 01:24:23 2020
Quant Method : Z:\V0ASRV\HPCHEM1\MSVOA_V\METHOD\SOMVLM121020WMA.M
Quant Title : VOC Analysis
QLast Update : Wed Dec 16 00:27:25 2020
Response via : Initial Calibration

