

Data Path : Z:\voasrv\HPCHEM1\MSVOA\_V\Data\VW121523\  
 Data File : VW033436.D  
 Acq On : 16 Dec 2023 15:40  
 Operator : SY/MD  
 Sample : 05802-10  
 Misc : 5.0mL/MSVOA\_V/WATER  
 ALS Vial : 67 Sample Multiplier: 1

Instrument :  
 MSVOA\_V  
 ClientSampleId :  
 F7B47

Quant Time: Dec 16 20:28:54 2023  
 Quant Method : Z:\voasrv\HPCHEM1\MSVOA\_V\Method\SFAMVLM121523WMA.M  
 Quant Title : VOC Analysis  
 QLast Update : Sat Dec 16 20:24:09 2023  
 Response via : Initial Calibration

| Compound                      | R.T.    | QIon  | Response | Conc     | Units | Dev(Min) |
|-------------------------------|---------|-------|----------|----------|-------|----------|
| -----                         |         |       |          |          |       |          |
| Internal Standards            |         |       |          |          |       |          |
| 1) 1,4-Difluorobenzene        | 5.535   | 114   | 213081   | 50.000   | ug/L  | 0.00     |
| 28) Chlorobenzene-d5          | 8.786   | 117   | 226572   | 50.000   | ug/L  | 0.00     |
| 58) 1,4-Dichlorobenzene-d4    | 11.188  | 152   | 94022    | 50.000   | ug/L  | 0.00     |
| System Monitoring Compounds   |         |       |          |          |       |          |
| 4) Vinyl Chloride-d3          | 1.278   | 65    | 69606    | 42.419   | ug/L  | 0.00     |
| Spiked Amount                 | 50.000  | Range | 60 - 135 | Recovery | =     | 84.840%  |
| 7) Chloroethane-d5            | 1.532   | 69    | 65942    | 44.780   | ug/L  | 0.00     |
| Spiked Amount                 | 50.000  | Range | 70 - 130 | Recovery | =     | 89.560%  |
| 11) 1,1-Dichloroethene-d2     | 2.060   | 65    | 34580    | 42.334   | ug/L  | 0.00     |
| Spiked Amount                 | 50.000  | Range | 60 - 125 | Recovery | =     | 84.660%  |
| 21) 2-Butanone-d5             | 3.783   | 46    | 78727    | 95.195   | ug/L  | 0.00     |
| Spiked Amount                 | 100.000 | Range | 40 - 130 | Recovery | =     | 95.200%  |
| 24) Chloroform-d              | 4.249   | 84    | 152493   | 46.980   | ug/L  | 0.00     |
| Spiked Amount                 | 50.000  | Range | 70 - 125 | Recovery | =     | 93.960%  |
| 26) 1,2-Dichloroethane-d4     | 4.941   | 65    | 98428    | 49.363   | ug/L  | 0.00     |
| Spiked Amount                 | 50.000  | Range | 70 - 125 | Recovery | =     | 98.720%  |
| 32) Benzene-d6                | 4.960   | 84    | 269980   | 44.242   | ug/L  | 0.00     |
| Spiked Amount                 | 50.000  | Range | 70 - 125 | Recovery | =     | 88.480%  |
| 36) 1,2-Dichloropropane-d6    | 5.989   | 67    | 86755    | 46.843   | ug/L  | 0.00     |
| Spiked Amount                 | 50.000  | Range | 70 - 120 | Recovery | =     | 93.680%  |
| 41) Toluene-d8                | 7.246   | 98    | 229593   | 41.118   | ug/L  | 0.00     |
| Spiked Amount                 | 50.000  | Range | 80 - 120 | Recovery | =     | 82.240%  |
| 43) trans-1,3-Dichloroprop... | 7.555   | 79    | 37701    | 41.536   | ug/L  | 0.00     |
| Spiked Amount                 | 50.000  | Range | 60 - 125 | Recovery | =     | 83.080%  |
| 47) 2-Hexanone-d5             | 8.024   | 63    | 55711    | 89.434   | ug/L  | 0.00     |
| Spiked Amount                 | 100.000 | Range | 45 - 130 | Recovery | =     | 89.430%  |
| 56) 1,1,2,2-Tetrachloroeth... | 10.153  | 84    | 127252   | 47.805   | ug/L  | 0.00     |
| Spiked Amount                 | 50.000  | Range | 65 - 120 | Recovery | =     | 95.620%  |
| 66) 1,2-Dichlorobenzene-d4    | 11.561  | 152   | 95629    | 54.125   | ug/L  | 0.00     |
| Spiked Amount                 | 50.000  | Range | 80 - 120 | Recovery | =     | 108.240% |

Target Compounds Qvalue

(#) = qualifier out of range (m) = manual integration (+) = signals summed

Data Path : Z:\voasrv\HPCHEM1\MSVOA\_V\Data\VV121523\  
Data File : VV033436.D  
Acq On : 16 Dec 2023 15:40  
Operator : SY/MD  
Sample : 05802-10  
Misc : 5.0mL/MSVOA\_V/WATER  
ALS Vial : 67 Sample Multiplier: 1

Instrument :  
MSVOA\_V  
ClientSampleId :  
F7B47

Quant Time: Dec 16 20:28:54 2023  
Quant Method : Z:\voasrv\HPCHEM1\MSVOA\_V\Method\SFAMVLM121523WMA.M  
Quant Title : VOC Analysis  
QLast Update : Sat Dec 16 20:24:09 2023  
Response via : Initial Calibration

