

Data File : VV014071.D
Acq On : 16 Dec 2019 14:47
Operator : SY/MD
Sample : K6275-04
Misc : 25.0mL/MSVOA_V/WATER
ALS Vial : 10 Sample Multiplier: 1

Instrument :
MSVOA_V
ClientSampleId :

Quant Time: Dec 17 06:30:18 2019
Quant Method : Z:\VOASRV\HPCHEM1\MSVOA_V\METHOD\SOMVTR112219WMA.M
Quant Title : TRACE VOA SOM01.0
QLast Update : Tue Dec 17 05:42:45 2019
Response via : Initial Calibration

Internal Standards	R.T.	QIon	Response	Conc	Units	Dev(Min)
1) 1,4-Difluorobenzene	5.66	114	375373	5.00	ug/L	0.00
28) Chlorobenzene-d5	8.89	117	384985	5.00	ug/L	0.00
60) 1,4-Dichlorobenzene-d4	11.29	152	168082	5.00	ug/L	0.00

System Monitoring Compounds

4) Vinyl Chloride-d3	1.32	65	100111	4.31	ug/L	0.00
Spiked Amount	5.000	Range	40 - 130	Recovery	=	86.20%
7) Chloroethane-d5	1.58	69	92404	4.68	ug/L	0.00
Spiked Amount	5.000	Range	65 - 130	Recovery	=	93.60%
11) 1,1-Dichloroethene-d2	2.13	63	125177	3.20	ug/L	0.00
Spiked Amount	5.000	Range	60 - 125	Recovery	=	64.00%
20) 2-Butanone-d5	3.95	46	300061	62.13	ug/L	0.00
Spiked Amount	50.000	Range	40 - 130	Recovery	=	124.26%
24) Chloroform-d	4.40	84	185108	3.95	ug/L	0.00
Spiked Amount	5.000	Range	70 - 125	Recovery	=	79.00%
26) 1,2-Dichloroethane-d4	5.08	65	107798	4.84	ug/L	0.00
Spiked Amount	5.000	Range	70 - 130	Recovery	=	96.80%
32) Benzene-d6	5.10	84	365375	3.68	ug/L	0.00
Spiked Amount	5.000	Range	70 - 125	Recovery	=	73.60%
36) 1,2-Dichloropropane-d6	6.12	67	127472	4.31	ug/L	0.00
Spiked Amount	5.000	Range	60 - 140	Recovery	=	86.20%
41) Toluene-d8	7.36	98	279613	3.00	ug/L	0.00
Spiked Amount	5.000	Range	70 - 130	Recovery	=	60.00%#
43) trans-1,3-Dichloropropene-	7.66	79	43510	3.78	ug/L	0.00
Spiked Amount	5.000	Range	55 - 130	Recovery	=	75.60%
46) 2-Hexanone-d5	8.14	63	204551	48.13	ug/L	0.00
Spiked Amount	50.000	Range	45 - 130	Recovery	=	96.26%
57) 1,1,2,2-Tetrachloroethane-	10.26	84	101888	4.96	ug/L	0.00
Spiked Amount	5.000	Range	65 - 120	Recovery	=	99.20%
64) 1,2-Dichlorobenzene-d4	11.67	152	137932	4.65	ug/L	0.00
Spiked Amount	5.000	Range	80 - 120	Recovery	=	93.00%

Target Compounds

					Qvalue
13) Acetone	2.21	43	13363m	3.957	ug/L
17) Methyl tert-butyl Ether	2.81	73	6429	0.130	ug/L # 73
22) cis-1,2-Dichloroethene	3.96	96	10736	0.449	ug/L 99
25) Chloroform	4.43	83	27581	0.639	ug/L 97
34) Trichloroethene	5.96	95	39489	1.516	ug/L 95

(#) = qualifier out of range (m) = manual integration (+) = signals summed

Data File : VV014071.D

```
Acq On       : 16 Dec 2019   14:47
Operator     : SY/MD
Sample       : K6275-04
Misc         : 25.0mL/MSVOA_V/WATER
ALS Vial     : 10      Sample Multiplier: 1
```

Instrument :
MSVOA_V
ClientSampleId :

Quant Time: Dec 17 06:30:18 2019
Quant Method : Z:\VOASRV\HPCHEM1\MSVOA_V\METHOD\SOMVTR112219WMA.M
Quant Title : TRACE VOA SOM01.0
QLast Update : Tue Dec 17 05:42:45 2019
Response via : Initial Calibration

