

Data Path : Z:\VOASRV\HPCHEM1\MSVOA_V\DATA\VV121619\
 Data File : VV014064.D
 Acq On : 16 Dec 2019 11:59
 Operator : SY/MD
 Sample : K6200-14
 Misc : 25.0mL/MSVOA_V/WATER
 ALS Vial : 3 Sample Multiplier: 1

Instrument :
 MSVOA_V
 ClientSampleId :
 VHBLK02

Quant Time: Dec 17 05:55:26 2019
 Quant Method : Z:\VOASRV\HPCHEM1\MSVOA_V\METHOD\SOMVTR112219WMA.M
 Quant Title : TRACE VOA SOM01.0
 QLast Update : Tue Dec 17 05:42:45 2019
 Response via : Initial Calibration

Internal Standards	R.T.	QIon	Response	Conc	Units	Dev(Min)
1) 1,4-Difluorobenzene	5.66	114	385871	5.00	ug/L	0.00
28) Chlorobenzene-d5	8.89	117	369964	5.00	ug/L	0.00
60) 1,4-Dichlorobenzene-d4	11.29	152	164609	5.00	ug/L	0.00

System Monitoring Compounds

4) Vinyl Chloride-d3	1.32	65	160993	6.74	ug/L	0.00
Spiked Amount	5.000	Range	40 - 130	Recovery	=	134.80%#
7) Chloroethane-d5	1.58	69	130004	6.41	ug/L	0.00
Spiked Amount	5.000	Range	65 - 130	Recovery	=	128.20%
11) 1,1-Dichloroethene-d2	2.13	63	200384	4.98	ug/L	0.00
Spiked Amount	5.000	Range	60 - 125	Recovery	=	99.60%
20) 2-Butanone-d5	3.95	46	196437	39.57	ug/L	0.00
Spiked Amount	50.000	Range	40 - 130	Recovery	=	79.14%
24) Chloroform-d	4.40	84	236975	4.92	ug/L	0.00
Spiked Amount	5.000	Range	70 - 125	Recovery	=	98.40%
26) 1,2-Dichloroethane-d4	5.07	65	115899	5.06	ug/L	0.00
Spiked Amount	5.000	Range	70 - 130	Recovery	=	101.20%
32) Benzene-d6	5.09	84	497942	5.22	ug/L	0.00
Spiked Amount	5.000	Range	70 - 125	Recovery	=	104.40%
36) 1,2-Dichloropropane-d6	6.11	67	146170	5.14	ug/L	0.00
Spiked Amount	5.000	Range	60 - 140	Recovery	=	102.80%
41) Toluene-d8	7.35	98	413333	4.61	ug/L	0.00
Spiked Amount	5.000	Range	70 - 130	Recovery	=	92.20%
43) trans-1,3-Dichloropropene-	7.66	79	47358	4.28	ug/L	0.00
Spiked Amount	5.000	Range	55 - 130	Recovery	=	85.60%
46) 2-Hexanone-d5	8.14	63	150485	36.85	ug/L	0.00
Spiked Amount	50.000	Range	45 - 130	Recovery	=	73.70%
57) 1,1,2,2-Tetrachloroethane-	10.26	84	85571	4.34	ug/L	0.00
Spiked Amount	5.000	Range	65 - 120	Recovery	=	86.80%
64) 1,2-Dichlorobenzene-d4	11.67	152	157372	5.42	ug/L	0.00
Spiked Amount	5.000	Range	80 - 120	Recovery	=	108.40%

Target Compounds					Qvalue
13) Acetone	2.21	43	5689	1.639 ug/L #	36
16) Methylene chloride	2.53	84	16591	0.653 ug/L	91

(#) = qualifier out of range (m) = manual integration (+) = signals summed

Data Path : Z:\V0ASRV\HPCHEM1\MSVOA_V\DATA\VV121619\
Data File : VV014064.D
Acq On : 16 Dec 2019 11:59
Operator : SY/MD
Sample : K6200-14
Misc : 25.0mL/MSVOA_V/WATER
ALS Vial : 3 Sample Multiplier: 1

Instrument :
MSVOA_V
ClientSampleId :
VHBLK02

Quant Time: Dec 17 05:55:26 2019
Quant Method : Z:\V0ASRV\HPCHEM1\MSVOA_V\METHOD\SOMVTR112219WMA.M
Quant Title : TRACE VOA SOM01.0
QLast Update : Tue Dec 17 05:42:45 2019
Response via : Initial Calibration

