

Data Path : Z:\VOASRV\HPCHEM1\MSVOA_V\DATA\VV121619\
 Data File : VV014067.D
 Acq On : 16 Dec 2019 13:14
 Operator : SY/MD
 Sample : VIBLK45
 Misc : 25.0mL/MSVOA_V/WATER
 ALS Vial : 6 Sample Multiplier: 1

Instrument :
 MSVOA_V
 ClientSampleId :
 VIBLK45

Quant Time: Dec 17 06:16:20 2019
 Quant Method : Z:\VOASRV\HPCHEM1\MSVOA_V\METHOD\SOMVTR112219WMA.M
 Quant Title : TRACE VOA SOM01.0
 QLast Update : Tue Dec 17 05:42:45 2019
 Response via : Initial Calibration

Internal Standards	R.T.	QIon	Response	Conc	Units	Dev(Min)
1) 1,4-Difluorobenzene	5.66	114	375962	5.00	ug/L	0.00
28) Chlorobenzene-d5	8.89	117	369207	5.00	ug/L	0.00
60) 1,4-Dichlorobenzene-d4	11.29	152	160450	5.00	ug/L	0.00

System Monitoring Compounds

4) Vinyl Chloride-d3	1.32	65	151683	6.52	ug/L	0.00
Spiked Amount	5.000	Range	40 - 130	Recovery	=	130.40%#
7) Chloroethane-d5	1.58	69	124222	6.29	ug/L	0.00
Spiked Amount	5.000	Range	65 - 130	Recovery	=	125.80%
11) 1,1-Dichloroethene-d2	2.13	63	189908	4.85	ug/L	0.00
Spiked Amount	5.000	Range	60 - 125	Recovery	=	97.00%
20) 2-Butanone-d5	3.93	46	223250	46.15	ug/L	-0.02
Spiked Amount	50.000	Range	40 - 130	Recovery	=	92.30%
24) Chloroform-d	4.40	84	224148	4.78	ug/L	0.00
Spiked Amount	5.000	Range	70 - 125	Recovery	=	95.60%
26) 1,2-Dichloroethane-d4	5.08	65	110820	4.96	ug/L	0.00
Spiked Amount	5.000	Range	70 - 130	Recovery	=	99.20%
32) Benzene-d6	5.10	84	463039	4.86	ug/L	0.00
Spiked Amount	5.000	Range	70 - 125	Recovery	=	97.20%
36) 1,2-Dichloropropane-d6	6.11	67	136897	4.82	ug/L	0.00
Spiked Amount	5.000	Range	60 - 140	Recovery	=	96.40%
41) Toluene-d8	7.36	98	372431	4.17	ug/L	0.00
Spiked Amount	5.000	Range	70 - 130	Recovery	=	83.40%
43) trans-1,3-Dichloropropene-	7.66	79	46294	4.19	ug/L	0.00
Spiked Amount	5.000	Range	55 - 130	Recovery	=	83.80%
46) 2-Hexanone-d5	8.13	63	135626	33.28	ug/L	0.00
Spiked Amount	50.000	Range	45 - 130	Recovery	=	66.56%
57) 1,1,2,2-Tetrachloroethane-	10.26	84	78500	3.99	ug/L	0.00
Spiked Amount	5.000	Range	65 - 120	Recovery	=	79.80%
64) 1,2-Dichlorobenzene-d4	11.67	152	144844	5.11	ug/L	0.00
Spiked Amount	5.000	Range	80 - 120	Recovery	=	102.20%

Target Compounds					Qvalue
13) Acetone	2.19	43	8105	2.396 ug/L #	43
16) Methylene chloride	2.53	84	11680	0.472 ug/L	88

(#) = qualifier out of range (m) = manual integration (+) = signals summed

Data Path : Z:\V0ASRV\HPCHEM1\MSVOA_V\DATA\VV121619\
Data File : VV014067.D
Acq On : 16 Dec 2019 13:14
Operator : SY/MD
Sample : VIBLK45
Misc : 25.0mL/MSVOA_V/WATER
ALS Vial : 6 Sample Multiplier: 1

Instrument :
MSVOA_V
ClientSampleId :
VIBLK45

Quant Time: Dec 17 06:16:20 2019
Quant Method : Z:\V0ASRV\HPCHEM1\MSVOA_V\METHOD\SOMVTR112219WMA.M
Quant Title : TRACE VOA SOM01.0
QLast Update : Tue Dec 17 05:42:45 2019
Response via : Initial Calibration

