

Data Path : Z:\VOASRV\HPCHEM1\MSVOA_V\DATA\VV121619\
 Data File : VV014069.D
 Acq On : 16 Dec 2019 14:01
 Operator : SY/MD
 Sample : K6200-14
 Misc : 25.0mL/MSVOA_V/WATER
 ALS Vial : 8 Sample Multiplier: 1

Instrument :
 MSVOA_V
 ClientSampleId :
 VHBLK02

Quant Time: Dec 17 06:21:46 2019
 Quant Method : Z:\VOASRV\HPCHEM1\MSVOA_V\METHOD\SOMVTR112219WMA.M
 Quant Title : TRACE VOA SOM01.0
 QLast Update : Tue Dec 17 05:42:45 2019
 Response via : Initial Calibration

Internal Standards	R.T.	QIon	Response	Conc	Units	Dev(Min)
1) 1,4-Difluorobenzene	5.66	114	392523	5.00	ug/L	0.00
28) Chlorobenzene-d5	8.89	117	398668	5.00	ug/L	0.00
60) 1,4-Dichlorobenzene-d4	11.29	152	169496	5.00	ug/L	0.00

System Monitoring Compounds

4) Vinyl Chloride-d3	1.32	65	150544	6.20	ug/L	0.00
Spiked Amount	5.000	Range	40 - 130	Recovery	=	124.00%
7) Chloroethane-d5	1.58	69	124562	6.04	ug/L	0.00
Spiked Amount	5.000	Range	65 - 130	Recovery	=	120.80%
11) 1,1-Dichloroethene-d2	2.13	63	189789	4.64	ug/L	0.00
Spiked Amount	5.000	Range	60 - 125	Recovery	=	92.80%
20) 2-Butanone-d5	3.94	46	258025	51.09	ug/L	0.00
Spiked Amount	50.000	Range	40 - 130	Recovery	=	102.18%
24) Chloroform-d	4.40	84	219465	4.48	ug/L	0.00
Spiked Amount	5.000	Range	70 - 125	Recovery	=	89.60%
26) 1,2-Dichloroethane-d4	5.08	65	116086	4.98	ug/L	0.00
Spiked Amount	5.000	Range	70 - 130	Recovery	=	99.60%
32) Benzene-d6	5.10	84	470519	4.57	ug/L	0.00
Spiked Amount	5.000	Range	70 - 125	Recovery	=	91.40%
36) 1,2-Dichloropropane-d6	6.11	67	138974	4.53	ug/L	0.00
Spiked Amount	5.000	Range	60 - 140	Recovery	=	90.60%
41) Toluene-d8	7.36	98	387818	4.02	ug/L	0.00
Spiked Amount	5.000	Range	70 - 130	Recovery	=	80.40%
43) trans-1,3-Dichloropropene-	7.66	79	49946	4.19	ug/L	0.00
Spiked Amount	5.000	Range	55 - 130	Recovery	=	83.80%
46) 2-Hexanone-d5	8.13	63	169143	38.44	ug/L	0.00
Spiked Amount	50.000	Range	45 - 130	Recovery	=	76.88%
57) 1,1,2,2-Tetrachloroethane-	10.26	84	89903	4.23	ug/L	0.00
Spiked Amount	5.000	Range	65 - 120	Recovery	=	84.60%
64) 1,2-Dichlorobenzene-d4	11.67	152	150050	5.01	ug/L	0.00
Spiked Amount	5.000	Range	80 - 120	Recovery	=	100.20%

Target Compounds					Qvalue
13) Acetone	2.20	43	8879	2.514 ug/L #	31
16) Methylene chloride	2.54	84	17118	0.662 ug/L	95

(#) = qualifier out of range (m) = manual integration (+) = signals summed

Data Path : Z:\V0ASRV\HPCHEM1\MSVOA_V\DATA\VV121619\
Data File : VV014069.D
Acq On : 16 Dec 2019 14:01
Operator : SY/MD
Sample : K6200-14
Misc : 25.0mL/MSVOA_V/WATER
ALS Vial : 8 Sample Multiplier: 1

Instrument :
MSVOA_V
ClientSampleId :
VHBLK02

Quant Time: Dec 17 06:21:46 2019
Quant Method : Z:\V0ASRV\HPCHEM1\MSVOA_V\METHOD\SOMVTR112219WMA.M
Quant Title : TRACE VOA SOM01.0
QLast Update : Tue Dec 17 05:42:45 2019
Response via : Initial Calibration

