


Data Path : Z:\voasrv\HPCHEM1\MSVOA_V\Data\VV121619\
 Data File : VV014061.D
 Acq On : 16 Dec 2019 09:16
 Operator : SY/MD
 Sample : BFB42
 Misc : 25.0mL/MSVOA_V/WATER
 ALS Vial : 1 Sample Multiplier: 1

Instrument :
 MSVOA_V
 ClientSampleId :
 BFB42

Integration File: rteint.p

Method : Z:\VOASRV\HPCHEM1\MSVOA_V\METHOD\SOMVTR112219WMA.M
 Title : TRACE VOA SOM01.0
 Last Update : Tue Dec 17 05:42:45 2019



AutoFind: Scans 2806, 2807, 2808; Background Corrected with Scan 2792

Target Mass	Rel. to Mass	Lower Limit%	Upper Limit%	Rel. Abn%	Raw Abn	Result Pass/Fail
50	95	15	40	16.5	40986	PASS
75	95	30	80	48.0	119469	PASS
95	95	100	100	100.0	248981	PASS
96	95	5	9	6.7	16675	PASS
173	174	0.00	2	0.0	0	PASS
174	95	50	120	97.1	241749	PASS
175	174	5	9	7.3	17578	PASS
176	174	95	101	96.5	233365	PASS
177	176	5	9	6.4	15033	PASS

m/z	Abundance
35.10	345.0
36.00	3737.0
37.10	10928.0
38.10	10175.0
39.10	4313.0
40.00	948.0
41.50	236.0
41.90	134.0
43.10	176.0
44.10	1700.0
45.00	1915.0
46.10	315.0
47.00	3093.0
48.00	1254.0
49.00	8700.0
50.00	42144.0
51.10	13437.0
52.10	733.0
53.00	547.0
54.00	169.0
54.90	590.0
56.00	3249.0
57.00	5217.0
58.10	238.0
60.00	2496.0
61.00	10668.0
62.10	11307.0
63.00	9678.0
64.10	1858.0
65.00	2112.0
66.00	339.0
67.10	797.0
68.00	22440.0
69.00	22952.0
70.00	1583.0
71.00	123.0
72.10	1332.0
73.10	10167.0
74.00	38736.0
75.00	122056.0
76.10	10082.0
77.00	1324.0
77.90	1219.0
78.90	6325.0
80.00	1930.0
81.00	7061.0
82.00	1465.0
83.00	251.0
85.90	428.0
87.00	9397.0
88.00	9785.0
91.00	990.0
92.00	7448.0
93.00	10972.0
94.00	28088.0
95.00	250688.0
96.10	17416.0
97.00	513.0
103.80	773.0
104.90	351.0
105.90	812.0
106.70	212.0
107.20	150.0
108.10	139.0
111.10	172.0
112.00	126.0
113.00	197.0
114.90	324.0
115.90	959.0
117.00	1816.0
117.90	975.0
118.90	1645.0
121.20	119.0
121.70	110.0
123.00	155.0
124.00	146.0
125.80	106.0
126.80	104.0
127.90	892.0
128.80	473.0
129.90	944.0
131.00	338.0
134.80	313.0
137.00	308.0
139.40	110.0
139.90	239.0
140.90	2461.0
141.80	317.0
143.00	3981.0
144.10	210.0
144.90	1275.0
145.90	571.0

Instrument :
MSVOA_V
ClientSampleId :
BFB42

147.20	259.0
147.90	578.0
148.70	174.0
149.90	356.0
151.70	225.0
152.80	278.0
154.00	215.0
155.00	550.0
156.00	172.0
156.80	432.0
157.70	109.0
158.90	426.0
160.70	287.0
171.00	848.0
171.90	3178.0
174.00	233280.0
175.00	17264.0
176.00	229568.0
177.00	15311.0
178.00	337.0
206.70	102.0

Instrument :
MSVOA_V
ClientSampleId :
BFB42

m/z	Abundance
35.10	294.0
36.00	3962.0
37.10	11178.0
38.10	10162.0
39.00	4423.0
40.00	804.0
41.60	239.0
43.10	163.0
44.10	1931.0
45.00	2105.0
46.10	304.0
47.00	3002.0
48.00	1393.0
49.10	9115.0
50.10	42016.0
51.10	13622.0
52.10	667.0
52.90	510.0
54.00	194.0
54.90	503.0
56.00	3677.0
57.00	5523.0
58.10	305.0
60.00	2533.0
61.00	11159.0
62.10	11510.0
63.00	9969.0
64.10	1782.0
65.00	2209.0
66.00	316.0
67.10	793.0
68.00	23488.0
69.00	23368.0
70.10	1582.0
71.10	157.0
72.10	1351.0
73.10	10374.0
74.00	39152.0
75.00	123424.0
76.10	9893.0
76.90	1385.0
77.90	1271.0
78.90	6451.0
80.00	2127.0
81.00	6965.0
82.00	1412.0
82.90	212.0
83.10	183.0
84.90	101.0
85.90	400.0
87.00	9830.0
88.00	9851.0
89.80	109.0
90.90	967.0
92.10	7645.0
93.00	11311.0
94.00	28904.0
95.00	257920.0
96.10	17216.0
97.00	577.0
103.90	802.0
104.90	344.0
105.90	1011.0
106.80	207.0
108.10	133.0
109.90	126.0
111.10	183.0
111.80	130.0
112.90	226.0
114.90	353.0
115.90	1065.0
116.90	1789.0
117.90	931.0
118.90	1633.0
121.70	144.0
123.90	145.0
125.90	111.0
127.10	111.0
127.90	925.0
128.90	452.0
129.90	960.0
131.00	359.0
134.90	344.0
137.00	331.0
140.00	244.0
140.90	2581.0
141.80	331.0
143.00	3997.0
144.90	1371.0
145.90	646.0
146.80	192.0
147.20	185.0

Instrument :
MSVOA_V
ClientSampleId :
BFB42

147.90	674.0
148.80	194.0
149.90	410.0
151.70	246.0
152.90	230.0
154.00	256.0
155.10	555.0
155.90	165.0
156.90	467.0
158.90	367.0
160.80	333.0
171.00	736.0
171.90	3371.0
174.00	250624.0
175.00	18080.0
176.00	241536.0
177.00	15504.0
178.00	349.0
280.80	100.0

Instrument :
MSVOA_V
ClientSampleId :
BFB42

m/z	Abundance
36.00	3983.0
37.10	10495.0
38.10	9162.0
39.00	4195.0
40.00	803.0
41.60	171.0
42.10	149.0
42.90	197.0
44.10	1828.0
45.00	1836.0
46.10	284.0
47.00	2682.0
48.00	1428.0
49.10	8741.0
50.10	38800.0
51.00	12378.0
52.00	544.0
52.90	428.0
54.00	219.0
55.00	411.0
56.00	3359.0
57.00	5550.0
58.10	339.0
59.00	111.0
60.00	2344.0
61.00	10678.0
62.10	10612.0
63.00	9210.0
64.00	1542.0
65.00	1975.0
66.00	257.0
67.00	684.0
68.00	22320.0
69.00	22240.0
70.10	1475.0
71.10	181.0
72.10	1198.0
73.10	9526.0
74.00	36024.0
75.00	112928.0
76.10	8917.0
76.90	1286.0
77.80	1129.0
78.90	5994.0
80.00	2149.0
81.00	6472.0
82.00	1280.0
82.70	195.0
83.30	153.0
86.00	426.0
87.00	9351.0
88.00	9038.0
89.80	104.0
90.90	813.0
92.10	6796.0
93.00	10604.0
94.10	27112.0
95.00	238336.0
96.10	15394.0
97.00	602.0
103.80	722.0
105.10	328.0
105.90	1078.0
106.80	205.0
108.00	120.0
109.90	143.0
111.10	195.0
111.80	139.0
112.90	231.0
114.80	330.0
115.90	1090.0
116.90	1727.0
117.90	872.0
118.90	1537.0
123.80	157.0
124.90	116.0
125.90	121.0
127.10	111.0
128.00	847.0
128.90	379.0
129.90	892.0
130.90	380.0
134.80	381.0
136.90	357.0
140.10	223.0
140.90	2350.0
141.90	370.0
142.90	3444.0
143.90	230.0
144.90	1451.0
145.90	584.0
146.80	194.0

Instrument :
MSVOA_V
ClientSampleId :
BFB42

147.10	149.0
147.90	642.0
148.90	247.0
149.90	360.0
151.70	195.0
153.00	265.0
154.00	198.0
155.00	619.0
155.80	181.0
156.90	515.0
158.90	282.0
160.80	346.0
171.00	605.0
172.00	3367.0
174.00	241344.0
175.00	17392.0
176.00	228992.0
177.00	14284.0
178.00	311.0
215.20	107.0
234.70	102.0

Instrument :
MSVOA_V
ClientSampleId :
BFB42

m/z	Abundance
36.00	1921.0
36.90	224.0
37.90	896.0
38.90	143.0
39.90	688.0
41.70	114.0
42.30	134.0
43.00	106.0
44.00	646.0
45.00	172.0
45.70	118.0
47.70	104.0
48.80	118.0
50.90	141.0
58.00	109.0
61.10	123.0
67.30	101.0
68.70	138.0
69.10	160.0
74.00	118.0
114.40	105.0
117.70	134.0
132.90	108.0
147.10	134.0

Instrument :
MSVOA_V
ClientSampleId :
BFB42