

Data Path : Z:\VOASRV\HPCHEM1\MSVOA V\DATA\VV121620\
 Data File : VV019668.D
 Acq On : 16 Dec 2020 12:36
 Operator : SY/MD
 Sample : VV1216WBL01
 Misc : 5.0mL/MSVOA V/WATER
 ALS Vial : 1 Sample Multiplier: 1

Instrument :
 MSVOA_V
 ClientSampleId :
 VBLK70

Quant Time: Dec 17 00:38:02 2020
 Quant Method : Z:\VOASRV\HPCHEM1\MSVOA_V\METHOD\SOMVLM121020WMA.M
 Quant Title : VOC Analysis
 QLast Update : Thu Dec 17 00:30:07 2020
 Response via : Initial Calibration

Internal Standards	R.T.	QIon	Response	Conc	Units	Dev(Min)
1) 1,4-Difluorobenzene	5.62	114	645100	50.00	ug/L	0.00
28) Chlorobenzene-d5	8.86	117	604813	50.00	ug/L	0.00
60) 1,4-Dichlorobenzene-d4	11.25	152	298232	50.00	ug/L	0.00

System Monitoring Compounds

4) Vinyl Chloride-d3	1.31	65	223601	48.34	ug/L	0.00
Spiked Amount	50.000	Range	60 - 135	Recovery	=	96.68%
7) Chloroethane-d5	1.57	69	179373	49.16	ug/L	0.00
Spiked Amount	50.000	Range	70 - 130	Recovery	=	98.32%
11) 1,1-Dichloroethene-d2	2.11	63	316101	36.82	ug/L	0.00
Spiked Amount	50.000	Range	60 - 125	Recovery	=	73.64%
21) 2-Butanone-d5	3.88	46	303609	120.75	ug/L	0.00
Spiked Amount	100.000	Range	40 - 130	Recovery	=	120.75%
24) Chloroform-d	4.35	84	388587	46.61	ug/L	0.00
Spiked Amount	50.000	Range	70 - 125	Recovery	=	93.22%
26) 1,2-Dichloroethane-d4	5.03	65	267562	49.11	ug/L	0.00
Spiked Amount	50.000	Range	70 - 125	Recovery	=	98.22%
32) Benzene-d6	5.05	84	783934	48.60	ug/L	0.00
Spiked Amount	50.000	Range	70 - 125	Recovery	=	97.20%
36) 1,2-Dichloropropane-d6	6.07	67	250487	49.57	ug/L	0.00
Spiked Amount	50.000	Range	70 - 120	Recovery	=	99.14%
41) Toluene-d8	7.32	98	690952	47.58	ug/L	0.00
Spiked Amount	50.000	Range	80 - 120	Recovery	=	95.16%
43) trans-1,3-Dichloropropene-	7.62	79	124336	43.26	ug/L	0.00
Spiked Amount	50.000	Range	60 - 125	Recovery	=	86.52%
47) 2-Hexanone-d5	8.09	63	211301	106.55	ug/L	0.00
Spiked Amount	100.000	Range	45 - 130	Recovery	=	106.55%
57) 1,1,2,2-Tetrachloroethane-	10.22	84	309800	45.33	ug/L	0.00
Spiked Amount	50.000	Range	65 - 120	Recovery	=	90.66%
64) 1,2-Dichlorobenzene-d4	11.63	152	256692	46.04	ug/L	0.00
Spiked Amount	50.000	Range	80 - 120	Recovery	=	92.08%

Target Compounds	Ovalue

(#) = qualifier out of range (m) = manual integration (+) = signals summed

Data Path : Z:\V0ASRV\HPCHEM1\MSVOA V\DATA\VV121620\
Data File : VV019668.D
Acq On : 16 Dec 2020 12:36
Operator : SY/MD
Sample : VV1216WBL01
Misc : 5.0mL/MSVOA V/WATER
ALS Vial : 1 Sample Multiplier: 1

Instrument :
MSVOA_V
ClientSampleId :
VBLK70

Quant Time: Dec 17 00:38:02 2020
Quant Method : Z:\V0ASRV\HPCHEM1\MSVOA_V\METHOD\SOMVLM121020WMA.M
Quant Title : VOC Analysis
QLast Update : Thu Dec 17 00:30:07 2020
Response via : Initial Calibration

