

Data Path : Z:\VOASRV\HPCHEM1\MSVOA V\DATA\VV121620\
 Data File : VV019678.D
 Acq On : 16 Dec 2020 16:35
 Operator : SY/MD
 Sample : L5062-10
 Misc : 5.0mL/MSVOA V/WATER
 ALS Vial : 1 Sample Multiplier: 1

Instrument :
 MSVOA_V
 ClientSampleId :
 C0B21

Quant Time: Dec 17 01:24:19 2020
 Quant Method : Z:\VOASRV\HPCHEM1\MSVOA_V\METHOD\SOMVLM121020WMA.M
 Quant Title : VOC Analysis
 QLast Update : Thu Dec 17 00:30:07 2020
 Response via : Initial Calibration

Internal Standards	R.T.	QIon	Response	Conc	Units	Dev(Min)
1) 1,4-Difluorobenzene	5.62	114	612833	50.00	ug/L	0.00
28) Chlorobenzene-d5	8.86	117	568949	50.00	ug/L	0.00
60) 1,4-Dichlorobenzene-d4	11.25	152	258429	50.00	ug/L	0.00

System Monitoring Compounds

4) Vinyl Chloride-d3	1.31	65	210864	47.98	ug/L	0.00
Spiked Amount	50.000	Range	60 - 135	Recovery	=	95.96%
7) Chloroethane-d5	1.57	69	172402	49.74	ug/L	0.00
Spiked Amount	50.000	Range	70 - 130	Recovery	=	99.48%
11) 1,1-Dichloroethene-d2	2.11	63	300684	36.87	ug/L	0.00
Spiked Amount	50.000	Range	60 - 125	Recovery	=	73.74%
21) 2-Butanone-d5	3.90	46	296124	123.98	ug/L	0.02
Spiked Amount	100.000	Range	40 - 130	Recovery	=	123.98%
24) Chloroform-d	4.35	84	370760	46.81	ug/L	0.00
Spiked Amount	50.000	Range	70 - 125	Recovery	=	93.62%
26) 1,2-Dichloroethane-d4	5.04	65	257852	49.82	ug/L	0.00
Spiked Amount	50.000	Range	70 - 125	Recovery	=	99.64%
32) Benzene-d6	5.05	84	751114	49.50	ug/L	0.00
Spiked Amount	50.000	Range	70 - 125	Recovery	=	99.00%
36) 1,2-Dichloropropane-d6	6.07	67	238840	50.25	ug/L	0.00
Spiked Amount	50.000	Range	70 - 120	Recovery	=	100.50%
41) Toluene-d8	7.32	98	665069	48.68	ug/L	0.00
Spiked Amount	50.000	Range	80 - 120	Recovery	=	97.36%
43) trans-1,3-Dichloropropene-	7.62	79	107897	39.91	ug/L	0.00
Spiked Amount	50.000	Range	60 - 125	Recovery	=	79.82%
47) 2-Hexanone-d5	8.09	63	196467	105.31	ug/L	0.00
Spiked Amount	100.000	Range	45 - 130	Recovery	=	105.31%
57) 1,1,2,2-Tetrachloroethane-	10.22	84	299004	46.51	ug/L	0.00
Spiked Amount	50.000	Range	65 - 120	Recovery	=	93.02%
64) 1,2-Dichlorobenzene-d4	11.63	152	237132	49.09	ug/L	0.00
Spiked Amount	50.000	Range	80 - 120	Recovery	=	98.18%

Target Compounds	Ovalue

(#) = qualifier out of range (m) = manual integration (+) = signals summed

Data Path : Z:\V0ASRV\HPCHEM1\MSVOA V\DATA\VV121620\
Data File : VV019678.D
Acq On : 16 Dec 2020 16:35
Operator : SY/MD
Sample : L5062-10
Misc : 5.0mL/MSVOA V/WATER
ALS Vial : 1 Sample Multiplier: 1

Instrument :
MSVOA_V
ClientSampleId :
C0B21

Quant Time: Dec 17 01:24:19 2020
Quant Method : Z:\V0ASRV\HPCHEM1\MSVOA_V\METHOD\SOMVLM121020WMA.M
Quant Title : VOC Analysis
QLast Update : Thu Dec 17 00:30:07 2020
Response via : Initial Calibration

