

Data Path : Z:\voasrv\HPCHEM1\MSVOA_V\Data\VW121621\
 Data File : VW024016.D
 Acq On : 16 Dec 2021 16:27
 Operator : SY/MD
 Sample : M5090-19
 Misc : 25.0mL/MSVOA_V/WATER
 ALS Vial : 18 Sample Multiplier: 1

Instrument :
 MSVOA_V
 ClientSampleId :
 BG3F5

Quant Time: Dec 17 02:01:41 2021
 Quant Method : Z:\voasrv\HPCHEM1\MSVOA_V\Method\SFAMVTR121021WMA.M
 Quant Title : TRACE VOA SFAM1.0
 QLast Update : Thu Dec 16 04:44:29 2021
 Response via : Initial Calibration

Compound	R.T.	QIon	Response	Conc	Units	Dev(Min)

Internal Standards						
1) 1,4-Difluorobenzene	5.619	114	95177	5.000	ug/L	0.00
28) Chlorobenzene-d5	8.853	117	102234	5.000	ug/L	0.00
58) 1,4-Dichlorobenzene-d4	11.249	152	57852	5.000	ug/L	0.00
System Monitoring Compounds						
4) Vinyl Chloride-d3	1.304	65	24047	4.700	ug/L	0.00
Spiked Amount	5.000	Range	40 - 130	Recovery	=	94.000%
7) Chloroethane-d5	1.568	69	20094	4.707	ug/L	0.00
Spiked Amount	5.000	Range	65 - 130	Recovery	=	94.200%
11) 1,1-Dichloroethene-d2	2.108	63	35926	3.597	ug/L	0.00
Spiked Amount	5.000	Range	60 - 125	Recovery	=	72.000%
20) 2-Butanone-d5	3.918	46	54622	54.642	ug/L	0.00
Spiked Amount	50.000	Range	40 - 130	Recovery	=	109.280%
24) Chloroform-d	4.349	84	52565	4.535	ug/L	0.00
Spiked Amount	5.000	Range	70 - 125	Recovery	=	90.800%
26) 1,2-Dichloroethane-d4	5.037	65	27343	5.067	ug/L	0.00
Spiked Amount	5.000	Range	70 - 130	Recovery	=	101.400%
32) Benzene-d6	5.050	84	101135	4.995	ug/L	0.00
Spiked Amount	5.000	Range	70 - 125	Recovery	=	99.800%
36) 1,2-Dichloropropane-d6	6.069	67	28851	4.723	ug/L	0.00
Spiked Amount	5.000	Range	60 - 140	Recovery	=	94.400%
41) Toluene-d8	7.316	98	89058	4.885	ug/L	0.00
Spiked Amount	5.000	Range	70 - 130	Recovery	=	97.800%
43) trans-1,3-Dichloroprop...	7.625	79	10652	4.203	ug/L	0.00
Spiked Amount	5.000	Range	55 - 130	Recovery	=	84.000%
46) 2-Hexanone-d5	8.091	63	47909	42.227	ug/L	0.00
Spiked Amount	50.000	Range	45 - 130	Recovery	=	84.460%
56) 1,1,2,2-Tetrachloroeth...	10.217	84	22867	4.512	ug/L	0.00
Spiked Amount	5.000	Range	65 - 120	Recovery	=	90.200%
66) 1,2-Dichlorobenzene-d4	11.625	152	39985	5.174	ug/L	0.00
Spiked Amount	5.000	Range	80 - 120	Recovery	=	103.400%
Target Compounds						Qvalue
22) cis-1,2-Dichloroethene	3.928	96	2054	0.275	ug/L	93
33) Benzene	5.111	78	6004	0.208	ug/L	100
42) Toluene	7.397	91	3942	0.121	ug/L	87
51) Chlorobenzene	8.882	112	4872	0.224	ug/L	89
52) Ethylbenzene	9.021	91	2123	0.062	ug/L	99
60) Isopropylbenzene	9.934	105	20763	0.611	ug/L	97
65) 1,4-Dichlorobenzene	11.281	146	2761	0.158	ug/L	88
67) 1,2-Dichlorobenzene	11.644	146	8751	0.550	ug/L	94

(#) = qualifier out of range (m) = manual integration (+) = signals summed

Data Path : Z:\voasrv\HPCHEM1\MSVOA_V\Data\VV121621\
Data File : VV024016.D
Acq On : 16 Dec 2021 16:27
Operator : SY/MD
Sample : M5090-19
Misc : 25.0mL/MSVOA_V/WATER
ALS Vial : 18 Sample Multiplier: 1

Instrument :
MSVOA_V
ClientSampleId :
BG3F5

Quant Time: Dec 17 02:01:41 2021
Quant Method : Z:\voasrv\HPCHEM1\MSVOA_V\Method\SFAMVTR121021WMA.M
Quant Title : TRACE VOA SFAM1.0
QLast Update : Thu Dec 16 04:44:29 2021
Response via : Initial Calibration

