

Data Path : Z:\voasrv\HPCHEM1\MSVOA_V\Data\VW121621\
 Data File : VW024045.D
 Acq On : 17 Dec 2021 04:53
 Operator : SY/MD
 Sample : M5133-18
 Misc : 25.0mL/MSVOA_V/WATER
 ALS Vial : 49 Sample Multiplier: 1

Instrument :
 MSVOA_V
 ClientSampleId :
 GBBJ8

Quant Time: Dec 17 05:35:01 2021
 Quant Method : Z:\voasrv\HPCHEM1\MSVOA_V\Method\SFAMVTR121021WMA.M
 Quant Title : TRACE VOA SFAM1.0
 QLast Update : Thu Dec 16 04:44:29 2021
 Response via : Initial Calibration

Compound	R.T.	QIon	Response	Conc	Units	Dev(Min)

Internal Standards						
1) 1,4-Difluorobenzene	5.619	114	93584	5.000	ug/L	0.00
28) Chlorobenzene-d5	8.854	117	97204	5.000	ug/L	0.00
58) 1,4-Dichlorobenzene-d4	11.249	152	47219	5.000	ug/L	0.00
System Monitoring Compounds						
4) Vinyl Chloride-d3	1.304	65	22887	4.549	ug/L	0.00
Spiked Amount	5.000	Range 40 - 130	Recovery	=	91.000%	
7) Chloroethane-d5	1.568	69	20526	4.890	ug/L	0.00
Spiked Amount	5.000	Range 65 - 130	Recovery	=	97.800%	
11) 1,1-Dichloroethene-d2	2.105	63	36051	3.671	ug/L	0.00
Spiked Amount	5.000	Range 60 - 125	Recovery	=	73.400%	
20) 2-Butanone-d5	3.905	46	44488	45.262	ug/L	0.00
Spiked Amount	50.000	Range 40 - 130	Recovery	=	90.520%	
24) Chloroform-d	4.349	84	53233	4.671	ug/L	0.00
Spiked Amount	5.000	Range 70 - 125	Recovery	=	93.400%	
26) 1,2-Dichloroethane-d4	5.034	65	27769	5.234	ug/L	0.00
Spiked Amount	5.000	Range 70 - 130	Recovery	=	104.600%	
32) Benzene-d6	5.050	84	99084	5.147	ug/L	0.00
Spiked Amount	5.000	Range 70 - 125	Recovery	=	103.000%	
36) 1,2-Dichloropropane-d6	6.072	67	28921	4.979	ug/L	0.00
Spiked Amount	5.000	Range 60 - 140	Recovery	=	99.600%	
41) Toluene-d8	7.317	98	85330	4.923	ug/L	0.00
Spiked Amount	5.000	Range 70 - 130	Recovery	=	98.400%	
43) trans-1,3-Dichloroprop...	7.625	79	10851	4.503	ug/L	0.00
Spiked Amount	5.000	Range 55 - 130	Recovery	=	90.000%	
46) 2-Hexanone-d5	8.092	63	39140	36.283	ug/L	0.00
Spiked Amount	50.000	Range 45 - 130	Recovery	=	72.560%	
56) 1,1,2,2-Tetrachloroeth...	10.217	84	20023	4.155	ug/L	0.00
Spiked Amount	5.000	Range 65 - 120	Recovery	=	83.200%	
66) 1,2-Dichlorobenzene-d4	11.622	152	35197	5.580	ug/L	0.00
Spiked Amount	5.000	Range 80 - 120	Recovery	=	111.600%	
Target Compounds						
					Qvalue	
13) Acetone	2.195	43	19599	24.924	ug/L	85
21) 2-Butanone	3.979	43	118160	105.166	ug/L	84
31) Carbon tetrachloride	4.828	117	229912	19.232	ug/L	99

(#) = qualifier out of range (m) = manual integration (+) = signals summed

Data Path : Z:\voasrv\HPCHEM1\MSVOA_V\Data\VV121621\
Data File : VV024045.D
Acq On : 17 Dec 2021 04:53
Operator : SY/MD
Sample : M5133-18
Misc : 25.0mL/MSVOA_V/WATER
ALS Vial : 49 Sample Multiplier: 1

Instrument :
MSVOA_V
ClientSampleId :
GBBJ8

Quant Time: Dec 17 05:35:01 2021
Quant Method : Z:\voasrv\HPCHEM1\MSVOA_V\Method\SFAMVTR121021WMA.M
Quant Title : TRACE VOA SFAM1.0
QLast Update : Thu Dec 16 04:44:29 2021
Response via : Initial Calibration

