

Data Path : Z:\voasrv\HPCHEM1\MSVOA_V\Data\VW121622\
 Data File : VW029694.D
 Acq On : 16 Dec 2022 13:44
 Operator : SY/MD
 Sample : VW1216WBL02
 Misc : 25.0mL/MSVOA_V/WATER
 ALS Vial : 4 Sample Multiplier: 1

Instrument :
 MSVOA_V
 ClientSampleId :
 VBLK288

Quant Time: Dec 16 20:18:18 2022
 Quant Method : Z:\voasrv\HPCHEM1\MSVOA_V\Method\SFAMVTR112322WMA.M
 Quant Title : TRACE VOA SFAM1.0
 QLast Update : Fri Dec 16 20:17:28 2022
 Response via : Initial Calibration

Compound	R.T.	QIon	Response	Conc	Units	Dev(Min)

Internal Standards						
1) 1,4-Difluorobenzene	5.609	114	229756	5.000	ug/L	0.00
28) Chlorobenzene-d5	8.844	117	211535	5.000	ug/L	0.00
58) 1,4-Dichlorobenzene-d4	11.236	152	108938	5.000	ug/L	0.00
System Monitoring Compounds						
4) Vinyl Chloride-d3	1.301	65	50544	3.575	ug/L	0.00
Spiked Amount	5.000	Range 40 - 130	Recovery	=	71.400%	
7) Chloroethane-d5	1.561	69	47490	4.588	ug/L	0.00
Spiked Amount	5.000	Range 65 - 130	Recovery	=	91.800%	
11) 1,1-Dichloroethene-d2	2.101	63	70993	3.138	ug/L	0.00
Spiked Amount	5.000	Range 60 - 125	Recovery	=	62.800%	
20) 2-Butanone-d5	3.912	46	117457	49.658	ug/L	0.00
Spiked Amount	50.000	Range 40 - 130	Recovery	=	99.320%	
24) Chloroform-d	4.339	84	111520	4.419	ug/L	0.00
Spiked Amount	5.000	Range 70 - 125	Recovery	=	88.400%	
26) 1,2-Dichloroethane-d4	5.024	65	50521	4.550	ug/L	0.00
Spiked Amount	5.000	Range 70 - 130	Recovery	=	91.000%	
32) Benzene-d6	5.040	84	227134	3.885	ug/L	0.00
Spiked Amount	5.000	Range 70 - 125	Recovery	=	77.800%	
36) 1,2-Dichloropropane-d6	6.063	67	73563	4.478	ug/L	0.00
Spiked Amount	5.000	Range 60 - 140	Recovery	=	89.600%	
41) Toluene-d8	7.307	98	197421	3.578	ug/L	0.00
Spiked Amount	5.000	Range 70 - 130	Recovery	=	71.600%	
43) trans-1,3-Dichloroprop...	7.619	79	22171	3.501	ug/L	0.00
Spiked Amount	5.000	Range 55 - 130	Recovery	=	70.000%	
46) 2-Hexanone-d5	8.088	63	114562	50.743	ug/L	0.00
Spiked Amount	50.000	Range 45 - 130	Recovery	=	101.480%	
56) 1,1,2,2-Tetrachloroeth...	10.207	84	52841	5.213	ug/L	0.00
Spiked Amount	5.000	Range 65 - 120	Recovery	=	104.200%	
66) 1,2-Dichlorobenzene-d4	11.612	152	78495	4.125	ug/L	0.00
Spiked Amount	5.000	Range 80 - 120	Recovery	=	82.400%	
Target Compounds						
					Qvalue	
16) Methylene chloride	2.500	84	8433	0.467	ug/L	97

(#) = qualifier out of range (m) = manual integration (+) = signals summed

Data Path : Z:\voasrv\HPCHEM1\MSVOA_V\Data\VV121622\
Data File : VV029694.D
Acq On : 16 Dec 2022 13:44
Operator : SY/MD
Sample : VV1216WBL02
Misc : 25.0mL/MSVOA_V/WATER
ALS Vial : 4 Sample Multiplier: 1

Instrument :
MSVOA_V
ClientSampleId :
VBLK288

Quant Time: Dec 16 20:18:18 2022
Quant Method : Z:\voasrv\HPCHEM1\MSVOA_V\Method\SFAMVTR112322WMA.M
Quant Title : TRACE VOA SFAM1.0
QLast Update : Fri Dec 16 20:17:28 2022
Response via : Initial Calibration

