

Data Path : Z:\VOASRV\HPCHEM1\MSVOA_V\DATA\VV121718\
 Data File : VV008986.D
 Acq On : 17 Dec 2018 18:05
 Operator : SY/MD
 Sample : J6387-15
 Misc : 5.0 mL/MSVOA_V/WATER
 ALS Vial : 19 Sample Multiplier: 1

Instrument :
 MSVOA_V
 ClientSampleId :
 COLB8

Quant Time: Dec 18 03:22:10 2018
 Quant Method : Z:\VOASRV\HPCHEM1\MSVOA_V\METHOD\SOMVLM121318WMA.M
 Quant Title : VOC Analysis
 QLast Update : Tue Dec 18 01:46:20 2018
 Response via : Initial Calibration

Internal Standards	R.T.	QIon	Response	Conc	Units	Dev(Min)
1) 1,4-Difluorobenzene	5.67	114	198799	50.00	ug/L	0.00
28) Chlorobenzene-d5	8.90	117	182340	50.00	ug/L	0.00
60) 1,4-Dichlorobenzene-d4	11.30	152	78467	50.00	ug/L	0.00

System Monitoring Compounds

4) Vinyl Chloride-d3	1.32	65	56673	48.81	ug/L	0.00
Spiked Amount	50.000	Range	60 - 135	Recovery	=	97.62%
7) Chloroethane-d5	1.57	69	47465	64.96	ug/L	0.00
Spiked Amount	50.000	Range	70 - 130	Recovery	=	129.92%
11) 1,1-Dichloroethene-d2	2.13	63	109307	39.87	ug/L	0.00
Spiked Amount	50.000	Range	60 - 125	Recovery	=	79.74%
21) 2-Butanone-d5	3.93	46	88451	114.11	ug/L	0.00
Spiked Amount	100.000	Range	40 - 130	Recovery	=	114.11%
24) Chloroform-d	4.41	84	121021	48.86	ug/L	0.00
Spiked Amount	50.000	Range	70 - 125	Recovery	=	97.72%
26) 1,2-Dichloroethane-d4	5.09	65	81527	51.67	ug/L	0.00
Spiked Amount	50.000	Range	70 - 125	Recovery	=	103.34%
32) Benzene-d6	5.10	84	251148	51.11	ug/L	0.00
Spiked Amount	50.000	Range	70 - 125	Recovery	=	102.22%
36) 1,2-Dichloropropane-d6	6.12	67	77077	52.20	ug/L	0.00
Spiked Amount	50.000	Range	70 - 120	Recovery	=	104.40%
41) Toluene-d8	7.36	98	234070	50.75	ug/L	0.00
Spiked Amount	50.000	Range	80 - 120	Recovery	=	101.50%
43) trans-1,3-Dichloropropene-	7.67	79	36588	48.86	ug/L	0.00
Spiked Amount	50.000	Range	60 - 125	Recovery	=	97.72%
47) 2-Hexanone-d5	8.14	63	54267	99.30	ug/L	0.00
Spiked Amount	100.000	Range	45 - 130	Recovery	=	99.30%
57) 1,1,2,2-Tetrachloroethane-	10.26	84	92603	48.58	ug/L	0.00
Spiked Amount	50.000	Range	65 - 120	Recovery	=	97.16%
64) 1,2-Dichlorobenzene-d4	11.68	152	82008	50.67	ug/L	0.00
Spiked Amount	50.000	Range	80 - 120	Recovery	=	101.34%

Target Compounds					Qvalue
34) Trichloroethene	5.96	95	56029	38.615 ug/L	99

(#) = qualifier out of range (m) = manual integration (+) = signals summed

Data Path : Z:\V0ASRV\HPCHEM1\MSVOA_V\DATA\VV121718\
Data File : VV008986.D
Acq On : 17 Dec 2018 18:05
Operator : SY/MD
Sample : J6387-15
Misc : 5.0 mL/MSVOA_V/WATER
ALS Vial : 19 Sample Multiplier: 1

Instrument :
MSVOA_V
ClientSampleId :
COLB8

Quant Time: Dec 18 03:22:10 2018
Quant Method : Z:\V0ASRV\HPCHEM1\MSVOA_V\METHOD\SOMVLM121318WMA.M
Quant Title : VOC Analysis
QLast Update : Tue Dec 18 01:46:20 2018
Response via : Initial Calibration

