

Data Path : Z:\VOASRV\HPCHEM1\MSVOA_V\DATA\VV121718\
 Data File : VV008997.D
 Acq On : 17 Dec 2018 22:29
 Operator : SY/MD
 Sample : J6383-07
 Misc : 5.0 mL/MSVOA_V/WATER
 ALS Vial : 30 Sample Multiplier: 1

Instrument :
 MSVOA_V
 ClientSampleId :
 VHBLK01

Quant Time: Dec 18 03:53:49 2018
 Quant Method : Z:\VOASRV\HPCHEM1\MSVOA_V\METHOD\SOMVLM121318WMA.M
 Quant Title : VOC Analysis
 QLast Update : Tue Dec 18 03:36:51 2018
 Response via : Initial Calibration

Internal Standards	R.T.	QIon	Response	Conc	Units	Dev(Min)
1) 1,4-Difluorobenzene	5.67	114	200284	50.00	ug/L	0.00
28) Chlorobenzene-d5	8.90	117	186024	50.00	ug/L	0.00
60) 1,4-Dichlorobenzene-d4	11.30	152	78742	50.00	ug/L	0.00

System Monitoring Compounds

4) Vinyl Chloride-d3	1.32	65	54309	46.43	ug/L	0.00
Spiked Amount	50.000	Range	60 - 135	Recovery	=	92.86%
7) Chloroethane-d5	1.56	69	47944	65.13	ug/L	0.00
Spiked Amount	50.000	Range	70 - 130	Recovery	=	130.26%#
11) 1,1-Dichloroethene-d2	2.13	63	105098	38.05	ug/L	0.00
Spiked Amount	50.000	Range	60 - 125	Recovery	=	76.10%
21) 2-Butanone-d5	3.93	46	90562	115.97	ug/L	0.00
Spiked Amount	100.000	Range	40 - 130	Recovery	=	115.97%
24) Chloroform-d	4.41	84	118816	47.61	ug/L	0.00
Spiked Amount	50.000	Range	70 - 125	Recovery	=	95.22%
26) 1,2-Dichloroethane-d4	5.08	65	80838	50.85	ug/L	0.00
Spiked Amount	50.000	Range	70 - 125	Recovery	=	101.70%
32) Benzene-d6	5.10	84	249930	49.86	ug/L	0.00
Spiked Amount	50.000	Range	70 - 125	Recovery	=	99.72%
36) 1,2-Dichloropropane-d6	6.12	67	74802	49.66	ug/L	0.00
Spiked Amount	50.000	Range	70 - 120	Recovery	=	99.32%
41) Toluene-d8	7.36	98	229189	48.70	ug/L	0.00
Spiked Amount	50.000	Range	80 - 120	Recovery	=	97.40%
43) trans-1,3-Dichloropropene-	7.67	79	35288	46.19	ug/L	0.00
Spiked Amount	50.000	Range	60 - 125	Recovery	=	92.38%
47) 2-Hexanone-d5	8.13	63	57712	103.51	ug/L	0.00
Spiked Amount	100.000	Range	45 - 130	Recovery	=	103.51%
57) 1,1,2,2-Tetrachloroethane-	10.26	84	92127	47.37	ug/L	0.00
Spiked Amount	50.000	Range	65 - 120	Recovery	=	94.74%
64) 1,2-Dichlorobenzene-d4	11.68	152	81717	50.32	ug/L	0.00
Spiked Amount	50.000	Range	80 - 120	Recovery	=	100.64%

Target Compounds	Qvalue

(#) = qualifier out of range (m) = manual integration (+) = signals summed

Data Path : Z:\VOASRV\HPCHEM1\MSVOA_V\DATA\VV121718\
Data File : VV008997.D
Acq On : 17 Dec 2018 22:29
Operator : SY/MD
Sample : J6383-07
Misc : 5.0 mL/MSVOA_V/WATER
ALS Vial : 30 Sample Multiplier: 1

Instrument :
MSVOA_V
ClientSampleId :
VHBLK01

Quant Time: Dec 18 03:53:49 2018
Quant Method : Z:\VOASRV\HPCHEM1\MSVOA_V\METHOD\SOMVLM121318WMA.M
Quant Title : VOC Analysis
QLast Update : Tue Dec 18 03:36:51 2018
Response via : Initial Calibration

