

Data Path : Z:\VOASRV\HPCHEM1\MSVOA_V\DATA\VV121718\
 Data File : VV009004.D
 Acq On : 18 Dec 2018 01:17
 Operator : SY/MD
 Sample : J6422-07
 Misc : 5.0 mL/MSVOA_V/WATER
 ALS Vial : 37 Sample Multiplier: 1

Instrument :
 MSVOA_V
 ClientSampleId :
 COLD1

Quant Time: Dec 18 04:14:05 2018
 Quant Method : Z:\VOASRV\HPCHEM1\MSVOA_V\METHOD\SOMVLM121318WMA.M
 Quant Title : VOC Analysis
 QLast Update : Tue Dec 18 03:36:51 2018
 Response via : Initial Calibration

Internal Standards	R.T.	QIon	Response	Conc	Units	Dev(Min)
1) 1,4-Difluorobenzene	5.67	114	189057	50.00	ug/L	0.00
28) Chlorobenzene-d5	8.90	117	177387	50.00	ug/L	0.00
60) 1,4-Dichlorobenzene-d4	11.30	152	73721	50.00	ug/L	0.00

System Monitoring Compounds

4) Vinyl Chloride-d3	1.32	65	53477	48.44	ug/L	0.00
Spiked Amount	50.000	Range	60 - 135	Recovery	=	96.88%
7) Chloroethane-d5	1.56	69	43748	62.96	ug/L	0.00
Spiked Amount	50.000	Range	70 - 130	Recovery	=	125.92%
11) 1,1-Dichloroethene-d2	2.13	63	105260	40.37	ug/L	0.00
Spiked Amount	50.000	Range	60 - 125	Recovery	=	80.74%
21) 2-Butanone-d5	3.93	46	86902	117.89	ug/L	0.00
Spiked Amount	100.000	Range	40 - 130	Recovery	=	117.89%
24) Chloroform-d	4.41	84	119849	50.88	ug/L	0.00
Spiked Amount	50.000	Range	70 - 125	Recovery	=	101.76%
26) 1,2-Dichloroethane-d4	5.08	65	79469	52.96	ug/L	0.00
Spiked Amount	50.000	Range	70 - 125	Recovery	=	105.92%
32) Benzene-d6	5.10	84	246733	51.61	ug/L	0.00
Spiked Amount	50.000	Range	70 - 125	Recovery	=	103.22%
36) 1,2-Dichloropropane-d6	6.12	67	75339	52.45	ug/L	0.00
Spiked Amount	50.000	Range	70 - 120	Recovery	=	104.90%
41) Toluene-d8	7.36	98	229130	51.06	ug/L	0.00
Spiked Amount	50.000	Range	80 - 120	Recovery	=	102.12%
43) trans-1,3-Dichloropropene-	7.67	79	34226	46.98	ug/L	0.00
Spiked Amount	50.000	Range	60 - 125	Recovery	=	93.96%
47) 2-Hexanone-d5	8.13	63	56621	106.50	ug/L	0.00
Spiked Amount	100.000	Range	45 - 130	Recovery	=	106.50%
57) 1,1,2,2-Tetrachloroethane-	10.26	84	92707	49.99	ug/L	0.00
Spiked Amount	50.000	Range	65 - 120	Recovery	=	99.98%
64) 1,2-Dichlorobenzene-d4	11.68	152	81213	53.41	ug/L	0.00
Spiked Amount	50.000	Range	80 - 120	Recovery	=	106.82%

Target Compounds	Qvalue

(#) = qualifier out of range (m) = manual integration (+) = signals summed

Data Path : Z:\V0ASRV\HPCHEM1\MSVOA_V\DATA\VV121718\
Data File : VV009004.D
Acq On : 18 Dec 2018 01:17
Operator : SY/MD
Sample : J6422-07
Misc : 5.0 mL/MSVOA_V/WATER
ALS Vial : 37 Sample Multiplier: 1

Instrument :
MSVOA_V
ClientSampleId :
COLD1

Quant Time: Dec 18 04:14:05 2018
Quant Method : Z:\V0ASRV\HPCHEM1\MSVOA_V\METHOD\SOMVLM121318WMA.M
Quant Title : VOC Analysis
QLast Update : Tue Dec 18 03:36:51 2018
Response via : Initial Calibration

