

Data Path : Z:\VOASRV\HPCHEM1\MSVOA_V\DATA\VV121718\
 Data File : VV008992.D
 Acq On : 17 Dec 2018 20:29
 Operator : SY/MD
 Sample : J6387-21
 Misc : 5.0 mL/MSVOA_V/WATER
 ALS Vial : 25 Sample Multiplier: 1

Instrument :
 MSVOA_V
 ClientSampleId :
 C0LC4

Quant Time: Dec 18 03:37:05 2018
 Quant Method : Z:\VOASRV\HPCHEM1\MSVOA_V\METHOD\SOMVLM121318WMA.M
 Quant Title : VOC Analysis
 QLast Update : Tue Dec 18 01:46:20 2018
 Response via : Initial Calibration

Internal Standards	R.T.	QIon	Response	Conc	Units	Dev(Min)
1) 1,4-Difluorobenzene	5.67	114	198625	50.00	ug/L	0.00
28) Chlorobenzene-d5	8.90	117	187531	50.00	ug/L	0.00
60) 1,4-Dichlorobenzene-d4	11.30	152	79258	50.00	ug/L	0.00

System Monitoring Compounds

4) Vinyl Chloride-d3	1.32	65	53026	45.71	ug/L	0.00
Spiked Amount	50.000	Range	60 - 135	Recovery	=	91.42%
7) Chloroethane-d5	1.57	69	41435	56.76	ug/L	0.00
Spiked Amount	50.000	Range	70 - 130	Recovery	=	113.52%
11) 1,1-Dichloroethene-d2	2.13	63	107992	39.42	ug/L	0.00
Spiked Amount	50.000	Range	60 - 125	Recovery	=	78.84%
21) 2-Butanone-d5	3.93	46	90775	117.21	ug/L	0.00
Spiked Amount	100.000	Range	40 - 130	Recovery	=	117.21%
24) Chloroform-d	4.41	84	118433	47.85	ug/L	0.00
Spiked Amount	50.000	Range	70 - 125	Recovery	=	95.70%
26) 1,2-Dichloroethane-d4	5.08	65	80572	51.11	ug/L	0.00
Spiked Amount	50.000	Range	70 - 125	Recovery	=	102.22%
32) Benzene-d6	5.10	84	251547	49.78	ug/L	0.00
Spiked Amount	50.000	Range	70 - 125	Recovery	=	99.56%
36) 1,2-Dichloropropane-d6	6.12	67	75527	49.74	ug/L	0.00
Spiked Amount	50.000	Range	70 - 120	Recovery	=	99.48%
41) Toluene-d8	7.36	98	233516	49.23	ug/L	0.00
Spiked Amount	50.000	Range	80 - 120	Recovery	=	98.46%
43) trans-1,3-Dichloropropene-	7.67	79	35762	46.43	ug/L	0.00
Spiked Amount	50.000	Range	60 - 125	Recovery	=	92.86%
47) 2-Hexanone-d5	8.13	63	56994	101.40	ug/L	0.00
Spiked Amount	100.000	Range	45 - 130	Recovery	=	101.40%
57) 1,1,2,2-Tetrachloroethane-	10.26	84	93334	47.60	ug/L	0.00
Spiked Amount	50.000	Range	65 - 120	Recovery	=	95.20%
64) 1,2-Dichlorobenzene-d4	11.68	152	81081	49.60	ug/L	0.00
Spiked Amount	50.000	Range	80 - 120	Recovery	=	99.20%

Target Compounds	Qvalue

(#) = qualifier out of range (m) = manual integration (+) = signals summed

Data Path : Z:\V0ASRV\HPCHEM1\MSVOA_V\DATA\VV121718\
Data File : VV008992.D
Acq On : 17 Dec 2018 20:29
Operator : SY/MD
Sample : J6387-21
Misc : 5.0 mL/MSVOA_V/WATER
ALS Vial : 25 Sample Multiplier: 1

Instrument :
MSVOA_V
ClientSampleId :
C0LC4

Quant Time: Dec 18 03:37:05 2018
Quant Method : Z:\V0ASRV\HPCHEM1\MSVOA_V\METHOD\SOMVLM121318WMA.M
Quant Title : VOC Analysis
QLast Update : Tue Dec 18 01:46:20 2018
Response via : Initial Calibration

