

Data Path : Z:\VOASRV\HPCHEM1\MSVOA_V\DATA\VV121718\
 Data File : VV009001.D
 Acq On : 18 Dec 2018 00:05
 Operator : SY/MD
 Sample : J6422-04
 Misc : 5.0 mL/MSVOA_V/WATER
 ALS Vial : 34 Sample Multiplier: 1

Instrument :
 MSVOA_V
 ClientSampleId :
 COLC8

Quant Time: Dec 18 04:08:01 2018
 Quant Method : Z:\VOASRV\HPCHEM1\MSVOA_V\METHOD\SOMVLM121318WMA.M
 Quant Title : VOC Analysis
 QLast Update : Tue Dec 18 03:36:51 2018
 Response via : Initial Calibration

Internal Standards	R.T.	QIon	Response	Conc	Units	Dev(Min)
1) 1,4-Difluorobenzene	5.67	114	195162	50.00	ug/L	0.00
28) Chlorobenzene-d5	8.90	117	180590	50.00	ug/L	0.00
60) 1,4-Dichlorobenzene-d4	11.30	152	75542	50.00	ug/L	0.00

System Monitoring Compounds

4) Vinyl Chloride-d3	1.32	65	52419	45.99	ug/L	0.00
Spiked Amount	50.000	Range	60 - 135	Recovery	=	91.98%
7) Chloroethane-d5	1.56	69	40599	56.60	ug/L	0.00
Spiked Amount	50.000	Range	70 - 130	Recovery	=	113.20%
11) 1,1-Dichloroethene-d2	2.13	63	99611	37.01	ug/L	0.00
Spiked Amount	50.000	Range	60 - 125	Recovery	=	74.02%
21) 2-Butanone-d5	3.93	46	89313	117.37	ug/L	0.00
Spiked Amount	100.000	Range	40 - 130	Recovery	=	117.37%
24) Chloroform-d	4.40	84	115044	47.31	ug/L	0.00
Spiked Amount	50.000	Range	70 - 125	Recovery	=	94.62%
26) 1,2-Dichloroethane-d4	5.08	65	79226	51.15	ug/L	0.00
Spiked Amount	50.000	Range	70 - 125	Recovery	=	102.30%
32) Benzene-d6	5.10	84	244803	50.30	ug/L	0.00
Spiked Amount	50.000	Range	70 - 125	Recovery	=	100.60%
36) 1,2-Dichloropropane-d6	6.12	67	73282	50.11	ug/L	0.00
Spiked Amount	50.000	Range	70 - 120	Recovery	=	100.22%
41) Toluene-d8	7.36	98	221254	48.43	ug/L	0.00
Spiked Amount	50.000	Range	80 - 120	Recovery	=	96.86%
43) trans-1,3-Dichloropropene-	7.67	79	33658	45.38	ug/L	0.00
Spiked Amount	50.000	Range	60 - 125	Recovery	=	90.76%
47) 2-Hexanone-d5	8.13	63	55621	102.76	ug/L	0.00
Spiked Amount	100.000	Range	45 - 130	Recovery	=	102.76%
57) 1,1,2,2-Tetrachloroethane-	10.26	84	89636	47.48	ug/L	0.00
Spiked Amount	50.000	Range	65 - 120	Recovery	=	94.96%
64) 1,2-Dichlorobenzene-d4	11.67	152	79041	50.73	ug/L	0.00
Spiked Amount	50.000	Range	80 - 120	Recovery	=	101.46%

Target Compounds	Qvalue

(#) = qualifier out of range (m) = manual integration (+) = signals summed

Data Path : Z:\V0ASRV\HPCHEM1\MSVOA_V\DATA\VV121718\
Data File : VV009001.D
Acq On : 18 Dec 2018 00:05
Operator : SY/MD
Sample : J6422-04
Misc : 5.0 mL/MSVOA_V/WATER
ALS Vial : 34 Sample Multiplier: 1

Instrument :
MSVOA_V
ClientSampleId :
C0LC8

Quant Time: Dec 18 04:08:01 2018
Quant Method : Z:\V0ASRV\HPCHEM1\MSVOA_V\METHOD\SOMVLM121318WMA.M
Quant Title : VOC Analysis
QLast Update : Tue Dec 18 03:36:51 2018
Response via : Initial Calibration

